Manual Supplement

Manual Title:52120A Getting Started Supplement Issue:4Part Number:3977724Issue Date:8/19Print Date:February 2012Page Count:2

Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual.



52120A Getting Started Manual Supplement

Change #1, 61183, 742

On page 1, replace the third bullet with:

• Offer enhanced to 140 ppm when used in closed-loop mode with a 6105A Electrical Power Standard

On page 3, delete the 10th bullet:

On page 5, replace the *How to Place and Rack Mount Product* section with:

How to Place and Rack Mount the Product

Always operate the Product in controlled electromagnetic environments such as calibration and measurement laboratories. Where RF transmitters, like mobile telephones, are not used.

This Product can be used on a bench-top or in a rack. Fluke does not have a rack mount kit for the 52120A. Fluke recommends that a drawer from the manufacturer of the rack be used.

Note

There must be space on the sides of the Product for sufficient air flow.

On page 7, replace Table 3 with:

Table 3. Line Power Cord Types Available from Fluke

Country	Fluke Part Number
UK	1998167
Europe	1998171
Australia, New Zealand	1998198
China	4121791
USA, Japan	1998209
Brazil	3841358
Other (no plug fitted)	1998211

On page 7, delete Figure 1.

8/19 1

On page 12, under **Product Connection Cables** replace the paragraph with:

Five signal cables are shipped with the Product. All are rated for 600 V. Two interchangeable low current cables, with 4 mm plugs, are used on the Product inputs or on the 2A or 20A outputs. Three heavy-duty cables, with 6 mm plugs, are used only on the 120A outputs. The short black cable loops Hi and Lo output terminals for clamp meter tests to a maximum of 120 A. The long red and black cables connect the high current outputs of the Product to the load.

To prevent an accidental disconnect, the heavy-duty cables have snap-in connectors. To operate the snap-in feature, push the connector in to the socket until the rubber insulation touches the Product and you hear a soft click. This locks the connector in the socket. To remove the connector, push in fully, then remove it. To bypass the snap-in mechanism, push the connector fully in to the socket. It is not necessary to push in on the connector to remove it when the snap-in mechanism is bypassed.

On page 15, following the **Caution** add:

Replaceable Parts

Table 9 is a list of replaceable parts. To order parts, see the How to Contact Fluke on page 1.

Table 9. Replaceable Parts

Part	Fluke Part No.
Rear Interconnect Cable	4101345
52120A-4412 Cable, High Current 52120A (set of three)	4101350
Lead Kit, 52120A Volt/Current (set of two)	4044919

On page 15 under *General Specifications* replace **Indoor use only** with:

Indoor use onlyPollution degree 2

Change #2, 64263, 470, 471

On page 1-5, add the following to the **Symbols** table:

Intrusion Protection IEC60529: IP20 Indoor use only.

Conforms to relevant South Korean EMC Standards.

On page 1-16, remove **Transient overvoltage**, **Shock and Vibration**, **Electromagnetic Environment**, **Electromagnetic Compatibility**, **Indoor use only**, and **Agency Approvals** and replace with:

2 8/19