

# Manual Supplement

Manual Title:	5500A Service	Supplement Issue:	1
Part Number:	105798	Issue Date:	1/07
Print Date:	August 1995	Page Count:	1
Revision/Date:	6, 7/06		

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

## Change #1, 39296

On page 1-17, under *Temperature Calibration (RTD) Specifications*, under **RTD Type** change:

From: Pt 395, 100  $\Omega$

To: Pt 385, 100  $\Omega$