



# Certificate of Compliance

**Certificate:** 2118607 (LR 44340C)

**Master Contract:** 162029

**Project:** 2668223

**Date Issued:** October 18, 2013

**Issued to:** **Fluke Corporation**  
6920 Seaway Blvd  
Everett, WA 98203  
USA  
Attention: Joe Swanzy

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Paul Tanaka*

**Issued by:** Paul Tanaka, AScT.

## **PRODUCTS**

**CLASS 8721 06** - LABORATORY EQUIPMENT - Electrical

**CLASS 8721 86** - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

Calibrators, Models 5700A and 5720A: cord connected, rated 100, 110, 115, 120, 200, 220, 230, 240 V, 47-63 Hz, 300 VA max.

Calibrator, Model 5730A; cord connected rated 100-120V, 220-240V, 47-63Hz, 300VA max.

### Notes:

1. The above models are cord connected, Equipment Class I, Pollution Degree 2, Installation category II;
2. Mode of operation: Continuous;
3. Environmental Conditions: Extended/as specified by manufacturer;



**Certificate:** 2118607 (LR 44340C)

**Master Contract:** 162029

**Project:** 2668223

**Date Issued:** October 18, 2013

---

## **APPLICABLE REQUIREMENTS**

### **CSA Standards:**

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

### **UL Standards:**

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

### **Reference Standards:**

EN 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

EN 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

## **CONDITIONS OF ACCEPTABILITY**

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.