

Manual Supplement

Manual Title:	5790A Operator	Supplement Issue:	2
Part Number:	893284	Issue Date:	11/16
Print Date:	January 1992	Page Count:	1
Revision/Date:	4, 8/12		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 45822, 47394

On page 1-17, under **General Specifications**, change:

From: Storage.....40 °C to 70 °C

To: Storage.....-40 °C to 70 °C

Add,

Confidence Level.....99 %

Change #2, 362

On page 6-33, under Response, remove:

- 4 = Measurement is unsettled ("U" annunciator is lit)

On page 6-38, under VAL?, replace the Response with:

Response

1. (Float) The magnitude of the measurement in volts.
2. (Float) The frequency of the measurement in Hertz.
3. (Integer) A code describing the measurement as follows. If more than one conditions is true, the highest number only is returned.
 - 0 = Measurement conditions valid, no other condition below is true
 - 1 = Frequency underrange
 - 2 = Frequency overrange
 - 3 = Measurement settled, but digital filter not full
 - 4 = Measurement is unsettled ("U" annunciator is lit)
 - 5 = Value is underrange ("LLLLLLLL" on display)
 - 6 = Value is overrange ("HHHHHHHH" on display)
 - 7 = Value is invalid ("-----" on display)