

Keysight 11853A 50 Ohm Type-N Accessory Kit

NOTICE: This document contains references to Agilent Technologies. Agilent's former Test and Measurement business has become Keysight Technologies. For more information, go to www.keysight.com.



Notices

© Keysight Technologies 1962 - 2014

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies as governed by United States and international copyright laws.

Manual Part Number

11853-90003

Edition

August 2014

Supersedes: February 2005

Keysight Technologies,
1400 Fountaingrove Parkway
Santa Rosa, CA 95403

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Keysight disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Keysight shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Keysight and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or sub-contract, Software is delivered and

licensed as “Commercial computer software” as defined in DFAR 252.227-7014 (June 1995), or as a “commercial item” as defined in FAR 2.101(a) or as “Restricted computer software” as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Keysight Technologies’ standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

11853A Accessory Kit

General Information

To obtain optimum performance from this accessory kit, observe these simple precautions:

- Keep the protective plastic end-caps on the components when not in use.
- Make connections carefully to avoid misalignment and connector damage, which will result in inaccurate measurements.
- Keep the connectors free of dirt and any particles.
- When you clean the connectors, try using compressed air first. Do not use abrasives. If further cleaning is required, refer to the *Connector Care for RF and Microwave Coaxial Connectors* document. It can be viewed online by searching for part number 08510-90064 at www.keysight.com.

Description

The 11853A 50 ohm Type-N Accessory kit contains the RF components required for measurement of devices with 50 ohm, type-N connectors, and for using a network analyzer with a type-N configuration. Each kit contains a type-N (female) short, a type-N (male) short, two type-N (male) barrels, two type-N (female) barrels, and a storage case.

Contents

The 11853A kit contains the following:

- Type-N (f) short (Keysight part number 85032-60009)
- Type-N (m) short (Keysight part number 85032-60008)
- Type-N (m) barrell (Keysight part number 1250-1475), quantity 2
- Type-N (f) barrell (Keysight part number 1250-1472), quantity 2
- Storage box, (Keysight part number 9211-1582)
- Storage box foam pad, (Keysight part number 9220-2550)
- Operating and Service Manual (Keysight part number 11853-90003)

Performance Characteristics

Return loss of adapters: typically 40 dB up to 1.3 GHz.

General Characteristics

Weight

Net: 0.91 kg (2 lb.)

Shipping: 1.36 kg (3 lb.)

Dimensions

Width: 168 mm

Depth: 114 mm

Height: 51 mm

(6 5/8 inches x 4 1/2 inches x 2 inches)

Replaceable Parts

There are no replaceable components for the components. A worn or damaged component must be replaced in whole.

Equipment and Supplies

The following equipment and supplies are required for the maintenance and use of, but are not supplied with, your 11853A accessories kit.

Item	Part Number
Type-N gage set (part of the 85032F calibration kit)	85054-60059
Torque wrench, 3/4", 136 N-cm (12 in-lb)	8710-1766
Document: <i>Connector Care for RF and Microwave Coaxial Connectors</i>	08510-90064

Connector Maintenance

For information on cleaning and gaging connectors, refer to the *Connector Care for RF and Microwave Coaxial Connectors* document. It can be viewed online by searching for part number 08510-90064 at www.keysight.com.

Contacting Keysight

Assistance with test and measurement needs and information on finding a local Keysight office are available on the Web at:

www.keysight.com/find/assist

NOTE In any correspondence or telephone conversation, refer to the Keysight product by its model number and full serial number. With this information, the Keysight representative can determine whether your product is still within its warranty period.

This information is subject to change without notice.
© Keysight Technologies 1962 - 2014
August 2014



11853-90003
www.keysight.com