



Agilent Technologies

Agilent Technologies (Malaysia) Sdn.Bhd.
(012767-W)
Bayan Lepas Free Industrial Zone
11900 Penang, Malaysia



5962-0476

Certificate Of Calibration

ANSI/NCSL Z540-1-1994

Certificate No: PENANG513933-581617-1

Manufacturer: Agilent Technologies

Description: Nano Volt/Micro Ohm Meter

Model No: 34420A

Serial No: MY42003588

Options Installed With Specifications: N/A

Date of Calibration: 13-MAR-2010

Temperature: (23 +/- 2) deg. C

Humidity: (20 to 70)% RH

Procedure: 34420-90010

This certifies that the above product was calibrated in compliance with ANSI/NCSL Z540-1-1994 and a quality system registered to ISO 9001:2008, using applicable Agilent Technologies' procedures.

As Received: Factory tested. No incoming data available.

As Shipped Conditions: At the completion of the calibration, measured values were IN SPECIFICATION at the points tested.

These calibration procedures and test points are those recommended in a procedure developed by Agilent. Measurement uncertainty is accounted for in the determination of the test limits for compliance to specifications.

Remarks or special requirements:

This calibration includes an attached measurement report

Traceability Information: Traceability is to the International System of Units (SI), consensus standards or ratio type measurements through national standards realized and maintained by the NIST U.S., NRC Canada, NMIJ Japan, KRISS Korea, Euramet members (NPL, PTB, etc.), NML-SIRIM in Malaysia or other National Measurement Institutes signatories to the CIPM MRA. Supporting documentation relative to traceability is available for review by appointment. This report shall not be reproduced, except in full, without prior written approval of the calibration facility.

Calibration Equipment Used:

Date used: Date equipment used in this calibration.

<u>Model Number</u>	<u>Model Description</u>	<u>Trace Number</u>	<u>Date Used</u>	<u>Cal Due Date</u>
FLU 742A-100	Resistance Standard	PZ00062	13-MAR-2010	30-SEP-2010
FLU 742A-10	Resistance Standard	PZ00061	13-MAR-2010	30-SEP-2010
FLU 742A-100K	Resistance Standard	PZ00063	13-MAR-2010	01-OCT-2010
FLU 742A-1	Resistance Standard	PZ00060	13-MAR-2010	30-SEP-2010
FLU 5720A	Calibrator	PZ00034	13-MAR-2010	29-APR-2010
3458A	8-1/2 Digit Multimeter	PZ00005	13-MAR-2010	08-SEP-2010

Print Date: 06-OCT-2011

Agilent Technologies

DD	MM	YY	BY:
CAL	13	03	10 SE
DUE			

Tay Eng Su
Quality Manager