

Manufacturer	HEWLETT-PACKARD	Calibration date	June 25 2018
Model Number	3458A	Ambient Temperature	24.7 °C
Serial	3458B	Relative Humidity	34.80 %
ID Number	IM5700A	Pressure	1007.0 hPa
Notes	Test A7Mark,TA9,A10	Test type	M5700

This note is test dummy text block for further use. It allow to include user information for further reference

Reference standard	Mfg	Model	Options	Serial / Unc	CEID	Calibration date	Due date
TEST MFC	Fluke	5700A	None	x26	ID02	05/17/2018	06/17/2018
DMM	HP	3458A	001,X02	MY45040325	XD2	04/23/2018	07/23/2018
STDR	ESI	SR104	10000.0011 KΩ	±1.00 ppm	XR04	05/25/2016	05/25/2018
STDR	xDevs.com/Fluke	SL935	1.00005616 Ω	±0.17 ppm	XR03	10/04/2018	10/04/2019
STDR	xDevs.com/Fluke	SL935	9999.9747 kΩ	±0.33 ppm	XR02	10/04/2018	10/04/2019
DC STD	xDevs.com	792X[2]	10.000009 VDC	±2.2 ppm	XD01	02/16/2018	08/16/2018

MFC last calibrated	1 days ago	MFC since DCV ZERO	1 days ago
MFC since WBFLAT	1 days ago	MFC since WBGAIN	1 days ago
MFC Confidence level	24h 95%	MFC Calibrate date	Debug
MFC Calibrate date Zero	Debug	Calibrate date WB Flatness	Debug
Calibrate date WB Gain	Debug	CAL CONST 6.5V reference voltage	1
CAL CONST 13V reference voltage	1	CAL CONST 22V range positive zero	1
CAL CONST 22V range negative zero	1	CAL CONST DAC Linearity	1
CAL CONST 10KOHM true output resistance	1	CAL CONST 10KOHM standard resistance	1
CAL CONST, Zero calibration temperature	1	CAL CONST, All calibration temp	1

This note is test MFC dummy text block for further use.
Calibrator was warmed up >8 hours.

Main DC Voltage ranges performance test.

Checks zero offset and +/-FS calibration on all ranges

The following test for the offset voltage specification using MFC 0V source in 4-wire ext sense mode as reference.

DCV gain range points verify gain of the DC voltage function, using uncorrected 24-hour MFC output. DC voltage offset of DUT is nulled before FS tests.

Test Description	Expected Value	Measured Value	Measurement Uncertainty	Lower Limit	Upper Limit	Deviation	DUT Spec	Test Status
Short 0 mVDC	0.0000000E+00	-1.26 µV	0.75 µV	-0.910 µV	0.910 µV	N/A	0.16 µV	FAIL
Short 0.0 VDC	0.0000000E+00	-1.13 µV	0.75 µV	-0.900 µV	0.900 µV	N/A	0.15 µV	FAIL
Short 00.0 VDC	0.0000000E+00	-0.53 µV	0.75 µV	-1.070 µV	1.070 µV	N/A	0.32 µV	PASS
Short 000.0 VDC	0.0000000E+00	-17.16 µV	0.75 µV	-14.750 µV	14.750 µV	N/A	14.00 µV	FAIL
Short 0000.0 VDC	0.0000000E+00	64.80 µV	0.75 µV	-41.750 µV	41.750 µV	N/A	41.00 µV	FAIL
DCV Test	0.1V-1000V	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
0.1 VDC (0.10 Range)	0.1000000	0.099999604	3.81 ppm	0.099999069	0.10000093	-3.959 ppm	5.50 ppm	PASS 42.53 %
-0.1 VDC (0.10 Range)	-0.1000000	-0.099999834	3.81 ppm	-0.10000093	-0.099999069	-1.659 ppm	5.50 ppm	PASS 17.82 %
0.1 VDC (1.00 Range)	0.1000000	0.10000035	2.45 ppm	0.099999575	0.10000043	3.528 ppm	1.80 ppm	PASS 83.01 %
0.2 VDC (1.00 Range)	0.2000000	0.20000043	2.45 ppm	0.19999915	0.20000085	2.147 ppm	1.80 ppm	PASS 50.53 %
1.0 VDC (1.00 Range)	1.0000000	1.0000006	2.45 ppm	0.99999575	1.0000042	0.575 ppm	1.80 ppm	PASS 13.53 %
-0.1 VDC (1.00 Range)	-0.1000000	-0.099999489	2.45 ppm	-0.10000043	-0.099999575	-5.112 ppm	1.80 ppm	FAIL 120.28 %
-0.2 VDC (1.00 Range)	-0.2000000	-0.19999955	2.45 ppm	-0.20000085	-0.19999915	-2.251 ppm	1.80 ppm	PASS 52.98 %
-1.0 VDC (1.00 Range)	-1.0000000	-0.9999996	2.45 ppm	-1.0000042	-0.99999575	-0.395 ppm	1.80 ppm	PASS 9.31 %
1.0 VDC (10.00 Range)	1.0000000	1.000001	1.47 ppm	0.99999798	1.000002	0.979 ppm	0.55 ppm	PASS 48.47 %
2.0 VDC (10.00 Range)	2.0000000	2.000002	1.47 ppm	1.999996	2.000004	0.125 ppm	0.55 ppm	PASS 6.19 %
10.0 VDC (10.00 Range)	10.0000000	10.000002	1.47 ppm	9.9999798	10.00002	0.219 ppm	0.55 ppm	PASS 10.84 %
-1.0 VDC (10.00 Range)	-1.0000000	-0.99999921	1.47 ppm	-1.000002	-0.99999798	-0.792 ppm	0.55 ppm	PASS 39.20 %
-2.0 VDC (10.00 Range)	-2.0000000	-1.9999984	1.47 ppm	-2.000004	-1.999996	-0.795 ppm	0.55 ppm	PASS 39.38 %
-10.0 VDC (10.00 Range)	-10.0000000	-9.9999994	1.47 ppm	-10.00002	-9.9999798	-0.065 ppm	0.55 ppm	PASS 3.21 %
10 VDC (100.00 Range)	10.0000000	10.000014	2.36 ppm	9.9999484	10.000052	1.440 ppm	2.80 ppm	PASS 27.91 %
20 VDC (100.00 Range)	20.0000000	20.000015	2.36 ppm	19.999897	20.000103	0.731 ppm	2.80 ppm	PASS 14.18 %
100 VDC (100.00 Range)	100.0000000	100.00007	2.36 ppm	99.999484	100.00052	0.718 ppm	2.80 ppm	PASS 13.91 %
-10 VDC (100.00 Range)	-10.0000000	-10.000005	2.36 ppm	-10.000052	-9.9999484	0.549 ppm	2.80 ppm	PASS 10.64 %
-20 VDC (100.00 Range)	-20.0000000	-20.000013	2.36 ppm	-20.000103	-19.999897	0.671 ppm	2.80 ppm	PASS 13.00 %
-100 VDC (100.00 Range)	-100.0000000	-100.00004	2.36 ppm	-100.00052	-99.999484	0.359 ppm	2.80 ppm	PASS 6.96 %
100 VDC (1000.00 Range)	100.0000000	100.0001	2.85 ppm	99.999455	100.00055	1.017 ppm	2.60 ppm	PASS 18.66 %
200 VDC (1000.00 Range)	200.0000000	200.00008	2.85 ppm	199.99891	200.00109	0.406 ppm	2.60 ppm	PASS 7.44 %
1000 VDC (1000.00 Range)	1000.0000000	1000.002	2.85 ppm	999.98255	1000.0175	2.027 ppm	2.60 ppm	PASS 11.62 %
-100 VDC (1000.00 Range)	-100.0000000	-100.00007	2.85 ppm	-100.00055	-99.999455	0.691 ppm	2.60 ppm	PASS 12.68 %
-200 VDC (1000.00 Range)	-200.0000000	-200.00001	2.85 ppm	-200.00109	-199.99891	0.068 ppm	2.60 ppm	PASS 1.25 %
-1000 VDC (1000.00 Range)	-1000.0000000	-1000.0023	2.85 ppm	-1000.0175	-999.98255	2.285 ppm	2.60 ppm	PASS 34.89 %

Additional test for **combined DUT+MFC** DC Voltage Integral Linearity (INL) using fixed 10V range. Integral linearity is a measure of the device's deviation from ideal linear behaviour.

DCV Linearity	1V Range	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
1.0999999	1.0999999	1.1000003	2.73 ppm	1.099996	1.100004	0.38 ppm	0.55 ppm	PASS 11.54 %
0.9999999	0.9999999	1.0000005	2.73 ppm	0.9999966	1.000003	0.60 ppm	0.55 ppm	PASS 18.41 %
0.9000000	0.9000000	0.9000005	2.73 ppm	0.899997	0.900003	0.60 ppm	0.55 ppm	PASS 18.16 %
0.8888888	0.8888888	0.8888895	2.73 ppm	0.8888859	0.8888917	0.76 ppm	0.55 ppm	PASS 23.28 %
0.8000000	0.8000000	0.8000006	2.73 ppm	0.7999974	0.8000026	0.80 ppm	0.55 ppm	PASS 24.54 %
0.7777777	0.7777777	0.7777783	2.73 ppm	0.7777751	0.7777803	0.77 ppm	0.55 ppm	PASS 23.34 %
0.7000000	0.7000000	0.7000005	2.73 ppm	0.6999977	0.7000023	0.72 ppm	0.55 ppm	PASS 21.98 %
0.6666666	0.6666666	0.6666672	2.73 ppm	0.6666644	0.6666688	0.87 ppm	0.55 ppm	PASS 26.37 %
0.6000000	0.6000000	0.6000004	2.73 ppm	0.599998	0.600002	0.71 ppm	0.55 ppm	PASS 21.69 %
0.5555555	0.5555555	0.5555559	2.73 ppm	0.5555537	0.5555573	0.73 ppm	0.55 ppm	PASS 22.17 %
0.5000000	0.5000000	0.5000005	2.73 ppm	0.4999984	0.5000016	0.96 ppm	0.55 ppm	PASS 29.29 %
0.4444444	0.4444444	0.4444449	2.73 ppm	0.4444429	0.4444459	1.05 ppm	0.55 ppm	PASS 32.11 %
0.4000000	0.4000000	0.4000005	2.73 ppm	0.3999987	0.4000013	1.17 ppm	0.55 ppm	PASS 35.75 %
0.3333333	0.3333333	0.3333336	2.73 ppm	0.3333322	0.3333344	0.97 ppm	0.55 ppm	PASS 29.48 %
0.3000000	0.3000000	0.3000004	2.73 ppm	0.299999	0.300001	1.21 ppm	0.55 ppm	PASS 37.02 %
0.2222222	0.2222222	0.2222225	2.73 ppm	0.2222215	0.2222229	1.19 ppm	0.55 ppm	PASS 36.21 %
0.2000000	0.2000000	0.2000002	2.73 ppm	0.1999993	0.2000007	0.82 ppm	0.55 ppm	PASS 25.07 %
0.1234567	0.1234567	0.1234569	2.73 ppm	0.1234563	0.1234571	1.88 ppm	0.55 ppm	PASS 57.42 %
0.1111111	0.1111111	0.1111113	2.73 ppm	0.1111107	0.1111115	1.36 ppm	0.55 ppm	PASS 41.33 %
0.1000000	0.1000000	0.1000002	2.73 ppm	0.09999967	0.1000003	1.73 ppm	0.55 ppm	PASS 52.76 %
0.0987654	0.0987654	0.0987655	2.73 ppm	0.09876508	0.09876572	1.12 ppm	0.55 ppm	PASS 34.08 %
0.0111111	0.0111111	0.0111111	2.73 ppm	0.01111106	0.01111114	1.67 ppm	0.55 ppm	PASS 50.84 %
-0.0111111	-0.0111111	-0.0111106	2.73 ppm	-0.01111114	-0.01111106	-42.95 ppm	0.55 ppm	FAIL 1309.36 %
-0.0987654	-0.0987654	-0.0987650	2.73 ppm	-0.09876572	-0.09876508	-3.85 ppm	0.55 ppm	FAIL 117.50 %
-0.1000000	-0.1000000	-0.0999996	2.73 ppm	-0.1000003	-0.09999967	-4.04 ppm	0.55 ppm	FAIL 123.11 %
-0.1111111	-0.1111111	-0.1111106	2.73 ppm	-0.1111115	-0.1111107	-4.22 ppm	0.55 ppm	FAIL 128.64 %
-0.1234567	-0.1234567	-0.1234562	2.73 ppm	-0.1234571	-0.1234563	-3.78 ppm	0.55 ppm	FAIL 115.16 %
-0.2000000	-0.2000000	-0.1999996	2.73 ppm	-0.2000007	-0.1999993	-2.25 ppm	0.55 ppm	PASS 68.47 %
-0.2222222	-0.2222222	-0.2222219	2.73 ppm	-0.2222229	-0.2222215	-1.55 ppm	0.55 ppm	PASS 47.19 %
-0.3000000	-0.3000000	-0.2999997	2.73 ppm	-0.300001	-0.299999	-1.01 ppm	0.55 ppm	PASS 30.79 %
-0.3333333	-0.3333333	-0.3333331	2.73 ppm	-0.3333344	-0.3333322	-0.54 ppm	0.55 ppm	PASS 16.40 %
-0.4000000	-0.4000000	-0.3999997	2.73 ppm	-0.4000013	-0.3999987	-0.65 ppm	0.55 ppm	PASS 19.96 %
-0.4444444	-0.4444444	-0.4444442	2.73 ppm	-0.4444459	-0.4444429	-0.45 ppm	0.55 ppm	PASS 13.82 %
-0.5000000	-0.5000000	-0.4999998	2.73 ppm	-0.5000016	-0.4999984	-0.49 ppm	0.55 ppm	PASS 14.82 %
-0.5555555	-0.5555555	-0.5555554	2.73 ppm	-0.5555573	-0.5555537	-0.23 ppm	0.55 ppm	PASS 7.05 %
-0.6000000	-0.6000000	-0.5999999	2.73 ppm	-0.600002	-0.599998	-0.23 ppm	0.55 ppm	PASS 6.88 %
-0.6666666	-0.6666666	-0.6666664	2.73 ppm	-0.6666688	-0.6666644	-0.30 ppm	0.55 ppm	PASS 9.11 %
-0.7000000	-0.7000000	-0.6999998	2.73 ppm	-0.7000023	-0.6999977	-0.28 ppm	0.55 ppm	PASS 8.43 %
-0.7777777	-0.7777777	-0.7777774	2.73 ppm	-0.7777803	-0.7777751	-0.35 ppm	0.55 ppm	PASS 10.71 %
-0.8000000	-0.8000000	-0.7999998	2.73 ppm	-0.8000026	-0.7999974	-0.26 ppm	0.55 ppm	PASS 7.88 %
-0.8888888	-0.8888888	-0.8888886	2.73 ppm	-0.8888917	-0.8888859	-0.26 ppm	0.55 ppm	PASS 8.03 %
-0.9000000	-0.9000000	-0.8999997	2.73 ppm	-0.900003	-0.899997	-0.29 ppm	0.55 ppm	PASS 8.88 %
-0.9999999	-0.9999999	-0.9999996	2.73 ppm	-1.000003	-0.9999966	-0.29 ppm	0.55 ppm	PASS 8.87 %
-1.0999999	-1.0999999	-1.0999995	2.73 ppm	-1.100004	-1.099996	-0.38 ppm	0.55 ppm	PASS 11.54 %
DCV Linearity	10V Range	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
10.999999	10.999999	11.0000000	2.73 ppm	10.99996	11.00004	0.09 ppm	0.55 ppm	PASS 2.88 %
10.101010	10.101010	10.1010116	2.73 ppm	10.10098	10.10104	0.15 ppm	0.55 ppm	PASS 4.72 %
10.000000	10.000000	10.0000017	2.73 ppm	9.999967	10.00003	0.17 ppm	0.55 ppm	PASS 5.05 %
9.999999	9.999999	10.0000007	2.73 ppm	9.999966	10.00003	0.17 ppm	0.55 ppm	PASS 5.10 %
9.000000	9.000000	9.0000017	2.73 ppm	8.99997	9.00003	0.19 ppm	0.55 ppm	PASS 5.65 %
8.888888	8.888888	8.8888897	2.73 ppm	8.888859	8.888917	0.19 ppm	0.55 ppm	PASS 5.78 %
8.000000	8.000000	8.0000013	2.73 ppm	7.999974	8.000026	0.16 ppm	0.55 ppm	PASS 5.00 %
7.777777	7.777777	7.7777783	2.73 ppm	7.777751	7.777803	0.17 ppm	0.55 ppm	PASS 5.19 %
7.000000	7.000000	7.0000007	2.73 ppm	6.999977	7.000023	0.10 ppm	0.55 ppm	PASS 3.04 %
6.666666	6.666666	6.6666672	2.73 ppm	6.666644	6.666688	0.18 ppm	0.55 ppm	PASS 5.46 %
6.000000	6.000000	6.0000009	2.73 ppm	5.99998	6.00002	0.16 ppm	0.55 ppm	PASS 4.79 %
5.555555	5.555555	5.5555550	2.73 ppm	5.555537	5.555573	0.01 ppm	0.55 ppm	PASS 0.23 %
5.000000	5.000000	5.0000001	2.73 ppm	4.999984	5.000016	0.02 ppm	0.55 ppm	PASS 0.62 %
4.444444	4.444444	4.4444439	2.73 ppm	4.444429	4.444459	-0.02 ppm	0.55 ppm	PASS 0.76 %
4.000000	4.000000	3.9999996	2.73 ppm	3.999987	4.000013	-0.10 ppm	0.55 ppm	PASS 3.06 %
3.333333	3.333333	3.3333324	2.73 ppm	3.333322	3.333344	-0.17 ppm	0.55 ppm	PASS 5.19 %

3.000000	3.000000	2.9999994	2.73 ppm	2.99999	3.00001	-0.21 ppm	0.55 ppm	PASS 6.40 %
2.222222	2.222222	2.2222215	2.73 ppm	2.222215	2.222229	-0.24 ppm	0.55 ppm	PASS 7.41 %
2.000000	2.000000	1.9999991	2.73 ppm	1.999993	2.000007	-0.46 ppm	0.55 ppm	PASS 13.87 %
1.111111	1.111111	1.1111096	2.73 ppm	1.111107	1.111115	-1.23 ppm	0.55 ppm	PASS 37.45 %
1.000000	1.000000	0.9999988	3.86 ppm	0.9999956	1.000004	-1.21 ppm	0.55 ppm	PASS 27.39 %
0.555555	0.555555	0.5555532	7.27 ppm	0.5555507	0.5555593	-3.15 ppm	0.55 ppm	PASS 40.34 %
-0.555555	-0.555555	-0.5555549	7.27 ppm	-0.5555593	-0.5555507	-0.24 ppm	0.55 ppm	PASS 3.12 %
-1.000000	-1.000000	-0.9999999	3.86 ppm	-1.000004	-0.9999956	-0.08 ppm	0.55 ppm	PASS 1.71 %
-1.111111	-1.111111	-1.1111108	2.73 ppm	-1.111115	-1.111107	-0.16 ppm	0.55 ppm	PASS 4.77 %
-2.000000	-2.000000	-2.0000001	2.73 ppm	-2.000007	-1.999993	0.03 ppm	0.55 ppm	PASS 0.82 %
-2.222222	-2.222222	-2.2222222	2.73 ppm	-2.222229	-2.222215	0.09 ppm	0.55 ppm	PASS 2.60 %
-3.000000	-3.000000	-3.0000002	2.73 ppm	-3.00001	-2.99999	0.05 ppm	0.55 ppm	PASS 1.67 %
-3.333333	-3.333333	-3.3333333	2.73 ppm	-3.333344	-3.333322	0.09 ppm	0.55 ppm	PASS 2.71 %
-4.000000	-4.000000	-4.0000004	2.73 ppm	-4.000013	-3.999987	0.09 ppm	0.55 ppm	PASS 2.82 %
-4.444444	-4.444444	-4.4444442	2.73 ppm	-4.444459	-4.444429	0.04 ppm	0.55 ppm	PASS 1.32 %
-5.000000	-5.000000	-5.0000001	2.73 ppm	-5.000016	-4.999984	0.02 ppm	0.55 ppm	PASS 0.51 %
-5.555555	-5.555555	-5.5555551	2.73 ppm	-5.555573	-5.555537	0.02 ppm	0.55 ppm	PASS 0.49 %
-6.000000	-6.000000	-6.0000003	2.73 ppm	-6.00002	-5.99998	0.06 ppm	0.55 ppm	PASS 1.70 %
-6.666666	-6.666666	-6.6666663	2.73 ppm	-6.666688	-6.666644	0.05 ppm	0.55 ppm	PASS 1.45 %
-7.000000	-7.000000	-7.0000004	2.73 ppm	-7.000023	-6.999977	0.05 ppm	0.55 ppm	PASS 1.64 %
-7.777777	-7.777777	-7.7777776	2.73 ppm	-7.777803	-7.777751	0.07 ppm	0.55 ppm	PASS 2.23 %
-8.000000	-8.000000	-8.0000007	2.73 ppm	-8.000026	-7.999974	0.09 ppm	0.55 ppm	PASS 2.84 %
-8.888888	-8.888888	-8.8888886	2.73 ppm	-8.888917	-8.888859	0.07 ppm	0.55 ppm	PASS 2.12 %
-9.000000	-9.000000	-9.0000009	2.73 ppm	-9.00003	-8.99997	0.10 ppm	0.55 ppm	PASS 2.98 %
-9.999999	-9.999999	-9.9999993	2.73 ppm	-10.00003	-9.999966	0.03 ppm	0.55 ppm	PASS 0.94 %
-10.000000	-10.000000	-10.0000005	2.73 ppm	-10.00003	-9.999967	0.05 ppm	0.55 ppm	PASS 1.51 %
-10.101010	-10.101010	-10.1010110	2.73 ppm	-10.10104	-10.10098	0.10 ppm	0.55 ppm	PASS 3.10 %
-10.999999	-10.999999	-10.9999998	2.73 ppm	-11.00004	-10.99996	0.07 ppm	0.55 ppm	PASS 2.17 %
DCV Linearity	100V Range	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
100.99999	100.99999	101.0000798	2.73 ppm	100.99966	101.00032	0.89 ppm	0.55 ppm	PASS 19.79 %
100.10101	100.10101	100.1010920	2.73 ppm	100.10068	100.10134	0.82 ppm	0.55 ppm	PASS 18.29 %
100.00000	100.00000	100.0000758	2.73 ppm	99.999672	100.00033	0.76 ppm	0.55 ppm	PASS 16.93 %
99.99999	99.99999	100.0000586	2.73 ppm	99.999662	100.00032	0.69 ppm	0.55 ppm	PASS 15.32 %
90.00000	90.00000	90.0000659	2.73 ppm	89.999705	90.000295	0.73 ppm	0.55 ppm	PASS 22.34 %
88.88888	88.88888	88.8889401	2.73 ppm	88.888588	88.889172	0.68 ppm	0.55 ppm	PASS 20.61 %
80.00000	80.00000	80.0000583	2.73 ppm	79.999738	80.000262	0.73 ppm	0.55 ppm	PASS 22.22 %
77.77777	77.77777	77.7778251	2.73 ppm	77.777515	77.778025	0.71 ppm	0.55 ppm	PASS 21.58 %
70.00000	70.00000	70.0000464	2.73 ppm	69.99977	70.00023	0.66 ppm	0.55 ppm	PASS 20.21 %
66.66666	66.66666	66.6667054	2.73 ppm	66.666441	66.666879	0.68 ppm	0.55 ppm	PASS 20.74 %
60.00000	60.00000	60.0000406	2.73 ppm	59.999803	60.000197	0.68 ppm	0.55 ppm	PASS 20.65 %
55.55555	55.55555	55.5555877	2.73 ppm	55.555368	55.555732	0.68 ppm	0.55 ppm	PASS 20.68 %
50.00000	50.00000	50.0000346	2.73 ppm	49.999836	50.000164	0.69 ppm	0.55 ppm	PASS 21.11 %
44.44444	44.44444	44.4444661	2.73 ppm	44.444294	44.444586	0.59 ppm	0.55 ppm	PASS 17.91 %
40.00000	40.00000	40.0000216	2.73 ppm	39.999869	40.000131	0.54 ppm	0.55 ppm	PASS 16.43 %
33.33333	33.33333	33.3333437	2.73 ppm	33.333221	33.333439	0.41 ppm	0.55 ppm	PASS 12.49 %
30.00000	30.00000	30.0000084	2.73 ppm	29.999902	30.000098	0.28 ppm	0.55 ppm	PASS 8.57 %
22.22222	22.22222	22.2222214	2.73 ppm	22.222147	22.222293	0.06 ppm	0.55 ppm	PASS 1.89 %
20.00000	20.00000	19.9999951	2.73 ppm	19.999934	20.000066	-0.24 ppm	0.55 ppm	PASS 7.41 %
11.11111	11.11111	11.1110879	2.73 ppm	11.111075	11.111147	-2.07 ppm	0.55 ppm	PASS 63.25 %
10.00000	10.00000	9.9999875	2.73 ppm	9.9999672	10.000033	-1.25 ppm	0.55 ppm	PASS 38.12 %
9.87654	9.87654	9.8765185	2.73 ppm	9.8765106	9.8765754	-2.48 ppm	0.55 ppm	PASS 75.56 %
-9.87654	-9.87654	-9.8765384	2.73 ppm	-9.8765754	-9.8765106	-0.46 ppm	0.55 ppm	PASS 14.11 %
-10.00000	-10.00000	-9.9999982	2.73 ppm	-10.000033	-9.9999672	-0.18 ppm	0.55 ppm	PASS 5.37 %
-11.11111	-11.11111	-11.1111066	2.73 ppm	-11.111147	-11.111075	-0.39 ppm	0.55 ppm	PASS 11.98 %
-20.00000	-20.00000	-20.0000019	2.73 ppm	-20.000066	-19.999934	0.09 ppm	0.55 ppm	PASS 2.82 %
-22.22222	-22.22222	-22.2222266	2.73 ppm	-22.222293	-22.222147	0.30 ppm	0.55 ppm	PASS 9.02 %
-30.00000	-30.00000	-30.0000133	2.73 ppm	-30.000098	-29.999902	0.44 ppm	0.55 ppm	PASS 13.56 %
-33.33333	-33.33333	-33.3333487	2.73 ppm	-33.333439	-33.333221	0.56 ppm	0.55 ppm	PASS 17.08 %
-40.00000	-40.00000	-40.0000250	2.73 ppm	-40.000131	-39.999869	0.63 ppm	0.55 ppm	PASS 19.06 %
-44.44444	-44.44444	-44.4444615	2.73 ppm	-44.444586	-44.444294	0.48 ppm	0.55 ppm	PASS 14.74 %
-50.00000	-50.00000	-50.0000267	2.73 ppm	-50.000164	-49.999836	0.53 ppm	0.55 ppm	PASS 16.31 %
-55.55555	-55.55555	-55.5555837	2.73 ppm	-55.555732	-55.555368	0.61 ppm	0.55 ppm	PASS 18.47 %
-60.00000	-60.00000	-60.0000343	2.73 ppm	-60.000197	-59.999803	0.57 ppm	0.55 ppm	PASS 17.44 %
-66.66666	-66.66666	-66.6666954	2.73 ppm	-66.666879	-66.666441	0.53 ppm	0.55 ppm	PASS 16.20 %
-70.00000	-70.00000	-70.0000381	2.73 ppm	-70.00023	-69.99977	0.54 ppm	0.55 ppm	PASS 16.60 %
-77.77777	-77.77777	-77.7778108	2.73 ppm	-77.778025	-77.777515	0.52 ppm	0.55 ppm	PASS 16.00 %

-80.00000	-80.00000	-80.0000385	2.73 ppm	-80.000262	-79.999738	0.48 ppm	0.55 ppm	PASS 14.66 %
-88.88888	-88.88888	-88.8889214	2.73 ppm	-88.889172	-88.888588	0.47 ppm	0.55 ppm	PASS 14.20 %
-90.00000	-90.00000	-90.0000377	2.73 ppm	-90.000295	-89.999705	0.42 ppm	0.55 ppm	PASS 12.77 %
-99.99999	-99.99999	-100.0000270	2.73 ppm	-100.00032	-99.999662	0.37 ppm	0.55 ppm	PASS 17.80 %
-100.00000	-100.00000	-100.0000415	2.73 ppm	-100.00033	-99.999672	0.42 ppm	0.55 ppm	PASS 19.97 %
-100.10101	-100.10101	-100.1010535	2.73 ppm	-100.10134	-100.10068	0.43 ppm	0.55 ppm	PASS 20.89 %
-100.99999	-100.99999	-101.0000321	2.73 ppm	-101.00032	-100.99966	0.42 ppm	0.55 ppm	PASS 20.14 %

4W test procedure for all test points that verify Gain of the OHMF function. 4-wire kelvin connection is used between DMM and MFC. 1GΩ resistance range is tested using the external standard, as MFC unable to provide this range value.

OHM Test	1 Ohm to 1 GOhm	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
1 Ω	0.9997848	0.99979148	27.0 ppm	9.9974981E-01	9.9981979E-01	6.680 ppm	8.0 ppm	PASS 19.09 %
1.9 Ω	1.8998263	1.89984	20.0 ppm	1.8997731E+00	1.8998795E+00	7.219 ppm	8.0 ppm	PASS 25.78 %
10 Ω	9.999695	9.9996817	4.0 ppm	9.9995750E+00	9.9998150E+00	-1.326 ppm	8.0 ppm	PASS 11.05 %
19 Ω	18.9982	18.998289	3.5 ppm	1.8998020E+01	1.8998380E+01	4.701 ppm	6.0 ppm	PASS 49.48 %
100 Ω	99.99801	99.997992	1.6 ppm	9.9997250E+01	9.9998770E+01	-0.177 ppm	6.0 ppm	PASS 2.33 %
190 Ω	189.98804	189.98812	1.6 ppm	1.8998732E+02	1.8998876E+02	0.413 ppm	2.2 ppm	PASS 10.88 %
1.0 kΩ	999.9383	999.93741	1.6 ppm	9.9993450E+02	9.9994210E+02	-0.893 ppm	2.2 ppm	PASS 23.51 %
1.9 kΩ	1899.8875	1899.8852	1.6 ppm	1.8998803E+03	1.8998947E+03	-1.185 ppm	2.2 ppm	PASS 31.19 %
10 kΩ	9999.581	9999.5602	1.6 ppm	9.9995430E+03	9.9996190E+03	-2.079 ppm	2.2 ppm	PASS 54.71 %
19 kΩ	18999.12	18999.062	1.6 ppm	1.8999048E+04	1.8999192E+04	-3.060 ppm	2.2 ppm	PASS 80.51 %
100 kΩ	99992.75	99992.302	1.6 ppm	9.9992370E+04	9.9993130E+04	-4.479 ppm	2.2 ppm	FAIL 117.86 %
190 kΩ	189998.12	189997.11	1.6 ppm	1.8999573E+05	1.9000051E+05	-5.299 ppm	11.0 ppm	PASS 42.06 %
1.0 MΩ	999879.7	999866.75	2.0 ppm	9.9986670E+05	9.9989270E+05	-12.955 ppm	11.0 ppm	PASS 99.65 %
1.9 MΩ	1899907	1899865.9	2.5 ppm	1.8997978E+06	1.9000162E+06	-21.654 ppm	55.0 ppm	PASS 37.66 %
10 MΩ	9998137	9997445.2	8.0 ppm	9.9975071E+06	9.9987669E+06	-69.192 ppm	55.0 ppm	FAIL 109.83 %
19 MΩ	18998461	18999845	16.0 ppm	1.8988468E+07	1.9008454E+07	72.874 ppm	510.0 ppm	PASS 13.85 %
100 MΩ	1.0000477E+08	1.0000867E+08	40.0 ppm	9.9949767E+07	1.0005977E+08	38.950 ppm	510.0 ppm	PASS 7.08 %

4W and 2W Zero test procedure for all test points that verify Zero offset of the OHMF function. 4-wire kelvin connection is used between DMM and MFC. 1GΩ resistance range is tested using the external standard, as MFC unable to provide this range value.

OHM ZERO 4W	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
10 Ω	Range 0.0000042 Ω	5.000e-05 Ω	-5e-05	5e-05	N/A	8.0000e-06 Ω	PASS
100 Ω	Range -0.0000425 Ω	5.500e-04 Ω	-0.00055	0.00055	N/A	2.2000e-06 Ω	PASS
1.0 kΩ	Range -0.0001255 Ω	5.500e-03 Ω	-0.0055	0.0055	N/A	2.2000e-06 Ω	PASS
10 kΩ	Range -0.0007947 Ω	5.500e-02 Ω	-0.055	0.055	N/A	2.2000e-06 Ω	PASS
100 kΩ	Range -0.0126352 Ω	5.500e-01 Ω	-0.55	0.55	N/A	2.2000e-06 Ω	PASS
1.0 MΩ	Range -0.6380376 Ω	5.500e+00 Ω	-5.5	5.5	N/A	2.2000e-06 Ω	PASS
10 MΩ	Range -5.8047627 Ω	5.500e+01 Ω	-55	55	N/A	2.2000e-06 Ω	PASS
100 MΩ	Range 1.5221380 Ω	5.500e+02 Ω	-550	550	N/A	2.2000e-06 Ω	PASS
1 GΩ	Range -6.2654542 Ω	5.500e+03 Ω	-5500	5500	N/A	2.2000e-06 Ω	PASS
OHM ZERO 2W	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
10 Ω	Range 0.4763256 Ω	5.500e-01 Ω	-0.55	0.55	N/A	8.0000e-06 Ω	PASS
100 Ω	Range 0.4754128 Ω	5.500e-01 Ω	-0.55	0.55	N/A	2.2000e-06 Ω	PASS
1.0 kΩ	Range 0.4748894 Ω	5.500e-01 Ω	-0.55	0.55	N/A	2.2000e-06 Ω	PASS
10 kΩ	Range 0.4850105 Ω	5.500e-01 Ω	-0.55	0.55	N/A	2.2000e-06 Ω	PASS
100 kΩ	Range 0.4852266 Ω	5.500e-01 Ω	-0.55	0.55	N/A	2.2000e-06 Ω	PASS
1.0 MΩ	Range 0.4839184 Ω	5.500e+00 Ω	-5.5	5.5	N/A	2.2000e-06 Ω	PASS
10 MΩ	Range 0.8605791 Ω	5.500e+01 Ω	-55	55	N/A	2.2000e-06 Ω	PASS
100 MΩ	Range -0.5307211 Ω	5.500e+02 Ω	-550	550	N/A	2.2000e-06 Ω	PASS
1 GΩ	Range -0.6127249 Ω	5.500e+03 Ω	-5500	5500	N/A	2.2000e-06 Ω	PASS

Procedure for all test points in the AC performance verification for ANAlog mode. AC-measurements does not suffer from TEMF offsets, test connection can be made using shielded leads terminated with dual banana plugs. MFC main AC output is used as reference source

ACV ANA Test	1V-10V	DUT	w/Guardband	Low Limit	Hi limit	Units	Measured	24h spec	Result
1.0 VAC @ 50.0 kHz	1.0	1.0000878	129.09	0.99955091	1.00044909	VAC	87.757 ppm	320.0 ppm	PASS 19.54 %
1.0 VAC @ 1.0 MHz	1.0	1.0111754	0.2500 %	0.9874	1.0126	VAC	1.1175 %	1.0100 %	PASS 88.69 %
10 VAC @ 10 Hz	10	9.9812981	2085	9.97805	10.02195	VAC	-1870.189 ppm	110.0 ppm	PASS 85.20 %
10 VAC @ 200 Hz	10	10.000503	73.18	9.9983682	10.0016318	VAC	50.278 ppm	90.0 ppm	PASS 30.81 %
10 VAC @ 500 Hz	10	10.000453	73.18	9.9983682	10.0016318	VAC	45.287 ppm	90.0 ppm	PASS 27.75 %
10 VAC @ 50.0 kHz	10	10.000193	129.09	9.9955091	10.0044909	VAC	19.332 ppm	320.0 ppm	PASS 4.30 %
10 VAC @ 1.0 MHz	10	10.093548	0.3000 %	9.869	10.131	VAC	0.9355 %	1.0100 %	PASS 71.41 %

Procedure for all test points in the AC performance verification for SYNCronous mode. This is highest AC accuracy test. AC-measurements does not suffer from TEMF offsets, test connection can be made using shielded leads terminated with dual banana plugs. MFC main AC output is used as reference source

ACV SYNC Test	DUT	w/Guardband	Low Limit	Hi limit	Measured	24h spec	Result, % spec
0.01 V AC+DC @ 10 Hz	0.010000992	372.33	0.009990	0.010010	99.164 ppm	600.0 ppm	PASS 10.20 %
0.01 V AC+DC @ 20 Hz	0.010000094	372.33	0.009990	0.010010	94.033 ppm	600.0 ppm	PASS 9.67 %
0.01 V AC+DC @ 40 Hz	0.010000721	372.33	0.009990	0.010010	72.078 ppm	600.0 ppm	PASS 7.41 %
0.01 V AC+DC @ 100 Hz	0.010000208	372.33	0.009993	0.010007	20.792 ppm	310.0 ppm	PASS 3.05 %
0.01 V AC+DC @ 1.0 kHz	0.010000513	372.33	0.009993	0.010007	51.293 ppm	310.0 ppm	PASS 7.52 %
0.01 V AC+DC @ 10.0 kHz	0.010002033	372.33	0.009992	0.010008	203.347 ppm	410.0 ppm	PASS 25.99 %
0.01 V AC+DC @ 20.0 kHz	0.010000997	372.33	0.009992	0.010008	99.675 ppm	410.0 ppm	PASS 12.74 %
0.01 V AC+DC @ 50.0 kHz	0.0099829042	0.0613 %	0.009983	0.010017	-0.1710 %	0.1110 %	PASS 99.24 %
0.01 V AC+DC @ 100.0 kHz	0.0099575463	0.1200 %	0.009937	0.010063	-0.4245 %	0.5110 %	PASS 67.28 %
0.01 V AC+DC @ 300.0 kHz	0.0098173112	0.1800 %	0.009580	0.010420	-1.8269 %	4.0200 %	PASS 43.50 %
0.01 V AC+DC @ 500.0 kHz	0.0097191237	0.2900 %	0.006766	0.013234	-2.8088 %	32.0500 %	PASS 8.69 %
0.01 V AC+DC @ 1.0 MHz	0.0091449539	0.4400 %	0.006751	0.013249	-8.5505 %	32.0500 %	PASS 26.32 %
0.1 V AC+DC @ 10 Hz	0.10000104	422.72	0.099947	0.100053	10.358 ppm	110.0 ppm	PASS 1.94 %
0.1 V AC+DC @ 20 Hz	0.099995054	206.36	0.099968	0.100032	-49.461 ppm	110.0 ppm	PASS 15.63 %
0.1 V AC+DC @ 40 Hz	0.099993854	206.36	0.099968	0.100032	-61.463 ppm	110.0 ppm	PASS 19.43 %
0.1 V AC+DC @ 100 Hz	0.099992575	121.36	0.099979	0.100021	-74.248 ppm	90.0 ppm	PASS 35.13 %
0.1 V AC+DC @ 1.0 kHz	0.099994851	121.36	0.099979	0.100021	-51.494 ppm	90.0 ppm	PASS 24.36 %
0.1 V AC+DC @ 10.0 kHz	0.099995796	121.36	0.099972	0.100028	-42.039 ppm	160.0 ppm	PASS 14.94 %
0.1 V AC+DC @ 20.0 kHz	0.099990933	121.36	0.099972	0.100028	-90.668 ppm	160.0 ppm	PASS 32.22 %
0.1 V AC+DC @ 50.0 kHz	0.099970247	345.45	0.099933	0.100067	-297.526 ppm	320.0 ppm	PASS 44.71 %
0.1 V AC+DC @ 100.0 kHz	0.099922147	886.36	0.099829	0.100171	-778.527 ppm	820.0 ppm	PASS 45.63 %
0.1 V AC+DC @ 300.0 kHz	0.099424918	0.1100 %	0.099580	0.100420	-0.5751 %	0.3100 %	FAIL 136.92 %
0.1 V AC+DC @ 500.0 kHz	0.099536134	0.1700 %	0.098820	0.101180	-0.4639 %	1.0100 %	PASS 39.31 %
0.1 V AC+DC @ 1.0 MHz	0.098419867	0.3500 %	0.098640	0.101360	-1.5801 %	1.0100 %	FAIL 116.19 %
1.0 V AC+DC @ 10 Hz	1.0000783	436.36	0.999454	1.000546	78.269 ppm	110.0 ppm	PASS 14.33 %
1.0 V AC+DC @ 20 Hz	1.0000135	141.36	0.999749	1.000251	13.464 ppm	110.0 ppm	PASS 5.36 %
1.0 V AC+DC @ 40 Hz	0.9999959	141.36	0.999749	1.000251	-4.100 ppm	110.0 ppm	PASS 1.63 %
1.0 V AC+DC @ 100 Hz	0.99998496	62.72	0.999847	1.000153	-15.040 ppm	90.0 ppm	PASS 9.85 %
1.0 V AC+DC @ 1.0 kHz	1.0000026	62.72	0.999847	1.000153	2.553 ppm	90.0 ppm	PASS 1.67 %
1.0 V AC+DC @ 10.0 kHz	0.99995857	62.72	0.999777	1.000223	-41.434 ppm	160.0 ppm	PASS 18.60 %
1.0 V AC+DC @ 20.0 kHz	0.99991463	62.72	0.999777	1.000223	-85.370 ppm	160.0 ppm	PASS 38.33 %
1.0 V AC+DC @ 50.0 kHz	0.99997721	129.09	0.999551	1.000449	-22.795 ppm	320.0 ppm	PASS 5.08 %
1.0 V AC+DC @ 100.0 kHz	1.0000231	266.36	0.998914	1.001086	23.124 ppm	820.0 ppm	PASS 2.13 %
1.0 V AC+DC @ 300.0 kHz	1.0010011	0.0468 %	0.996432	1.003568	0.1001 %	0.3100 %	PASS 28.06 %
1.0 V AC+DC @ 500.0 kHz	1.0025787	0.1200 %	0.988700	1.011300	0.2579 %	1.0100 %	PASS 22.82 %
1.0 V AC+DC @ 1.0 MHz	1.0053246	0.2500 %	0.987400	1.012600	0.5325 %	1.0100 %	PASS 42.26 %
10.0 V AC+DC @ 10 Hz	10.000809	403.63	9.993564	10.006436	80.937 ppm	240.0 ppm	PASS 12.58 %
10.0 V AC+DC @ 20 Hz	10.000215	141.36	9.996186	10.003814	21.505 ppm	240.0 ppm	PASS 5.64 %
10.0 V AC+DC @ 40 Hz	10.000039	141.36	9.996186	10.003814	3.868 ppm	240.0 ppm	PASS 1.01 %
10.0 V AC+DC @ 100 Hz	9.9999411	62.72	9.997173	10.002827	-5.892 ppm	220.0 ppm	PASS 2.08 %
10.0 V AC+DC @ 1.0 kHz	10.000095	62.72	9.997173	10.002827	9.535 ppm	220.0 ppm	PASS 3.37 %
10.0 V AC+DC @ 10.0 kHz	9.9994558	62.72	9.997173	10.002827	-54.419 ppm	220.0 ppm	PASS 19.25 %
10.0 V AC+DC @ 20.0 kHz	9.9992391	62.72	9.997173	10.002827	-76.092 ppm	220.0 ppm	PASS 26.91 %
10.0 V AC+DC @ 50.0 kHz	9.9986496	129.09	9.995009	10.004991	-135.039 ppm	370.0 ppm	PASS 27.06 %
10.0 V AC+DC @ 100.0 kHz	9.9963032	0.0248 %	9.985318	10.014682	-0.0370 %	0.1220 %	PASS 25.18 %
10.0 V AC+DC @ 300.0 kHz	9.9826126	0.0577 %	9.953227	10.046773	-0.1739 %	0.4100 %	PASS 37.17 %
10.0 V AC+DC @ 500.0 kHz	9.9915778	0.1400 %	9.835000	10.165000	-0.0842 %	1.5100 %	PASS 5.10 %
10.0 V AC+DC @ 1.0 MHz	10.041325	0.3000 %	9.819000	10.181000	0.4133 %	1.5100 %	PASS 22.83 %
100.0 V AC+DC @ 1.0 kHz	100.00088	65.0	99.951500	100.048500	8.772 ppm	420.0 ppm	PASS 1.80 %
100.0 V AC+DC @ 10.0 kHz	99.995495	65.0	99.931500	100.068500	-45.054 ppm	620.0 ppm	PASS 6.58 %
100.0 V AC+DC @ 20.0 kHz	99.981768	65.0	99.931500	100.068500	-182.318 ppm	620.0 ppm	PASS 26.62 %
100.0 V AC+DC @ 50.0 kHz	99.942167	0.0170 %	99.860998	100.139002	-0.0578 %	0.1220 %	PASS 41.61 %
100.0 V AC+DC @ 100.0 kHz	99.926534	0.0400 %	99.657997	100.342003	-0.0735 %	0.3020 %	PASS 21.48 %
700.0 V AC+DC @ 1.0 kHz	700.0498	78.64	699.650952	700.349048	71.147 ppm	420.0 ppm	PASS 14.03 %

Procedure for all test points that verify Gain of the DC current DCI function. Both +/-FS points are tested.

2-wire connection at LO and DCI is used between DMM and MFC.

DCI gain range points verify gain of the DC current function, using corrected 24-hour MFC output.

DCI Test	100nA-1A	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result
Zero µADC	0	-4.7676196E-10	71.82 ppm	0	0	Z-check	410 ppm	INFO
50 nADC	5E-08	4.9592379E-08	71.82 ppm	4.997591E-08	5.002409E-08	-8152.416 ppm	410 ppm	INFO
100 nADC	1E-07	9.958617E-08	71.82 ppm	9.995182E-08	1.000482E-07	-4138.298 ppm	410 ppm	FAIL 858.89 %
-100 nADC	-1E-07	-1.0048389E-07	71.82 ppm	-1.000482E-07	-9.995182E-08	4838.903 ppm	410 ppm	FAIL 1004.30 %
-50 nADC	-5E-08	-5.0474981E-08	71.82 ppm	-5.002409E-08	-4.997591E-08	9499.622 ppm	410 ppm	INFO
Zero µADC	0	-4.9760353E-10	71.82 ppm	0	0	Z-check	410 ppm	INFO
0.5 µADC	5E-07	4.9952603E-07	71.82 ppm	4.999391E-07	5.000609E-07	-947.949 ppm	50 ppm	FAIL 778.16 %
1.0 µADC	1E-06	9.9954878E-07	71.82 ppm	9.998782E-07	1.000122E-06	-451.220 ppm	50 ppm	FAIL 370.40 %
-1.0 µADC	-1E-06	-1.0004818E-06	71.82 ppm	-1.000122E-06	-9.998782E-07	481.759 ppm	50 ppm	FAIL 395.47 %
-0.5 µADC	-5E-07	-5.0039332E-07	71.82 ppm	-5.000609E-07	-4.999391E-07	786.631 ppm	50 ppm	FAIL 645.73 %
Zero 00 µADC	0	-4.1422766E-10	71.82 ppm	0	0	Z-check	410 ppm	INFO
5 µADC	5E-06	4.9995957E-06	71.82 ppm	4.999556E-06	5.000444E-06	-80.865 ppm	17 ppm	PASS 91.04 %
10 µADC	1E-05	9.9995807E-06	71.82 ppm	9.999112E-06	1.000089E-05	-41.927 ppm	17 ppm	PASS 47.20 %
-10 µADC	-1E-05	-1.0000354E-05	71.82 ppm	-1.000089E-05	-9.999112E-06	35.412 ppm	17 ppm	PASS 39.87 %
-5 µADC	-5E-06	-5.0003496E-06	71.82 ppm	-5.000444E-06	-4.999556E-06	69.915 ppm	17 ppm	PASS 78.71 %
Zero 000 µADC	0	-3.6671531E-10	71.82 ppm	0	0	Z-check	410 ppm	INFO
50 µADC	5E-05	4.9999457E-05	71.82 ppm	4.999561E-05	5.000439E-05	-10.865 ppm	16 ppm	PASS 12.37 %
100 µADC	0.0001	9.9999339E-05	71.82 ppm	9.999122E-05	0.0001000088	-6.606 ppm	16 ppm	PASS 7.52 %
-100 µADC	-0.0001	-0.0001	71.82 ppm	-0.0001000088	-9.999122E-05	0.028 ppm	16 ppm	PASS 0.03 %
-50 µADC	-5E-05	-5.0000267E-05	71.82 ppm	-5.000439E-05	-4.999561E-05	5.331 ppm	16 ppm	PASS 6.07 %
Zero mADC	0	-3.1826028E-10	33.64 ppm	0	0	Z-check	410 ppm	INFO
0.5 mADC	0.0005	0.00049999712	33.64 ppm	0.0004999762	0.0005000238	-5.754 ppm	14 ppm	PASS 12.08 %
1.0 mADC	0.001	0.0009999953	33.64 ppm	0.0009999524	0.001000048	-4.698 ppm	14 ppm	PASS 9.86 %
-1.0 mADC	-0.001	-0.00099999782	33.64 ppm	-0.001000048	-0.0009999524	-2.181 ppm	14 ppm	PASS 4.58 %
-0.5 mADC	-0.0005	-0.0004999991	33.64 ppm	-0.0005000238	-0.0004999762	-1.803 ppm	14 ppm	PASS 3.79 %
Zero 00 mADC	0	-2.7686894E-10	32.27 ppm	0	0	Z-check	410 ppm	INFO
5 mADC	0.005	0.0049999894	32.27 ppm	0.004999769	0.005000231	-2.112 ppm	14 ppm	PASS 4.56 %
10 mADC	0.01	0.0099999634	32.27 ppm	0.009999537	0.01000046	-3.659 ppm	14 ppm	PASS 7.91 %
-10 mADC	-0.01	-0.0099999478	32.27 ppm	-0.01000046	-0.009999537	-5.221 ppm	14 ppm	PASS 11.28 %
-5 mADC	-0.005	-0.0049999672	32.27 ppm	-0.005000231	-0.004999769	-6.561 ppm	14 ppm	PASS 14.18 %
Zero 000 mADC	0	-2.8164646E-10	53.32 ppm	0	0	Z-check	410 ppm	INFO
50 mADC	0.05	0.050000271	53.32 ppm	0.04999588	0.05000412	5.420 ppm	29 ppm	PASS 6.58 %
100 mADC	0.1	0.099999527	53.32 ppm	0.09999177	0.1000082	-4.729 ppm	29 ppm	PASS 5.74 %
-100 mADC	-0.1	-0.099999414	53.32 ppm	-0.1000082	-0.09999177	-5.859 ppm	29 ppm	PASS 7.12 %
-50 mADC	-0.05	-0.050000001	53.32 ppm	-0.05000412	-0.04999588	0.011 ppm	29 ppm	PASS 0.01 %
Zero ADC	0	-2.6830786E-10	115.22 ppm	0	0	Z-check	410 ppm	INFO
0.5 ADC	0.5	0.50000113	115.22 ppm	0.4998874	0.5001126	2.250 ppm	110 ppm	PASS 1.00 %
1.0 ADC	1	0.9999864	115.22 ppm	0.9997748	1.000225	-13.598 ppm	110 ppm	PASS 6.04 %
-1.0 ADC	-1	-1.0000029	115.22 ppm	-1.000225	-0.9997748	2.908 ppm	110 ppm	PASS 1.29 %
-0.5 ADC	-0.5	-0.49999863	115.22 ppm	-0.5001126	-0.4998874	-2.730 ppm	110 ppm	PASS 1.21 %

Procedure for all test points that verify Gain of the AC Current ACI function. Three frequency band points are tested, 50 Hz, 60 Hz and 1 kHz. 2-wire connection at LO and DCI is used between DMM and MFC.

ACI Test	200µA-2A	DUT	Source unc.	Low Limit	Hi limit	Measured	24h spec	Result, % spec
10 µA AC @ 50 Hz	1e-05	1.0015132E-05	0.0165 %	9.9893455e-06	1.00106545e-05	1513.237 ppm	0.0900 %	INFO
100 µA AC @ 50 Hz	0.0001	9.9984379E-05	0.0165 %	9.9893455e-05	0.000100106545	-156.212 ppm	0.0900 %	PASS 14.66 %
1.0 mA AC @ 50 Hz	0.001	0.00099997499	0.0165 %	0.00099903455	0.00100096545	-25.014 ppm	0.0800 %	PASS 2.59 %
10 mA AC @ 50 Hz	0.01	0.0099998373	0.0165 %	0.0099903455	0.0100096545	-16.272 ppm	0.0800 %	PASS 1.69 %
100 mA AC @ 50 Hz	0.1	0.10000257	0.0138 %	0.099906182	0.100093818	25.706 ppm	0.0800 %	PASS 2.74 %
1.0 A AC @ 50 Hz	1.0	0.99997752	0.0138 %	0.99886182	1.00113818	-0.0022 %	0.1000 %	PASS 1.98 %
10 µA AC @ 60 Hz	1e-05	1.0023788E-05	0.0138 %	9.9896182e-06	1.00103818e-05	2378.815 ppm	0.0900 %	INFO
100 µA AC @ 60 Hz	0.0001	9.9975248E-05	0.0138 %	9.9896182e-05	0.000100103818	-247.520 ppm	0.0900 %	PASS 23.84 %
1.0 mA AC @ 60 Hz	0.001	0.00099998506	0.0134 %	0.00099906636	0.00100093364	-14.941 ppm	0.0800 %	PASS 1.60 %
10 mA AC @ 60 Hz	0.01	0.010000092	0.0134 %	0.0099906636	0.0100093364	9.176 ppm	0.0800 %	PASS 0.98 %
100 mA AC @ 60 Hz	0.1	0.10000541	0.0308 %	0.099889182	0.100110818	54.057 ppm	0.0800 %	PASS 4.88 %
1.0 A AC @ 60 Hz	1.0	1.0000072	0.0308 %	0.99869182	1.00130818	0.0007 %	0.1000 %	PASS 0.55 %
10 µA AC @ 1.0 kHz	1e-05	1.0012834E-05	0.0165 %	9.9893455e-06	1.00106545e-05	1283.437 ppm	0.0900 %	INFO
100 µA AC @ 1.0 kHz	0.0001	9.9969459E-05	0.0165 %	9.9893455e-05	0.000100106545	-305.409 ppm	0.0900 %	PASS 28.66 %
1.0 mA AC @ 1.0 kHz	0.001	0.0010000554	0.0165 %	0.00099933455	0.00100066545	55.400 ppm	0.0500 %	PASS 8.33 %
10 mA AC @ 1.0 kHz	0.01	0.010000611	0.0165 %	0.0099933455	0.0100066545	61.124 ppm	0.0500 %	PASS 9.19 %
100 mA AC @ 1.0 kHz	0.1	0.10001144	0.0138 %	0.099936182	0.100063818	114.353 ppm	0.0500 %	PASS 17.92 %
1.0 A AC @ 1.0 kHz	1.0	1.0001534	0.0138 %	0.99866182	1.00133818	0.0153 %	0.1200 %	PASS 11.46 %

Test date	25 June 2018 12:42
UUT Internal TEMP?	1
Destructive overloads?	1

Lab temperature maintained +24°C ±2°C

Internal use only

Not validated