



Certificate Of Calibration

Certificate No: 3458AMY59350622

Manufacturer: Keysight Technologies

Description: 8.5 Digit Multimeter

Model No: 3458A

Serial No: MY59350622

Options Installed With Specifications: N/A

Date of Calibration: 15-DEC-2019

Humidity: (20 to 80)% RH

Temperature: (23 ± 2)°C

Procedure: 3458FINAL/229530179072

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures in compliance with a quality management system registered to ISO 9001:2015.

As Received Conditions: Factory tested. No incoming data available.

Action Taken: No corrective actions were necessary.

As Shipped Conditions: At the completion of the calibration, measured values were IN SPECIFICATION at the points tested.

Remarks or special requirements:

Notes:


- 1) This calibration report shall not be reproduced, except in full.

Traceability Information: Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Calibration Equipment Used:

Model Number	Model Description	<i>Date Used: Date equipment used in this Calibration</i>		
		Equipment ID	Date Used	Cal Due Date
5730A	Fluke Calibrator	4527502	15-DEC-2019	10-FEB-2020
5725A	Fluke Amplifier	2785003	15-DEC-2019	10-FEB-2020
3458A	3458A Reference Std	MY45041545	15-DEC-2019	10-FEB-2020

Print Date: 15-DEC-2019



Tay Eng Su
Quality Manager

Keysight Technologies

	DD	MM	YY	BY:
CAL	15	12	19	S.Y
DUE				