

# Keysight Technologies

## Understanding Measurement Risk

White Paper

### 1 Abstract

One key reason for performing a calibration is to assess a device as either in- or out-of-tolerance. Common calibration test scenarios compare a device parameter against that of a measurement standard by way of a measurement process. If the difference between the device parameter and the measurement standard is greater than the specified tolerance, the device is deemed out-of-tolerance. However, errors in the measurement process bring about the possibility of an incorrect assessment. An incorrect assessment may result in devices incorrectly declared as in-tolerance (false-accept) or incorrectly declared as out-of-tolerance (false-reject).

The risk of making an incorrect in- or out-of-tolerance assessment can be determined by evaluating probability density functions that incorporate a device's parameter population and the measurement error. This paper provides an intuitive explanation of these probability density functions drawing on Monte Carlo simulation to demonstrate the relationship between a device's true value and the corresponding measured value.

## 2.0 Introduction

In manufacturing facilities throughout the world, test engineers design measurement procedures for manufacturing purposes. It is common for test engineers to rely on the specifications of measuring equipment to assess the accuracy of the measurement procedures. This creates a dependency between the measuring equipment specifications and the quality of the manufacturing process. To maintain manufacturing process quality, the measuring equipment requires periodic calibration.

For the above scenario, one of the primary purposes of calibration is to verify that the measuring equipment performs at a level consistent with the equipment's specifications. In other words, is the measuring equipment in- or out-of-tolerance?

Frequently, calibration involves comparing a device parameter (that is, a parameter of the measuring equipment) against that of a measurement standard. For example, assume we wish to calibrate an RF power source with a power meter. The purpose of the calibration is to assess the RF power source error (the difference between the indicated power and the true power supplied by the source) and determine if it is less than a specified tolerance. If it were possible to use a perfect power meter and a perfect measuring procedure, determining the RF power source's error is simply a matter of noting the difference between the power meter's reading and the indicated value of the RF power source. However, since a real-world power meter is not perfect, knowing the exact RF power source error is not possible. Our lack of knowledge about the exact error is what gives rise to the possibility of declaring a device as in-tolerance when it is actually out-of-tolerance (false-accept) or, declaring a device as out-of-tolerance when it is actually in-tolerance (false-reject).

Michael Dobbert, Keysight Technologies

### 3 Device Error

From our example, the RF power source error is the value that we wish to compare against the tolerance limit. The RF power source displays the power level it purports to output. The RF power source error is the difference between the purported output power level and the actual true value [7] of the power. Expressed mathematically,

$$e_{dut} = n_{dut} - t \quad \text{Equation (1)}$$

where:

$$\begin{aligned} e_{dut} &= \text{RF power source error, which is the device under test error} \\ n_{dut} &= \text{nominal, or displayed RF power level} \\ t &= \text{true power level} \end{aligned}$$

To calibrate the RF power source, the power meter measures the true output power from the RF power source. However, the measured power indicated by the power meter differs from the true value because of measurement error. That is,

$$i_{std} = t + e_{std} \quad \text{Equation (2)}$$

where:

$$\begin{aligned} i_{std} &= \text{power indicated by the power meter} \\ t &= \text{true power level} \\ e_{std} &= \text{power meter measurement error} \end{aligned}$$

Combining Equations (1) and (2) yields:

$$n_{dut} - i_{std} = e_{dut} - e_{std} \quad \text{Equation (3)}$$

The left hand side of Equation (3) represents the observed calibration result. That is, the observed calibration result is the difference between the device under test, as displayed by our RF power source, and the calibration standard as indicated by our power meter. Representing the observed calibration result as  $y$ , we can rewrite Equation (3) as

$$y = e_{dut} - e_{std} \quad \text{Equation (4)}$$

where:

$$y = \text{observed calibration result}$$

Equation (4) relates the actual device under test error,  $e_{dut}$ , and the measurement error,  $e_{std}$ , to the observed calibration result,  $y$ . A device is in-tolerance if  $e_{dut}$  is less than the tolerance limit. That is, for a tolerance limit,  $L$ , a device is in-tolerance if:

$$-L \leq e_{dut} \leq L$$

At the time of calibration, however, the precise value of  $e_{std}$  is unknown. Consequently, it is not possible to determine the precise value of  $e_{dut}$ . What is known is the observed calibration result,  $y$ . Fortunately, if  $e_{std}$  is small relative to  $e_{dut}$ , the observed calibration result,  $y$ , is a reasonable approximation to  $e_{dut}$ . Therefore, a device is declared in-tolerance if:

$$-L \leq y \leq L$$

## 4 Random Variables, Tolerance Testing and Risk

Because at the time of calibration the precise value of the device under test error and the measurement error are unknown, it is necessary to represent  $e_{dut}$  and  $e_{std}$  as statistical quantities. Particularly,  $e_{dut}$  and  $e_{std}$  are random variables each characterized by a distribution. Assuming that the device under test belongs to a population, the distribution for  $e_{dut}$  represents the different error values for each device in that population. The distribution for the measurement error,  $e_{std}$  is the measurement uncertainty of the measurement process.

Tolerance testing involves using the observed result of the calibration to assess a device as either in- or out-of-tolerance. However, the actual in- or out-of-tolerance status of a device depends upon the actual value of the device under test error. Therefore, there is a chance that the observed result of the calibration leads to an incorrect assessment of the device as being either in- or out-of-tolerance. Risk is the probability of making an incorrect assessment.

## 5 Monte Carlo Simulation

We can use Monte Carlo methods to simulate the result of calibrating devices using calibration standards that include errors. As such, Monte Carlo simulation provides a way to visualize the relationship between the device under test error,  $e_{dut}$  and the calibration result,  $y$ , when treating the variables in Equation (4) as random variables. Monte Carlo simulation is also useful for identifying incorrect in- or out-of-tolerance assessments as well as determining the risk of making an incorrect assessment.

The Monte Carlo simulation operates as follows. Assuming a distribution for the population to which our device under test belongs, we can draw from it a large number of random samples. Each sample represents a value of the error,  $e_{dut}$  for a device under test. For each of these random samples, we can draw another random sample from a distribution representing the measurement error,  $e_{std}$  and subtract it from the sample for the device under test error to give a value for the calibration result,  $y$ .




Figure 1. Simulation of observed calibration results versus the device under test error.

The benefit of a Monte Carlo simulation is that values for  $e_{dut}$  and the corresponding values for  $y$  exist, unlike real-world calibrations where  $e_{dut}$  is unknowable. Figure 1 shows values for  $y$  plotted against the values for  $e_{dut}$  and illustrates graphically that  $y$  and  $e_{dut}$  are highly correlated. This is necessary in order to approximate a value for  $e_{dut}$  using a value for  $y$ .

## 5 Monte Carlo Simulation (continued)

Adding tolerance limits to the plot of  $y$  values versus  $e_{dut}$  values defines several different regions.




Figure 2. Tolerance limits defining false-accept and false reject regions.

The vertical lines in Figure 2 represent two-sided symmetrical tolerance limits about the device under test error,  $e_{dut}$ . Points between the vertical lines represent in-tolerance devices; points outside represent out-of-tolerance devices. The horizontal lines represent two-sided symmetrical tolerance limits about the calibration result,  $y$ . Points between these lines represent devices *observed* as in-tolerance while points outside represent devices *observed* as out-of-tolerance.

Points in the center region of Figure 2 represent devices that are both truly in-tolerance ( $-L \leq e_{dut} \leq L$ ) as well as observed in-tolerance ( $-L \leq y \leq L$ ). Similarly, points in the outer corners of Figure 2 represent devices that are both truly out-of-tolerance and observed out-of-tolerance. For these two cases, the observed calibration result leads to a correct assessment of a device's actual in- or out-of-tolerance status.

Figure 2 also shows points representing devices observed in-tolerance when, in fact, the device is actually out-of-tolerance. These points lie in the regions labeled *false-accept*. Additionally, Figure 2 shows points representing devices observed as out-of-tolerance when the device's true status is in-tolerance. These points lie in the regions labeled *false-reject*. Points in the false-accept and false-reject regions represent devices for which the observed calibration result leads to an incorrect in- or out-of-tolerance assessment.

Risk is the probability of a point lying in either the false-accept regions or the false-reject regions. With our Monte Carlo simulation, determining risk involves counting the number of points in a particular region and dividing the count by the number of points in the population, or a subset of the population. The divisor, in this case, depends upon the type of risk.

## 6 Types of False-Accept Risk

Different types of false-accept risk exist depending on various views individuals take regarding risk. This section explores three different view cases.

### 6.1 Case A

Case A defines risk as the unconditional probability of observing a device parameter as intolerance when it is actually out-of-tolerance. Case A risk is a quality metric for a calibration laboratory. It describes the likelihood of any given device leaving the calibration laboratory as out-of-tolerance.

Considering our Monte Carlo simulation, Case A risk is the probability of randomly choosing a point in the false-accept regions from the entire population. Determining this risk is a matter of counting the points in the false-accept regions and dividing by the total number of points in the population.

### 6.2 Case B

Case B defines risk as the conditional probability that a device parameter is out-of-tolerance *given* an in-tolerance observation. Calibration laboratories only put observed in-tolerance devices into service. Case B describes the risk of an observed in-tolerance device that is unknowingly out-of-tolerance. For calibration laboratory management, this risk defines the beginning-of-period measurement reliability for managed calibration equipment.

Determining this risk from our Monte Carlo simulation entails counting the points in the false-accept regions and dividing by the number of points where the calibration result is within the tolerance limits ( $-L \leq y \leq L$ ).

### 6.3 Case C

Case C defines risk as the conditional probability that a device parameter is out-of-tolerance *given* a particular observed calibration result. Since a calibration result is particular to a given device, Case C defines risk specific to that device.

For our Monte Carlo simulation, the region of interest is very narrow horizontal slice centered on the given calibration result. Risk, in this case, is the number of out-of-tolerance points divided by the total number of points in this narrow slice.

## 7 Joint Probability Density Function

Monte Carlo simulation is useful for visualizing how risk is determined, but a more practical method of determining risk is necessary. This more practical method is available using a joint probability density function appropriate for our calibration scenario. As stated earlier, we wish to treat the variables in Equation (4) as random variables that follow a distribution. We could choose from any number of possible distributions, but based upon the Principle of Maximum Entropy [6], our Monte Carlo simulation assumed that the device population and the measurement error follow Gaussian distributions<sup>1</sup>.

The probability density function assumed for the device population is

$$p_0(e_{dut}) = \frac{1}{\sqrt{2\pi}\sigma_0} e^{-\frac{(e_{dut})^2}{2\sigma_0^2}} \quad \text{Equation (5)}$$

where:

$\sigma_0$  = standard deviation of the population distribution

Recalling that the calibration result is the difference between the device under test error and the measurement error, the probability density function assumed for measurement error is a Gaussian distribution centered about a value for  $e_{dut}$ . The probability, as a function of  $y$ , given a value for  $e_{dut}$  is,

$$p_m(y - e_{dut}) = \frac{1}{\sqrt{2\pi}\sigma_m} e^{-\frac{(y - e_{dut})^2}{2\sigma_m^2}} \quad \text{Equation (6)}$$

where:

$\sigma_m$  = standard deviation of the measurement error (standard uncertainty)

1. With better knowledge about the population or the measurement uncertainty, we could choose different distributions or retain the Gaussian distributions, but choose those distributions whose mean value is non-zero.



## 7 Joint Probability Density Function (continued)




Figure 3. Distribution assumptions for device under test error and measurement error.

Figure 3 illustrates how the device population distribution and the measurement error distribution relate to our Monte Carlo simulation. (It is necessary, however, to imagine the measurement error distribution sliding diagonally across the plot for various values of  $e_{dut}$ .) Given the Gaussian assumptions, most points occur near the center of the plot as indicated by the point coloring. The point coloring indicates density with the light coloring representing high point density; dark coloring, low point density<sup>2</sup>.

The product of Equations (5) and (6) is the joint probability density function, which defines probability density over a two-dimensional surface area.

$$p(e_{dut}, y) = p_0(e_{dut}) p_m(y - e_{dut}) = \frac{1}{\sqrt{2\pi}\sigma_0} e^{-\frac{(e_{dut})^2}{2\sigma_0^2}} \cdot \frac{1}{\sqrt{2\pi}\sigma_m} e^{-\frac{(y - e_{dut})^2}{2\sigma_m^2}} \quad \text{Equation (7)}$$

The total probability for a given two-dimensional rectangular area or region is found by integrating the joint probability density function over the region. That is, the probability for a given region is,

$$P_R = \iint_R p_0(e_{dut}) p_m(y - e_{dut}) dA$$

where  $R$  defines a particular region on our calibration result versus device under test error plot. We can determine false-accept risk by integrating the joint probability density function over the false-accept regions of our plot.

2. For the point coloring, the method to determine density was to draw a grid, count the number of points in each grid square and divide by the grid square area.

## 8 Determining False-Accept Risk (Case A)

For two-sided tolerance limits, two false-accept regions exist (see Figure 4). Case A false-accept risk is found by integrating Equation (7) over the false-accept regions indicated by the dashed arrows in Figure 4.




Figure 4. False-accept regions for Case A risk.

Specifically,

$$P_{FA^+} = \iint_{FA^+} p_0(e_{dut}) p_m(y - e_{dut}) dA$$

where  $FA^+$  is the region:

$$L \leq e_{dut} \leq \infty \text{ and } -L \leq y \leq L$$

and

$$P_{FA^-} = \iint_{FA^-} p_0(e_{dut}) p_m(y - e_{dut}) dA$$

where  $FA^-$  is the region:

$$-\infty \leq e_{dut} \leq -L \text{ and } -L \leq y \leq L$$

The total risk is the sum of the probability for each of the two false-accept risk regions.

$$P_{Case A} = P_{FA^+} + P_{FA^-} \quad \text{Equation (8)}$$

## 9 Determining False-Accept Risk (Case B)

For Case A risk, integrating Equation (7) provides the unconditional probability for a given region. Case B risk, in contrast, is a conditional probability of false-accept. That is, rather than the probability relative to the entire population, Case B risk is the probability of false-accept given an observed in-tolerance calibration result ( $-L \leq y \leq L$ ). Conditional probability, in this case, is the (unconditional) probability of a region divided by the (unconditional) probability for a region representing the *condition*. Case B risk is the unconditional probability of false-accept divided by the unconditional probability of an observed in-tolerance calibration result.




Figure 5. Observed in-tolerance region for Case B risk.

As illustrated in Figure 5, the probability of an observed in-tolerance calibration result is,

$$P_{InTol} = \iint_{InTol} p_0(e_{dut}) p_m(y - e_{dut}) dA \quad \text{Equation (9)}$$

where *InTol* is the region:

$$-\infty \leq e_{dut} \leq \infty \text{ and } -L \leq y \leq L$$

The unconditional false-accept risk is given by PCase A. Case B risk is,

$$P_{Case B} = \frac{P_{Case A}}{P_{InTol}} \quad \text{Equation (10)}$$

## 10 Determining False-Accept Risk (Case C)

Case C risk is the conditional probability of false-accept given an observed calibration result. It is useful to think of Case C as similar to Case B. Case B only involved values in the region of  $(-L \leq y \leq L)$ . This region contains the false-accepts regions and defines the region applied to Equation (9). For Case C risk, this region is infinitely narrow about the observed calibration result. Figure 6 represents this infinitely narrow region as a line cutting horizontally through our calibration result versus device under test error plot.




Figure 6. False-accept regions for Case C risk.

Case C risk is the probability of false-accept, now defined as infinitely narrow horizontal regions positioned at the observed calibration result, divided by the probability of obtaining the observed calibration result. Integrating Equation (7) gives all of these probabilities. Once again, two false-accept regions exist. A region exists above the tolerance limit ( $L \leq e_{dut} \leq \infty$ ) and below the tolerance limit ( $-\infty \leq e_{dut} \leq -L$ ). Specifically for Case C, a known calibration result exists; therefore,  $y$  is constant.

Integrating Equation (7) over the regions indicated by the dashed arrows in Figure 6, assuming a constant value for  $y$ , gives,

$$P'_{FA^+} = \int_L^{\infty} p_0(e_{dut}) p_m(y - e_{dut}) de_{dut}$$

$$P'_{FA^-} = \int_{-\infty}^{-L} p_0(e_{dut}) p_m(y - e_{dut}) de_{dut}$$

$$P_y = \int_{-\infty}^{\infty} p_0(e_{dut}) p_m(y - e_{dut}) de_{dut}$$

For Case C, risk is a function of the calibration result.

$$P(y)_{Case\ C} = \frac{P'_{FA^+} + P'_{FA^-}}{P_y} \quad \text{Equation (11)}$$

## 11 Reducing False-Accept Risk

Using a device that is unknowingly out-of-tolerance has financial consequences. Therefore, minimizing false-accept risk (as well as false-reject risk) is an objective for metrologists.




Figure 7. Reducing false-accept risk.

Figure 7 illustrates three methods for reducing false-accept risk. Each method involves reducing the number of points in the false-accept regions. The upper left corner of Figure 7 represents the reference case. Each of the three remaining quadrants represents reduced false-accept probability from the reference.

## 11 Reducing False-Accept Risk (continued)

### 11.1 Reduce $\sigma_m$

Reducing the standard deviation of  $e_{std}$  is to reduce the measurement error. As measurement error lessens, the correlation between the calibration result,  $y$ , and the device under test error,  $e_{dut}$ , improves, making  $y$  a better approximation to  $e_{dut}$ . Likewise, reducing  $\sigma_m$  improves the Test Uncertainty Ratio (TUR) [8].

Represented graphically in the upper right corner of Figure 7, as measurement error lessens, calibration results in the false-accept region move vertically to be out-of-tolerance. As well, calibration results in the false-reject region move to be in-tolerance. The net result is fewer false-accept and false-reject points.

### 11.2 Reduce $\sigma_0$

Reducing the standard deviation of  $e_{dut}$  is to reduce the spread of the device population distribution. This is possible for a calibration laboratory that manages an inventory of devices. If we assume a device experiences random drift (or other mechanisms that cause a device parameter value to change semi-permanently), then the distribution representing the device population tends to spread over time. Managing an inventory of devices with a calibration process that adjusts (or repairs) devices limits the spread of the population<sup>3</sup>. It is possible, therefore, to control  $\sigma_0$  by selecting appropriate adjustment limits or by adjusting calibration intervals.

The lower left corner of Figure 7 illustrates the effect of reducing the spread of  $e_{dut}$ . As  $e_{dut}$  narrows, false-accept points move horizontally to be in-tolerance, while false-reject points move to be out-of-tolerance. Once again, the net result is fewer false-accept and false-reject points.

3. *Managing devices in this way is generally part of a program to set and manage measurement reliability targets [8] [9].*

## 11 Reducing False-Accept Risk (continued)

### 11.3 Guard-band

Instances arise where reducing false-accept risk is necessary and yet, reducing or further reducing  $e_{std}$  and  $e_{dut}$  is not practical. Applying a guard-band can reduce false-accept risk, but at the expense of increased false-reject risk.

Guard-banding is to create test limits for the calibration result more restrictive than the tolerance limits. The lower right corner of Figure 7 illustrates this situation where the test limits (indicated as  $TL$  and  $-TL$ ) are narrower than the tolerance limits.

Guard-banding changes the in- or out-of-tolerance decision point. As test limits becomes more restrictive, more points once considered in-tolerance are now considered out-of-tolerance. The result is fewer false-accept points but more false-reject points.

It is possible set test limits with sufficient guard-band to achieve desired maximum allowable levels of false-accept risk. Because the test limit is different from the tolerance limit, our risk calculations require integration over a modified region. Namely, the unconditional probability for the false-accept region above the tolerance limit is, (as a function of the test limit,  $TL$ ),

$$P(TL) = \iint_{GBR} p_0(e_{dut}) p_m(y - e_{dut}) dA \quad \text{Equation (12)}$$

where  $GBR$  is the guard-banded region,

$$L \leq e_{dut} \leq \infty \text{ and } -TL \leq y \leq TL$$

For Case A and B risk, it is necessary to find a value for  $TL$  such that the false-accept risk is no greater than the maximum allowable risk.

$$P(TL)_{FA^+} + P(TL)_{FA^-} \leq Risk_{max} \quad \text{(Case A)}$$

$$\frac{P(TL)_{FA^+} + P(TL)_{FA^-}}{P(TL)_{InTol}} \leq Risk_{max} \quad \text{(Case B)}$$

The guard-band is the difference between the test limit that results in the maximum allowable risk and the tolerance.

For Case C risk, it is necessary to find the calibration result for which the false-accept risk is no greater than the maximum allowable risk.

$$P(y)_{Case C} \leq Risk_{max} \quad \text{(Case C)}$$

For Case C, the difference between the tolerance limit and the calibration result that gives the maximum allowable risk is the guard-band.

## 12 Conclusion

Monte Carlo simulation is very useful for visualizing and understanding the subject of measurement risk. This paper presented three similar views of risk. Each depends upon two independent a priori probability distributions: the device under test population and the measurement error distribution. The device under test population, which, by applying the Principle of Maximum Entropy (PME), we determine using equation (5). Referring to the methods of the GUM [10], accounting for all known offset error and continuing with the PME, we determine the measurement error distribution using equation (6). Relying on the GUM, the PME and the joint probability density function provides a very good methodology for determining risk that is appropriate for the purposes of publication and regulation.

For other applications, in addition to publication and regulation, our methodology requires careful understanding of underlying assumptions. For instance, as shown in [5], the computed risk is sensitive to offset of either the device under test population or the measurement error distribution, which we have assumed to be negligible. Nevertheless, the method for determining risk using a joint probability density function is rigorous. The calculated risk value, however, is only as good as the assumptions for both distributions. If desired, we can construct joint probability density functions using distributions that more accurately reflect our knowledge of the device under test population and the measurement error.

Furthermore, we may wish to consider the following:

- Offset error, at times, is unavoidable. This is especially true for devices that operate over a range of settings or conditions. It is simply not possible to adjust all parameters for all conditions to nominal simultaneously. Equations (5) and (6) are easily adjusted to account for offset.
- Observed calibration history is useful for estimating device under test distributions. However, care is necessary to ensure the observed data is homogeneous, as indicated in [9]. Several factors that lead to observed multi-mode (i.e., non-homogeneous) distributions relate to device manufacture/model/option, device calibration interval, test station at which the device is tested, as well as functioning devices versus nonfunctioning (broken) devices.
- The random variables  $e_{dut}$  and  $e_{std}$  from equation (4) are not static over time and therefore, sampling issues need consideration (see [1]).
- We can extend the joint probability density function to included a weighting function allowing the determination of cost associated with false-accept.

$$Cost_R = \iint_R p_0(e_{dut}) p_m(y - e_{dut}) w(e_{dut}, y) dA$$

We can choose a weighting function that assigns higher cost values to devices significantly out-of-tolerance compared to devices just out-of-tolerance. Such a costing model requires a clear understanding of the device under test distribution, the measurement error and the weighting function.



### 13 Example

Computing risk depends on the tolerance limits, a value for the measurement error standard uncertainty,  $\sigma_m$ , and an estimate of the device under test population standard deviation,  $\sigma_\theta$ . For the purposes of the following example, an estimate<sup>4</sup> for  $\sigma_\theta$  is,

$$\sigma_\theta = \frac{L}{F^{-1}\left(\frac{1+p}{2}\right)} \quad \text{Equation (13)}$$

where  $F^{-1}$  is the inverse normal function, and  $p$  is the in-tolerance probability which is based on the observed history for a population of devices.

In this example, the tolerance limits for the output power accuracy of an RF power source are  $\pm 0.9$  dB. A power meter with a 95% expanded uncertainty of  $\pm 0.274$  dB is used to calibrate the RF power source. From the calibration history, we will assume that the in-tolerance probability is 80%. We wish to establish guard-bands to limit the false-accept risk to 2%.

From the above,

$$L = 0.9$$

$$\sigma_m = \frac{0.274}{1.96} = 0.14$$

$$p = 0.8$$

$$\sigma_\theta = \frac{0.9}{F^{-1}\left(\frac{1+0.8}{2}\right)} = 0.7$$

With values for  $L$ ,  $\sigma_\theta$ ,  $\sigma_m$ , we can use Equations (8) and (10) to determine Case A and B false-accept risk.

$$P_{FA^+} = \int_{0.9}^{\infty} \int_{-0.9}^{0.9} p(e_{dut}) p(y - e_{dut}) dy de_{dut}$$

$$P_{FA^-} = \int_{-\infty}^{-0.9} \int_{-0.9}^{0.9} p(e_{dut}) p(y - e_{dut}) dy de_{dut}$$

$$P_{Case A} = P_{FA^+} + P_{FA^-} = 2.370\%$$

$$P_{InTol} = \int_{-\infty}^{\infty} \int_{-0.9}^{0.9} p(e_{dut}) p(y - e_{dut}) dy de_{dut}$$

$$P_{Case B} = \frac{P_{Case A}}{P_{InTol}} = 2.996\%$$

4. The use of Equation (13) is by convention. Case C risk and guard-band can be computed using the software tool RiskGuard [2] [3] provided by Integrated Sciences Group, which estimates  $\sigma_\theta$  using Equation (13). Similar methods for estimating  $\sigma_\theta$  are found in [4].

### 13 Example (continued)

This example does not provide a calibration result. For that reason, Case C risk is not determined.

Both the Case A and Case B false-accept risks are greater than the desired 2%. The goal is to find a test limit, employing Equation (12), that achieves the desired false-accept risk. Doing so is an iterative process for which a numerical solver<sup>5</sup> is most efficient.

$$TL_{Case\ A} = 0.881\ dB$$

$$TL_{Case\ B} = 0.853\ dB$$

The guard-band in these cases is the difference between the test limit and the tolerance limit.

$$GB_{Case\ A} = 0.9 - 0.881 = 0.019\ dB$$

$$GB_{Case\ B} = 0.9 - 0.853 = 0.047\ dB$$

To determine Case C guard-band, it is necessary to find the calibration result,  $\gamma$ , using Equation (11), that achieves the desired false-accept risk. Doing so is once again an iterative process best executed using a numerical solver.

$$\gamma_{Case\ C} = 0.643\ dB$$

$$GB_{Case\ C} = 0.9 - 0.643 = 0.257\ dB$$

Table 1 summarizes the example.

Table 1 Example Summary.

	Case A	Case B	Case C
	Unconditional Probability of False-Accept	Probability of False-Accept Given Observed In-Tolerance	Probability of False-Accept Given a Measured Value
Probability of False-Accept	2.370 %	2.996 %	NA
2% Risk Guard-band	0.019 dB	0.047 dB	0.257 dB

5. For this example, the `fsolve` function from the MATLAB® Optimization Toolbox was used.

## 14 References

- [1] Brown, R., *Using a Manufacturer's Specification as a Type B Error Contribution*, NCSL International Workshop and Symposium, Nashville, TN, August, 2006
- [2] Castrup, H., *Estimating Guardbands Commensurate with Allowable Risk*, August 18, 2000
- [3] Castrup, H., *Analytical Metrology SPC Methods*, Proc. 1991 NCSL Workshop and Symposium, August 1991, Albuquerque, NM
- [4] Jackson, D., *Measurement Risk Analysis Methods*, Measurement Science Conference, 2005
- [5] Nicholas, R., *The Significance on Consumer and Producer Risk for Uncorrected Product and Measurement Biases*, NCSL International Workshop and Symposium, Nashville, TN, August, 2006
- [6] Wöger, W., *Probability assignment to systemic deviations by the Principle of Maximum Entropy*, IEEE Transactions on Instrumentation and Measurement., June 1987, IM-36, pg. 655-658
- [7] *International Vocabulary of Basic and General Terms in Metrology*, Geneva, International Organization for Standardization, 1993
- [8] ANSI/NCSL Z540.3-2006, *Requirements for the Calibration of Measuring and Test Equipment*, National Conference of Standard Laboratories International, Bolder, CO, 2006
- [9] NCSL RP-1, *Establishment and Adjustment of Calibration Intervals*, National Conference of Standard Laboratories International, Bolder, CO, 1996
- [10] *Guide to the Expression of Uncertainty in Measurement (GUM)*, BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, OIML – ISO, Geneva, 1995



myKeysight

[www.keysight.com/find/mykeysight](http://www.keysight.com/find/mykeysight)

A personalized view into the information most relevant to you.

For more information on Keysight Technologies' products, applications or services, please contact your local Keysight office. The complete list is available at: [www.keysight.com/find/contactus](http://www.keysight.com/find/contactus)

#### Americas

Canada	(877) 894 4414
Brazil	55 11 3351 7010
Mexico	001 800 254 2440
United States	(800) 829 4444

#### Asia Pacific

Australia	1 800 629 485
China	800 810 0189
Hong Kong	800 938 693
India	1 800 112 929
Japan	0120 (421) 345
Korea	080 769 0800
Malaysia	1 800 888 848
Singapore	1 800 375 8100
Taiwan	0800 047 866
Other AP Countries	(65) 6375 8100

#### Europe & Middle East

Austria	0800 001122
Belgium	0800 58580
Finland	0800 523252
France	0805 980333
Germany	0800 6270999
Ireland	1800 832700
Israel	1 809 343051
Italy	800 599100
Luxembourg	+32 800 58580
Netherlands	0800 0233200
Russia	8800 5009286
Spain	0800 000154
Sweden	0200 882255
Switzerland	0800 805353
	Opt. 1 (DE)
	Opt. 2 (FR)
	Opt. 3 (IT)
United Kingdom	0800 0260637

For other unlisted countries:  
[www.keysight.com/find/contactus](http://www.keysight.com/find/contactus)

(BP-05-19-14)