



Keysight Technologies Malaysia Sdn Bhd (463532-M)  
 (Formerly known as Agilent Technologies Microwave  
 Products (M) Sdn Bhd)  
 Bayan Lepas Free Industrial Zone  
 11900 Penang, Malaysia  
 Keysight Approved Calibration Provider #25865



## Certificate of Calibration

**Certificate No:** 4109227-4893557-1

**Manufacturer:** Keysight Technologies

**Description:** Electrometer/High Resistance Meter, 0.01fA,  
1000V

**Model No:** B2987A

**Serial No:** MY54320125

**Options Installed With Specifications:** NA

**Date of Calibration:** 07-AUG-2014

**Temperature:** (23 ± 5) °C

**Humidity:** (40 to 70)% RH

**Procedure:** B2987-99001-1

This certifies that the above product was calibrated by an approved Keysight calibration provider in accordance with procedures approved by Keysight Technologies.

**As Received:** Factory tested. No incoming data available.

**As Shipped Condition:** At the completion of the calibration, measured values were IN SPECIFICATION at the points tested.

These calibration procedures and test points are those recommended in a procedure approved by Keysight.

**Remarks or special requirements:**

For inquiries regarding this calibration please contact Keysight Technologies at the above address.

**Traceability Information:** Traceability is to the International System of Units (SI), consensus standards or ratio type measurements through national standards realized and maintained by the NIST U.S., NRC Canada, NMIJ Japan, KRIS Korea, Euramet members (NPL, PTB, etc.), NML-SIRIM in Malaysia or other National Measurement Institutes signatories to the CIPM MRA. Supporting documentation relative to traceability is available for review by appointment. This report shall not be reproduced, except in full, without prior written approval of the calibration facility.

**Print Date:** 07-AUG-2014

Keysight Provider #25865

	DD	MM	YY	BY
Cal	07	08	14	CT
Due				

Nithia Kumar  
 Sr. Quality Manager

Keysight Calibration Provider #25865



Keysight Technologies Malaysia Sdn Bhd (463532-M)  
(Formerly known as Agilent Technologies Microwave  
Products (M) Sdn Bhd)  
Bayan Lepas Free Industrial Zone  
11900 Penang, Malaysia  
Keysight Approved Calibration Provider #25865



## Certificate of Calibration

Certificate No: 4109227-4893557-1

### Calibration Equipment Used

*Date Used: Date equipment used in this calibration*

<u>Model Number</u>	<u>Model Description</u>	<u>Trace Number</u>	<u>Date Used</u>	<u>Cal Due Date</u>
3458A	8-1/2 Digit Multimeter	PD7763	07-AUG-2014	09-JAN-2015
16353H	STANDARD RESISTOR	PD7971	07-AUG-2014	06-JAN-2015
16353A	Standard Resistor Set	PD7970	07-AUG-2014	08-JAN-2015
16380C	Capacitance Standard	PD7978	07-AUG-2014	07-NOV-2014
16380A	Standard Air Capacitor Set	PD7973	07-AUG-2014	15-JAN-2015
16380A	Standard Air Capacitor Set	PD7972	07-AUG-2014	08-JAN-2015
16380C	Capacitance Standard	PD7974	07-AUG-2014	26-SEP-2014
16353H	STANDARD RESISTOR	PD7706	07-AUG-2014	25-FEB-2015
3458A	8-1/2 Digit Multimeter	PD7289	07-AUG-2014	17-JUN-2015
16353A	Standard Resistor Set	PD7683	07-AUG-2014	02-SEP-2014
FLU 714	THERMOCOUPLE CALIBRATOR	PD7990	07-AUG-2014	20-JAN-2015
B2962A	6.5 Digit Low Noise Power Source, 2ch	PD7991	07-AUG-2014	02-JUL-2015
B2962A	6.5 Digit Low Noise Power Source, 2ch	PD7992	07-AUG-2014	02-JUL-2015