



Keysight Technologies Malaysia Sdn Bhd (463532-M)
 Bayan Lepas Free Industrial Zone
 11900 Penang, Malaysia
 Keysight Approved Calibration Provider #25865



Certificate of Calibration

Certificate No: 4188250-5025769-1

Manufacturer: Keysight Technologies

Description: Electrometer/High Resistance Meter, 0.01fA, 1000V

Model No: B2987A

Serial No: MY54320308

Options Installed With Specifications: NA

Date of Calibration: 06-MAY-2016

Temperature: (23 ± 5) °C

Humidity: (40 to 70)% RH

Procedure: B2987-99001-1

This certifies that the above product was calibrated by an approved Keysight calibration provider in accordance with procedures approved by Keysight Technologies.

As Received: Factory tested. No incoming data available.

As Shipped Condition: At the completion of the calibration, measured values were IN SPECIFICATION at the points tested.

These calibration procedures and test points are those recommended in a procedure approved by Keysight.

Remarks or special requirements:

For inquiries regarding this calibration please contact Keysight Technologies at the above address.

Traceability Information: Traceability is to the International System of Units (SI), consensus standards or ratio type measurements through national standards realized and maintained by the NIST U.S., NRC Canada, NMIJ Japan, KRISS Korea, Euramet members (NPL, PTB, etc.), NML-SIRIM in Malaysia or other National Measurement Institutes signatories to the CIPM MRA. Supporting documentation relative to traceability is available for review by appointment. This report shall not be reproduced, except in full, without prior written approval of the calibration facility.

Calibration Equipment Used

Date Used: Date equipment used in this calibration

<u>Model Number</u>	<u>Model Description</u>	<u>Trace Number</u>	<u>Date Used</u>	<u>Cal Due Date</u>
16380A	Standard Air Capacitor Set	PD7305	06-MAY-2016	25-AUG-2016
16380C	Capacitance Standard	PD7559	06-MAY-2016	28-SEP-2016
3458A	8-1/2 Digit Multimeter	PD7664	06-MAY-2016	23-SEP-2016
16353A	Standard Resistor Set	PD7371	06-MAY-2016	19-OCT-2016
FLU 714	THERMOCOUPLE CALIBRATOR	PE7028	06-MAY-2016	21-AUG-2016
16353H	STANDARD RESISTOR	PD7296	06-MAY-2016	24-JUL-2016
B2962A	6.5 Digit Low Noise Power Source, 2ch	PD7991	06-MAY-2016	01-JUL-2016

Print Date: 06-MAY-2016

Keysight Provider #25865

	DD	MM	YY	BY
Cal	06	05	16	CT
Due				

Nithia Kumar
 Sr. Quality Manager

Keysight Calibration Provider #25865