

Keysight E8257D/57N/67D and E8663D PSG Signal Generators Removable Flash Memory Option 008/009 Upgrade Kit

Notice: This document contains references to Agilent. Please note that Agilent's Test and Measurement business has become Keysight Technologies. For more information, go to www.keysight.com.

Installation Note

Part Number E8251-90371
Printed in USA August 2014

Notice

The information contained in this document is subject to change without notice.

Keysight Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Keysight Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Removable Flash Memory Option 008/009 Upgrade Kit

Products Affected:	E8257D/57N/67D, E8663D
Serial Numbers:	E8257D: US4928/MY4928/SG4928 and above E8257N: US4928 and above E8267D: US/4829/MY4829/SG4829 and above E8663D: US4928/MY4928/SG4928 and above
Options:	
To Be Performed By:	(X) Agilent Technologies Service Center (X) Personnel Qualified by Agilent Technologies (X) Customer
Estimated Installation Time:	< 20 minutes
Estimated Verification Time:	None required

Introduction

This upgrade kit adds Option 008, Removable Flash Memory, to the E8257D/57N and E8663D PSG Signal Generators, or Option 009, Removable Flash Memory, to the E8267D Vector Signal Generator. This upgrade kit can only be used in signal generators with prefixes as listed above.

Installation includes completing the following major steps, in the order given.

1. Installing the IC 8G-BIT Compact Flash Memory Card.
2. Redeeming the license entitlement certificate and entering the license key into the instrument.
3. Verifying installation and activation.

Installation Kit Parts List

Item	Quantity	Description	Part Number
1	1	IC 8G-BIT COMPACT FLASH MEMORY	1819-0427
2	1	Installation Note, Option 008/009 Upgrade Kit	E8251-90371
3	1	Entitlement Certificate	n/a

Tools Required

- Straight Slot Screw Driver

Install the Flash Memory Card

1. Using a straight slot screw driver, loosen the latch screw and open the “Removable Compact Flash” door located on the rear panel.



2. Install the IC flash memory card provided with the upgrade kit.



3. Close the door and tighten the latch screw.

Activate Option 008/009

1. Follow the instructions printed on the Entitlement Certificate.

Verify Installation and Activation

1. Cycle instrument power if you have not done so after entering the license key.
2. Press **Utility > Instrument Info/Help Mode > Options Info** or **Diagnostic Info** and verify that 008/009 is listed in the option string.

NOTE Option 008/009 is not listed if the flash memory card is not installed.

Installation is complete and the instrument is ready for use.

