

# Manual Supplement

Manual Title: 1504 Users  
Revision/Date: 571202

Supplement Issue: 1  
Issue Date: 1/19  
Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

**FLUKE®**

**Calibration**

## Change #1, 573

On page 30, under *Command List Continued*, change:

From: IDN? read manufacturer, model number, serial number, and firmware version number (SCPIcompatible)

To: \*IDN? read manufacturer, model number, serial number, and firmware version number (SCPIcompatible)