

REMOTE CONTROL MANUAL

MODELS 9410/14/20/24/30/50
DUAL- AND QUAD-CHANNEL
DIGITAL OSCILLOSCOPES

Serial Number _____

May 1992

LeCroy

Corporate Headquarters

700 Chestnut Ridge Road
Chestnut Ridge, NY 10977-6499
Tel: (914) 425-2000, TWX: 710-577-2832

European Headquarters

2, rue Pré-de-la-Fontaine
P.O. Box 341
1217 Meyrin 1/Geneva, Switzerland
Tel.: (022) 719 21 11, Telex: 419 058

Copyright® May 1992, LeCroy. All rights reserved. Information in this publication supersedes all earlier versions. Specifications subject to change.

TABLE OF CONTENTS

1 General Information

Initial Inspection	1
Warranty	1
Product Assistance	1
Maintenance Agreements	1
Document Discrepancies	2
Service Procedure	2
Return Procedure	2

2 About Remote Control

GPIB Implementation Standard	3
Program Messages	3
Commands and Queries	4
Local and Remote State	5
Program Message Form	5
Command/Query Form	6
Response Message Form	9

3 GPIB Operation

GPIB Structure	11
Interface Capabilities	11
Addressing	12
GPIB Signals	12
IEEE 488.1 Standard Messages	13
Programming GPIB Transfers	15
Programming Service Requests	19
Instrument Polls	21
Driving a Hard-copy Device	25

4 RS-232-C Operation

Introduction	29
RS-232-C Pin Assignments	29
RS-232-C Configuration	30
Commands Simulating GPIB Commands	33

5 System Commands

Organization	35
Command Summary	35
Command Execution	37
Command Notation	37

6 Waveform Structure

Introduction	179
Logical Data Blocks of a Waveform	179
Inspect? Command	180
Waveform? Command	182
Waveform Command	187
More Control of Waveform Queries	188
High-speed Waveform Transfer	188

7 Status Registers

Overview of Status and Service Request Reporting	191
Status Byte Register (STB)	193
Standard Event Status Register (ESR)	194
Standard Event Status Enable Register (ESE)	195
Service Request Enable Register (SRE)	195
Parallel Poll Enable Register (PRE)	195
Internal State Change Status Register (INR)	195
Internal State Change Enable Register (INE)	196
Command Error Status Register (CMR)	196
Device Dependent Error Status Register (DDR)	196
Execution Error Status Register (EXR)	196
User Request Status Register (URR)	196

Appendix A

Example 1: Use of the Interactive GPIB Program 'IBIC'	199
Example 2: GPIB Program for IBM PC (High-level Function Calls)	200
Example 3: GPIB Program for IBM PC (Low-level Function Calls)	202

Appendix B

The Waveform Template	205
-----------------------	-----



INITIAL INSPECTION

It is recommended that the shipment be thoroughly inspected immediately upon delivery to the purchaser. All material in the container should be checked against the enclosed Packing List. LeCroy cannot accept responsibility for shortages in comparison with the Packing List unless notified promptly. If the shipment is damaged in any way, please contact the Customer Service Department or local field office immediately.

WARRANTY

LeCroy warrants its oscilloscope products to operate within specifications under normal use for a period of two years from the date of shipment. Spares, replacement parts and repairs are warranted for 90 days. The instrument's firmware is thoroughly tested and thought to be functional, but is supplied "as is" with no warranty of any kind covering detailed performance. Products not manufactured by LeCroy are covered solely by the warranty of the original equipment manufacturer.

In exercising this warranty, LeCroy will repair or, at its option, replace any product returned to the Customer Service Department or an authorized service facility within the warranty period, provided that the warrantor's examination discloses that the product is defective due to workmanship or materials and that the defect has not been caused by misuse, neglect, accident or abnormal conditions or operation.

The purchaser is responsible for transportation and insurance charges for the return of products to the servicing facility. LeCroy will return all in-warranty products with transportation prepaid.

This warranty is in lieu of all other warranties, expressed or implied, including but not limited to any implied warranty of merchantability, fitness, or adequacy for any particular purpose or use. LeCroy shall not be liable for any special, incidental, or consequential damages, whether in contract or otherwise.

PRODUCT ASSISTANCE

Answers to questions concerning installation, calibration, and use of LeCroy equipment are available from the Customer Service Department, 700 Chestnut Ridge Road, Chestnut Ridge, New York 10977-6499, U.S.A., tel. (914)578-6061, and 2, rue du Pré-de-la-Fontaine, 1217 Meyrin 1, Geneva, Switzerland, tel. (41)22/719 21 11, or your local field engineering office.

MAINTENANCE AGREEMENTS

LeCroy offers a selection of customer support services. Maintenance agreements provide extended warranty and allow the customer to budget maintenance costs after the initial two year warranty has expired. Other services such as installation, training, enhancements and on-site repair are available through specific Supplemental Support Agreements.

DOCUMENTATION DISCREPANCIES

LeCroy is committed to providing state-of-the-art instrumentation and is continually refining and improving the performance of its products. While physical modifications can be implemented quite rapidly, the corrected documentation frequently requires more time to produce. Consequently, this manual may not agree in every detail with the accompanying product. There may be small discrepancies in the values of components for the purposes of pulse shape, timing, offset, etc., and, occasionally, minor logic changes. Where any such inconsistencies exist, please be assured that the unit is correct and incorporates the most up-to-date circuitry. In a similar way the firmware may undergo revision when the instrument is serviced. Should this be the case, manual updates will be made available as necessary.

SERVICE PROCEDURE

Products requiring maintenance should be returned to the Customer Service Department or authorized service facility. LeCroy will repair or replace any product under warranty at no charge. The purchaser is only responsible for transportation charges.

For all LeCroy products in need of repair after the warranty period, the customer must provide a Purchase Order Number before repairs can be initiated. The customer will be billed for parts and labor for the repair, as well as for shipping.

RETURN PROCEDURE

To determine your nearest authorized service facility, contact the Customer Service Department or your field office. All products returned for repair should be identified by the model and serial numbers and include a description of the defect or failure, name and phone number of the user, and, in the case of products returned to the factory, a Return Authorization Number (RAN). The RAN may be obtained by contacting the Customer Service Department in New York, tel. (914)578-6061, in Geneva, tel. (41)22/719 21 11, or your nearest sales office.

Return shipments should be made prepaid. LeCroy will not accept C.O.D. or Collect Return Shipments. Air-freight is generally recommended. Wherever possible, the original shipping carton should be used. If a substitute carton is used, it should be rigid and be packed such that the product is surrounded with a minimum of four inches of excelsior or similar shock-absorbing material. In addressing the shipment, it is important that the Return Authorization Number be displayed on the outside of the container to ensure its prompt routing to the proper department within LeCroy.

Two modes of operation are available in the oscilloscope. The instrument may be operated either manually, by using the front-panel controls, or remotely by means of an external controller (which is usually a computer, but may be a simple terminal). This Remote Control Manual describes how to control the oscilloscope in the remote mode. For explanations on how to manually set front-panel controls, refer to the Operator's Manual.

The oscilloscope is remotely controlled via either the GPIB (General Purpose Interface Bus) or the RS-232-C communication ports. Whenever the rear-panel GPIB address switches are set between 0 and 30, control is via GPIB; when they are at 31 or above, control is via RS-232-C. The instrument can be fully controlled in remote mode. The only actions which cannot be performed remotely are switching on the instrument or setting the remote address.

This section introduces the basic remote control concepts which are common to both RS-232-C and GPIB. It also presents a brief description of remote control messages.

Sections 3 and 4 explain how to send program messages over the GPIB or the RS-232-C interfaces, respectively. Section 5 alphabetically lists all the remote control commands. Section 6 is a detailed description and tutorial of the transfer and format of waveforms, whereas Section 7 explains the use of status bytes for error reporting. Appendix A shows some complete programming examples. Appendix B contains a printout of a waveform template.

GPIB IMPLEMENTATION STANDARD

The remote commands conform to the GPIB IEEE 488.2 standard¹. This standard may be seen as an extension of the IEEE 488.1 standard which dealt mainly with electrical and mechanical issues. The IEEE 488.2 recommendations have also been adopted for RS-232-C communications whenever applicable.

PROGRAM MESSAGES

To remotely control the oscilloscope the controller must send program messages which conform to precise format structures. The instrument will execute all program messages which are in the correct form and ignore those where errors are detected.

1. ANSI/IEEE Std. 488.2-1987, "IEEE Standard Codes, Formats, Protocols, and Common Commands", The Institute of Electrical and Electronics Engineers Inc., 345 East 47th Street, New York, NY 10017, USA.

Warning or error messages are normally not reported by the instrument, unless the controller explicitly examines the relevant status register, or if the status enable registers have been set in such a way that the controller can be interrupted when an error occurs. The status registers are explained in Section 7.

During the development of the control program it is possible to observe all remote control transactions, including error messages, on an external monitor connected to the RS-232-C port. Refer to the command "COMM_HELP" for further details.

COMMANDS AND QUERIES

Program messages consist of one or several commands or queries. A command directs the instrument to change its state, e.g. to change its time base or vertical sensitivity. A query asks the instrument about its state. Very often, the same mnemonic is used for a command and a query, the query being identified by a <?> after the last character.

For example, to change the time base to 2 msec/div, the controller should send the following command to the instrument

```
TIME_DIV 2 MS
```

To ask the instrument about its time base, this query should be sent

```
TIME_DIV?
```

A query causes the instrument to send a response message. The control program should read this message with a "read" instruction to the GPIB or RS-232-C interface of the controller. The response message to the query above might be

```
TIME_DIV 10 NS
```

The portion of the query preceding the question mark is repeated as part of the response message. If desired, this text may be suppressed with the command "COMM_HEADER".

Depending on the state of the instrument and the computation to be done, the controller may have to wait up to several seconds for a response. Command interpretation does not have priority over other oscilloscope activities. It is therefore judicious to set the controller IO timeout conditions to 3 or more seconds. In addition, it must be remembered that an incorrect query message will not generate a response message.

LOCAL AND REMOTE STATE

As a rule, remote commands are only executed by the instrument when it is in the REMOTE state, whereas queries are always executed. A few commands which don't affect the state of the front panel are also executed in LOCAL (refer to the beginning of Section 5 for a list of these commands). When the instrument is in REMOTE, all front-panel controls are disabled, except the left-hand menu buttons, the intensity controls (which can be disabled with the command "INTENSITY") and the LOCAL button (which can be disabled by setting the instrument to LOCAL LOCKOUT). For an explanation on how to set the instrument to LOCAL, REMOTE or LOCAL LOCKOUT, refer to Section 3 for GPIB and to Section 4 for RS-232-C.

PROGRAM MESSAGE FORM

An instrument is remotely controlled with program messages which consist of one or several commands or queries, separated by semicolons <;> and ended by a terminator:

<command/query>;.....;<command/query> <terminator>

Upper and/or lower case characters can be used for program messages.

The instrument does not decode an incoming program message before a terminator has been received (exception: if the program message is longer than the 256 byte input buffer of the instrument, the oscilloscope starts analyzing the message when the buffer is full). The commands or queries are executed in the order in which they are transmitted.

In GPIB mode, the following are valid terminators:

- <NL> New-line character (i.e. the ASCII new-line character, whose decimal value is 10).
- <NL> <EOI> New-line character with a simultaneous <EOI> signal.
- <EOI> <EOI> signal together with the last character of the program message.

Note: The <EOI> signal is a dedicated GPIB interface line which can be set with a special call to the GPIB interface driver. Refer to the GPIB interface manufacturer's manual and support programs.

The <NL> <EOI> terminator is always used in response messages sent by the instrument to the controller.

In RS-232-C, the terminator may be defined by the user with the command "COMM_RS232". The default value is <CR>, i.e. the ASCII carriage return character, the decimal value of which is 13.

Examples

GRID DUAL

This program message consists of a single command which instructs the instrument to display a dual grid. The terminator is not shown since it is usually automatically added by the interface driver routine which writes to the GPIB (or RS-232).

BWL ON; DISPLAY OFF; DATE?

This program message consists of two commands, followed by a query. They instruct the instrument to turn on the bandwidth limit, turn off the display, and then ask for the current date. Again, the terminator is not shown.

COMMAND/QUERY FORM

The general form of a command or a query consists of a command header <header> which is optionally followed by one or several parameters <data> separated by commas:

<header>[?] <data>,...,<data>

The notation [?] shows that the question mark is optional (turning the command into a query). The detailed listing of all commands in Section 5 indicates which commands may also be queries. There is a space between the header and the first parameter. There are commas between parameters.

Example

DATE 15,OCT,1989,13,21,16

This command instructs the oscilloscope to set its date and time to 15 OCT 1989, 13:21:16. The command header "DATE" indicates the action, the 6 data values specify it in detail.

Header

The header is the mnemonic form of the operation to be performed by the oscilloscope. All command mnemonics are listed in alphabetic order in Section 5.

The majority of the command/query headers have a long form for optimum legibility and a short form for better transfer and decoding speed. The two forms are fully equivalent and can be used interchangeably. For example, the following two commands for switching to the automatic trigger mode are fully equivalent:

TRIG_MODE AUTO and TRMD AUTO

Some command/query mnemonics are imposed by the IEEE 488.2 standard. They are standardized so that different instruments present the same programming interface for similar functions. All these mnemonics begin with an asterisk <*>, e.g. the command “*RST” is the IEEE 488.2 imposed mnemonic for resetting the instrument, whereas “*TST?” instructs the instrument to perform an internal self-test and to report the outcome.

Header path

Some commands or queries apply to a sub-section of the oscilloscope, e.g. a single input channel or a trace on the display. In such cases, the header must be preceded by a path name that indicates the channel or trace to which the command applies. The header path normally consists of a 2-letter path name followed by a colon <:> which immediately precedes the command header.

Usually one of the waveform traces can be specified in the header path (refer to the individual commands listed in Section 5 for details on which values apply to a given command header):

- C1, C2 Channels 1 and 2
- C3, C4 Channels 3 and 4 (in 4-channel instruments)
- MC, MD Memory C and D
- FE, FF Function E and F
- EA, EB Expand A and B
- EX, EX10 External trigger

Example

C1:OFST -300 MV Set the offset of Channel 1 to -300 mV

Header paths need only be specified once. Subsequent commands whose header destination is not indicated are assumed to refer to the last defined path. For example, the following commands are identical:

C2:VDIV?; C2:OFST? What is the vertical sensitivity and the offset of channel 2?

C2:VDIV?; OFST? Same as above, without repeating the path.

Data

Whenever a command/query uses additional data values, they are expressed in terms of ASCII characters. There is a single exception: the transfer of waveforms with the command/query “WAVEFORM”, where the waveform may be expressed as a sequence of binary data values. Refer to Section 6 for a detailed explanation of the format of waveforms.

ASCII data can have the form of character, numeric, string or block data.

Character data

These are simple words or abbreviations for the indication of a specific action.

BANDWIDTH_LIMIT ON The data value "ON" indicates that the bandwidth limit should be turned on, rather than off.

In some commands, where as many as a dozen different parameters can be specified, or where not all parameters apply at the same time, the format requires pairs of data values. The first one names the parameter to be modified and the second gives its value. Only those parameter pairs to be changed need to be indicated.

HARDCOPY_SETUP DEV,HP7470A,PORT,GPIB,PSIZE,A4
Three pairs of parameters are specified. The first specifies the device as the H7470A plotter (or compatible), the second indicates the GPIB port and the third requests the A4 format for paper size. While the command "HARDCOPY_SETUP" allows many more parameters, they are either not relevant for plotters or they are left unchanged.

Numeric Data

The numeric data type is used to enter quantitative information. Numbers can be entered as integers, as fractions or in exponential representation.

EA:VPOS -5 Move the displayed trace of Expand A downwards by 5 divisions.

C2:OFST 3.56 Set the DC offset of Channel 2 to 3.56 V.

TDIV 5.0E-6 Adjust the time base to 5 μ sec/div.

Note: Numeric values may be followed by multipliers and units, modifying the value of the numerical expression. The following mnemonics are recognized:

EX	1E18	Exa-	PE	1E15	Peta-
T	1E12	Tera-	G	1E9	Giga-
MA	1E6	Mega-	K	1E3	kilo-
M	1E-3	milli-	U	1E-6	micro-
N	1E-9	nano-	PI	1E-12	pico-
F	1E-15	femto-	A	1E-18	atto-

For example, there are many ways of setting the time base of the instrument to 5 μ sec/div:

TDIV 5E-6 Exponential notation, without any suffix.

TDIV 5 US Suffix multiplier "U" for 1E-6, with the (optional) suffix "S" for seconds.

TDIV 5000 NS

TDIV 5000E-3 US

String Data

This data type enables the transfer of a (long) string of characters as a single parameter. String data are formed by simply enclosing any sequence of ASCII characters between simple or double quotes.

MESSAGE 'Connect probe to point J3'

The instrument displays this message in the Message field above the grid.

Block Data

These are binary data values coded in hexadecimal ASCII, i.e. 4-bit nibbles are translated into the digits 0,...9, A,...F and transmitted as ASCII characters. They are only used for the transfer of waveforms (command "WAVEFORM") and of the instrument configuration (command "PANEL_SETUP")

RESPONSE MESSAGE FORM

The instrument sends a response message to the controller, as an answer to a query. The format of such messages is the same as that of program messages, i.e. individual responses in the format of commands, separated by semicolons <> and ended by a terminator. They can be sent back to the instrument in the form in which they are received, and will be accepted as valid commands. In GPIB response messages, the <NL> <EOI> terminator is always used.

For example, if the controller sends the program message:

TIME_DIV?;TRIG_MODE NORM;C1:COUPLING? (terminator not shown)

the instrument might respond as follows:

```
TIME_DIV 50 NS;C1:COUPLING D50 (terminator not shown)
```

The response message only refers to the queries, i.e. "TRIG_MODE" is left out. If this response is sent back to the instrument, it is a valid program message for setting its time base to 50 nsec/div and the input coupling of Channel 1 to 50 Ω .

Whenever a response is expected from the instrument, the control program must instruct the GPIB or RS-232-C interface to read from the instrument. If the controller sends another program message without reading the response to the previous one, the response message in the output buffer of the instrument is discarded.

The instrument uses somewhat stricter rules for response messages than for the acceptance of program messages. Whereas the controller may send program messages in upper or lower case characters, response messages are always returned in upper case. Program messages may contain extraneous spaces or tabs (white space), response messages do not. Whereas program messages may contain a mixture of short and long command/query headers, response messages always use short headers as a default. However, the instrument can be forced with the command "COMM_HEADER" to use long headers or no headers at all. If the response header is omitted, the response transfer time is minimized, but such a response could not be sent back to the instrument again. In this case suffix units are also suppressed in the response.

If the trigger slope of Channel 1 is set to negative, the query "C1:TRSL?" could yield the following responses:

C1:TRIG_SLOPE NEG	header format: long
C1:TRSL NEG	header format: short
NEG	header format: off

Waveforms which are obtained from the instrument using the query "WAVEFORM?" constitute a special kind of response message. Their exact format can be controlled with the commands "COMM_FORMAT" and "COMM_ORDER", as explained in Section 6.

This section describes how to remotely control the oscilloscope via the GPIB. Topics discussed include interface capabilities, addressing, standard bus commands, and polling schemes.

GPIB STRUCTURE

The GPIB is like an ordinary computer bus, except that it interconnects independent devices via a cable bus whereas a computer has its circuit cards interconnected via a backplane bus. The GPIB carries program messages and interface messages:

- Program messages, often called device-dependent messages, contain programming instructions, measurement results, instrument status and waveform data. Their general form is described in Section 2.
- Interface messages manage the bus itself. They perform functions such as initializing the bus, addressing and unaddressing devices and setting remote and local modes.

Devices on the GPIB can be listeners, talkers, and/or controllers. A talker sends program messages to one or more listeners. A controller manages the flow of information on the bus by sending interface messages to the devices.

The oscilloscope can be a talker or a listener, but not a controller. The host computer, however, must be able to act as a listener, talker and controller. For details on how the controller configures the GPIB for specific functions, refer to the GPIB interface manufacturer's manual.

INTERFACE CAPABILITIES

The interface capabilities of the oscilloscope include the following IEEE 488.1 definitions:

AH1	Complete Acceptor Handshake
SH1	Complete Source Handshake
L4	Partial Listener Function
T5	Complete Talker Function
SR1	Complete Service Request Function
RL1	Complete Remote/Local Function
DC1	Complete Device Clear Function
DT1	Complete Device Trigger
PP1	Parallel Polling: remote configurability
C0	No Controller Functions
E2	Tri-state Drivers

3 GPIB Operation

ADDRESSING

Every device on the GPIB has an address. When the thumbwheel address switches on the rear panel of the oscilloscope are set to a value between 0 and 30, the instrument can be controlled via GPIB. When the switches are set to above 30, the instrument can execute talk-only operations on the GPIB, for example driving a GPIB plotter. In this case no controller is present and the instrument is directly connected to the plotter. Addresses above 30 also enable the instrument to be controlled via the RS-232-C port.

The instrument reads the address switches once at power on, or when the RESET button on the rear panel is pressed. If the address is changed during operation, the instrument must be powered again to enable the new address. The value of the GPIB address appears in the menu "Auxiliary Setups".

If the oscilloscope is addressed to talk, it will remain configured to talk until a universal untalk command (UNT), its own listen address (MLA), or another instrument's talk address is received.

Similarly, if the oscilloscope is addressed to listen, it will remain configured to listen until a universal unlisten command (UNL), or its own talker address (MTA) is received.

GPIB SIGNALS

The bus system consists of 16 signal lines and 8 ground or shield lines. The signal lines are divided into 3 groups:

- 8 data lines
- 3 handshake lines
- 5 interface management lines

Data Lines

The eight data lines, usually called DI01 through DI08, carry both program and interface messages. Most of the messages use the 7-bit ASCII code, in which case DI08 is unused.

Handshake Lines

These three lines control the transfer of message bytes between devices. The process is called a three-wire interlocked handshake and it guarantees that the message bytes on the data lines are sent and received without transmission error.

Interface Management Lines

The following five lines manage the flow of information across the interface.

ATN (ATteNtion): The controller drives the ATN line true when it uses the data lines to send interface messages such as talk and listen addresses or a device clear (DCL) message. When ATN is false, the bus is in the data mode for the transfer of program messages from talkers to listeners.

IFC (InterFace Clear): The controller sets the IFC line true to initialize the bus.

REN (Remote ENable): The controller uses this line to place devices in remote or local program mode.

SRQ (Service ReQuest): Any device can drive the SRQ line true to asynchronously request service from the controller. This is the equivalent of a single interrupt line on a computer bus.

EOI (End Or Identify): This line has two purposes. The talker uses it to mark the end of a message string. The controller uses it to tell devices to identify their response in a parallel poll (discussed later in this section).

I/O Buffers

The instrument has a 256-byte input buffer and a 256-byte output buffer. An incoming program message is not decoded before a message terminator has been received. However, if the input buffer becomes full (because the program message is longer than the buffer), the instrument starts analyzing the message. In this case data transmission is temporarily halted, and the controller may generate a timeout if the limit was set too low.

IEEE 488.1 STANDARD MESSAGES

The IEEE 488.1 standard specifies not only the mechanical and electrical aspects of the GPIB, but also the low-level transfer protocol, e.g. it defines how a controller addresses devices, turns them into talkers or listeners, resets them or puts them in the remote state. Such interface messages are executed with the interface management lines of the GPIB, usually with ATN true. All of these messages (except GET) are executed immediately upon reception and not in chronological order with normal commands.

Note: In addition to the IEEE 488.1 interface message standards, the IEEE 488.2 standard specifies some standardized program messages, i.e. command headers. They are identified with a leading asterisk <> and are listed among the commands in Section 5.*

The command list in Section 5 does not contain any command for clearing the input/output buffers or for setting the instrument to the remote state. This is because such commands are already specified as IEEE 488.1 standard messages. Refer to the GPIB interface manual of the host controller as well as to its support programs which should contain special calls for the execution of these messages.

The following describes those IEEE 488.1 standard messages which go beyond mere reconfiguration of the bus and which have an effect on the operation of the instrument.

Device Clear

In response to a universal Device CLear (DCL) or a Selected Device Clear message (SDC), the oscilloscope clears the input/output

buffers, aborts the interpretation of the current command (if any) and clears any pending commands. Status registers and status enable registers are not cleared. Although DCL has an immediate effect it can take several seconds to execute this command if the instrument is busy.

Group Execute Trigger

The Group Execute Trigger message (GET) causes the oscilloscope to arm the trigger system. It is functionally identical to the “*TRG” command.

Remote ENable

This interface message is executed when the controller holds the Remote ENable control line (REN) true and configures the instrument as a listener. The REMOTE LED on the front panel lights up to indicate that the instrument is set to the remote mode. All the front-panel controls are disabled except the left-hand menu buttons, the intensity controls and the LOCAL button. The menu indications on the left-hand side of the screen no longer appear since menus cannot now be operated manually. Whenever the controller returns the REN line to false, all instruments on the bus return to LOCAL. Individual instruments can be returned to LOCAL with the Go To Local message (see below).

As a rule, remote commands are only executed when the instrument is in the remote state, whereas queries are always executed. Local front-panel control may be regained by pressing the LOCAL push button, unless the instrument was placed in the Local Lockout (LLO) mode.

Local LOkout

The Local LOkout command (LLO) causes the LOCAL button on the front panel of the oscilloscope to be disabled. The LLO command can be sent in local or remote mode but only becomes effective once the instrument has been set to the remote mode.

Go To Local

The Go To Local message (GTL) causes the instrument to return to the local mode. All front-panel controls become active and the menus on the left-hand side of the screen reappear. Thereafter, whenever the instrument is addressed as a listener it will be immediately set to the remote state again.

Note that a GTL message does not clear the local lockout if it was set. Thus, whenever the instrument returns to the remote state the local lockout mode would immediately be effective again.

A command string should not be immediately followed by a GTL message. Since GTL is executed at once, the instrument may already be returned to the local state before the commands in the input buffer are interpreted. Therefore, the instrument may refuse to execute them if they require the instrument to be in REMOTE.

A safe way to ensure that all commands have been interpreted is to append a query (e.g. “*STB?”) to the command string and to wait for the response before sending a GTL.

InterFace Clear

The InterFace Clear message (IFC) initializes the GPIB but has no effect on the operation of the oscilloscope.

**PROGRAMMING
GPIB TRANSFERS**

To illustrate the GPIB programming concepts a number of examples written in BASICA are included in this section. It is assumed that the controller is IBM-PC compatible, running under DOS, and that it is equipped with a National Instruments² GPIB interface card. GPIB programming with other languages such as C or Pascal is quite similar.

If you use another computer or another GPIB interface, refer to the interface manual for installation procedures and subroutine calls similar to those described here.

**Configuring the
GPIB Hardware**

Check that the GPIB interface is properly installed in the computer. If it is not, follow the installation instructions of the interface manufacturer. In the case of the National Instruments interface, it is possible to modify the base I/O address of the board, the DMA channel number and the interrupt line setting using switches and jumpers. In our program examples, they are assumed to be left in their default positions.

Connect the oscilloscope to the computer with a GPIB interface cable. Set the GPIB address on the rear of the instrument to the required value. The program examples assume that it is set to 4. Remember to power the instrument up *after* setting the GPIB address.

**Configuring the
GPIB Driver Software**

The host computer needs an interface driver which handles the transactions between the user's programs and the interface board. In the case of the National Instruments interface, the installation procedure:

- copies the GPIB handler GPIB.COM into the boot directory.
- modifies the DOS system configuration file CONFIG.SYS to declare the presence of the GPIB handler.
- creates a sub-directory GPIB-PC.
- installs in GPIB-PC a number of files and programs which are useful for testing and reconfiguring the system, and for writing user programs.

2. National Instruments Corporation, 12109 Technology Boulevard, Austin, Texas 78727

The following files in the sub-directory GPIB-PC are of particular use:

IBIC.EXE allows interactive control of the GPIB via functions entered at the keyboard. Use of this program is highly recommended to anyone who is not familiar with GPIB programming or with the oscilloscope's remote commands. An example of the use of IBIC.EXE is shown in Appendix A.

DECL.BAS is a declaration file that contains code to be included at the beginning of any BASICA application program. Simple application programs can be quickly written by appending the user's instructions to DECL.BAS and executing the complete file.

IBCONF.EXE is an interactive program which allows inspection or modification of the current settings of the GPIB handler. To run IBCONF.EXE, refer to the National Instruments user's manual.

In the program examples in this section, it is assumed that the National Instruments GPIB driver GPIB.COM is in its default state, i.e. that the user has not modified it with IBCONF.EXE. This means that the interface board can be referred to by the symbolic name 'GPIB0' and that devices on the GPIB bus with addresses between 1 and 16 can be called by the symbolic names 'DEV1' to 'DEV16'.

Note: If you have a National Instruments PC2 interface card rather than PC2A, you must run IBCONF to declare the presence of this card rather than the default PC2A.

Simple Transfers

For a large number of remote control operations it is sufficient to use just 3 different subroutines (IBFIND, IBRD and IBWRT) provided by National Instruments. The following complete program reads the time-base setting of the oscilloscope and displays it on the terminal:

```
1-99      <DECL.BAS>
100      DEV$="DEV4"
110      CALL IBFIND(DEV$,SCOPE%)
120      CMD$="TDIV?"
130      CALL IBWRT(SCOPE%,CMD$)
140      CALL IBRD(SCOPE%,RD$)
150      PRINT RD$
160      END
```

Explanation

Lines 1 – 99 are a copy of the file DECL.BAS supplied by National Instruments. The first 6 lines are required for the initialization of the GPIB handler. The other lines are declarations which may be useful for larger programs, but are not really required code. The sample program above only uses the strings CMD\$ and RD\$ which are declared in DECL.BAS as arrays of 255 characters.

Note: DECL.BAS requires access to the file BIB.M during the GPIB initialization. BIB.M is one of the files supplied by National Instruments, and it must exist in the directory currently in use.

Note: The first 2 lines of DECL.BAS each contain a string "XXXXX" which must be replaced by the number of bytes which determine the maximum workspace for BASICA (computed by subtracting the size of BIB.M from the space currently available in BASICA). For example, if the size of BIB.M is 1200 bytes and when BASICA is loaded it reports "60200 bytes free", you should replace "XXXXX" by the value 59000 or less.

Lines 100 and 110 open the device "DEV4" and associate with it the descriptor "SCOPE%". All I/O calls from now on will refer to "SCOPE%". The default configuration of the GPIB handler recognizes "DEV4" and associates with it a device with GPIB address 4. If you want to use another GPIB address between 1 and 16, use the string "DEVx" with x = 1...16. If you want to use another name, run IBCONF.EXE to declare this name to the handler.

Lines 120 and 130 prepare the command string TDIV? and transfer it to the instrument. The command instructs it to respond with the current setting of the time base.

Line 140 reads the response of the instrument and places it into the character string RD\$.

Line 150 displays the response on the terminal.

When running this sample program, the oscilloscope will automatically be set to the remote state when IBWRT is executed, and will remain in that state. Pressing the LOCAL button on the front panel will return the oscilloscope to local mode if the GPIB handler was modified to inhibit Local Lockout (LLO).

Here is a slightly modified version of the sample program which checks if any error occurred during GPIB operation:

```
1-99      <DECL.BAS>
100      DEV$="DEV4"
110      CALL IBFIND(DEV$,SCOPE%)
120      CMD$="TDIV?"
130      CALL IBWRT(SCOPE%,CMD$)
140      IF ISTA% < 0 THEN GOTO 200
150      CALL IBRD(SCOPE%,RD$)
160      IF ISTA% < 0 THEN GOTO 250
170      PRINT RD$
180      IBLOC(SCOPE%)
190      END
200      PRINT "WRITE ERROR = ";IBERR%
210      END
250      PRINT "READ ERROR = ";IBERR%
260      END
```

The GPIB status word *ISTA%*, the GPIB error variable *IBERR%* and the count variable *IBCNT%* are defined by the GPIB handler and are updated with every GPIB function call. Refer to the National Instruments user's manual for details. The sample program above would report if the GPIB address of the instrument was set to a value other than 4. Line 180 resets the instrument to local with a call to the GPIB routine *IBLOC*.

Example 2 in Appendix A provides a more useful program which enables interactive setting and inspection of the front-panel controls as well as archiving and recalling of waveforms. Note that this program is written with just 7 different GPIB calls.

Some Additional Driver Calls

IBLOC is used to execute the IEEE 488.1 standard message Go To Local (GTL), i.e. it returns the instrument to the local state. The programming example above shows its use.

IBCLR executes the IEEE 488.1 standard message Selected Device Clear (SDC).

IBRDF and **IBWRTF** allow data to be read from GPIB to a file and data to be written from a file to GPIB respectively. Transferring data directly to or from a storage device does not limit the size of the data block, but it may be slower than transferring to the computer memory. Example 2 in Appendix A shows the use of these calls.

IBRDI and **IBWRTI** allow data to be read from GPIB to an integer array and data to be written from an integer array to GPIB. Since the integer array allows storage of up to 64 kilobytes (in BASIC), **IBRDI** and **IBWRTI** should be used for the transfer of large

data blocks to the computer memory, rather than IBRD or IBWRT which are limited to 256 bytes by the BASIC string length. Note that IBRDI and IBWRTI only exist for BASIC, since the function calls IBRD and IBWRT for more modern programming languages, such as C, are much less limited in the data block size.

IBTMO can be used to change the time-out value during program execution. The default value of the GPIB driver is 10 seconds, e.g. if the instrument does not respond to a IBRD call, IBRD will return with an error after the specified time.

IBTRG executes the IEEE 488.1 standard message Group Execute Trigger (GET), which causes the oscilloscope to arm the trigger system.

National Instruments supply a number of additional function calls. In particular, it is possible to use the so-called board level calls which allow a very detailed control of the GPIB. The use of such calls is shown in Example 3 of Appendix A.

PROGRAMMING SERVICE REQUESTS

When an oscilloscope is used in a remote application, events often occur asynchronously, i.e. at times that are unpredictable for the host computer. The most common case is waiting for a trigger after the instrument has been armed. The controller must wait until the acquisition is finished before it can read the acquired waveform. The simplest way of checking if a certain event has occurred is by continuously or periodically reading the status bit associated with it until the required transition is detected. Continuous status bit polling is described in more detail in the sub-section "Instrument Polls". For a complete explanation of the status bytes refer to Section 7.

A potentially more efficient way of detecting events occurring in the instrument is the use of the Service Request (SRQ). This GPIB interrupt line can be used to interrupt program execution in the controller. Therefore, the controller can execute other programs while waiting for the instrument. Unfortunately, not all interface manufacturers support the programming of interrupt service routines. In particular, National Instruments only supports the SRQ bit within the `ISTA%` status word. This requires the user to continuously or periodically check this word, either explicitly or with the function call `IBWAIT`. In the absence of real interrupt service routines the use of SRQ may not be very advantageous.

In the default state, after power-on, the Service ReQuest is disabled. The SRQ is enabled by setting the Service Request Enable register with the command "`*SRE`" and specifying which event should generate an SRQ. The oscilloscope will interrupt the con-

troller as soon as the selected event(s) occur by asserting the SRQ interface line. If several devices are connected to the GPIB, the controller may have to identify which instrument caused the interrupt by serial polling the various devices.

*Note: The SRQ bit is latched until the controller reads the Status Byte Register (STB). The action of reading the STB with the command "*STB?" clears the register contents except the MAV bit (bit 4) until a new event occurs. Service requesting may be disabled by clearing the SRE register ("*SRE 0").*

Example 1

To assert SRQ in response to the events "new signal acquired" or "return-to-local" (pressing the front-panel button LOCAL).

These events are tracked by the INR register which is reflected in the SRE register as the INB summary bit in position 0. Since the bit position 0 has the value 1, the command "*SRE 1" enables the generation of SRQ whenever the INB summary bit is set.

In addition, the events of the INR register which may be summarized in the INB bit must be specified. The event "new signal acquired" corresponds to INE bit 0 (value 1) while the event "return-to-local" is assigned to INE bit 2 (value 4). The total sum is 1+4=5. Thus the command "INE 5" is needed.

```
CMD$="INE 5;*SRE 1"  
CALL IBWRT(SCOPE%,CMD$)
```

Example 2

To assert SRQ when soft key 10 is pressed.

The event "soft key 10 pressed" is tracked by the URR register. Since the URR register is not directly reflected in STB but only in the ESR register (URR, bit position 6), the ESE enable register must be set first with the command "*ESE 64" to allow the URQ setting to be reported in STB. An SRQ request will now be generated provided that the ESB summary bit (bit position 5) in the SRE enable register is set ("*SRE 32").

```
CMD$="*ESE 64;*SRE 32"  
CALL IBWRT(SCOPE%,CMD$)
```

INSTRUMENT POLLS

State transitions occurring within the instrument can be remotely monitored by polling selected internal status registers. This subsection discusses a number of polling methods which may be used to detect the occurrence of a given event.

1. Continuous poll
2. Serial poll
3. Parallel poll
4. *IST poll

To emphasize the differences between these methods, the same example will be presented in each case, i.e. determining if a new acquisition has taken place. By far the simplest poll is the continuous poll. The other methods only make sense if interrupt service routines (servicing the SRQ line) are supported or if multiple devices on GPIB must be monitored simultaneously.

Continuous Poll

In continuous polling a status register is continuously monitored until a transition is observed. This is the most straightforward method for detecting state changes but may be impracticable in some situations, especially in multiple device configurations.

In the following example, the event "new signal acquired" is observed by continuously polling the INternal state change Register (INR) until the corresponding bit (in this case bit 0, i.e. value 1) is non-zero to indicate that a new waveform has been acquired. Reading INR clears it at the same time so that there is no need for an additional clearing action after a non-zero value has been detected. The command "CHDR OFF" instructs the instrument to omit any command headers when responding to a query. This simplifies the decoding of the response. The instrument would therefore send "1" rather than "INR 1".

```

CMD$="CHDR OFF"
CALL IBWRT(SCOPE%,CMD$)
MASK% = 1           'New Signal Bit has value 1
LOOP% = 1
WHILE LOOP%
    CMD$="INR?"
    CALL IBWRT(SCOPE%,CMD$)
    CALL IBRD(SCOPE%,RD$)
    NEWSIG% = VAL(RD$) AND MASK%
    IF NEWSIG% = MASK% THEN LOOP% = 0
WEND
    
```

Serial Poll

Serial polling takes place once the SRQ interrupt line has been asserted. The controller examines which instrument has generated the interrupt by inspecting the SRQ bit in the STB register of each instrument. Because service request is based on an interrupt mechanism, serial polling offers a reasonable compromise in terms of servicing speed in multiple device configurations.

In the following example, the command "INE 1" enables the event "new signal acquired" to be reported in the INR to the INB bit of the status byte STB. The command "*SRE 1" enables the INB of the status byte to generate an SRQ whenever it is set. The function call IBWAIT instructs the computer to wait until one of three conditions occur: &H8000 in the mask (MASK%) corresponds to a GPIB error, &H4000 to a time-out error and &H0800 to the detection of RQS (ReQuest for Service generated by the SRQ bit).

Whenever IBWAIT detects RQS it automatically performs a serial poll to find out which instrument generated the interrupt. It will only exit if there was a time-out or if the instrument "SCOPE%" generated SRQ. The additional function call IBRSP fetches the value of the status byte which may be further interpreted. For this example to function properly the value of 'Disable Auto Serial Polling' must be set 'off' in the GPIB handler (use IBCONF.EXE to check).

```
CMD$="*CLS; INE 1; *SRE 1"  
CALL IBWRT(SCOPE%,CMD$)  
MASK% = &HC800  
CALL IBWAIT(SCOPE%,MASK%)  
IF (IBSTA% AND &HC000) <> 0 THEN PRINT "GPIB or  
Time-out Error" : STOP  
CALL IBRSP(SCOPE%,SPR%)  
PRINT "Status Byte = ", SPR%
```

*Note: After the serial poll is completed, the RQS bit in the STB status register is cleared. Note that the other STB register bits remain set until they are cleared by means of a "*CLS" command or the instrument is reset. If these bits are not cleared, they cannot generate another interrupt.*

Serial polling is only an advantage if there are several instruments that may need attention. Board-level function calls can deal simultaneously with several instruments attached to the same interface board. Refer to the National Instruments user's manual.

Parallel Poll

Parallel polling is only an advantage if there are several instruments that may need attention.

In parallel polling, the controller simultaneously reads the Individual STatus bit (IST) of all the instruments to determine which one needs service. Since parallel polling allows up to eight different instruments to be polled at the same time, parallel polling is the fastest way to identify state changes of instruments supporting this capability.

When a parallel poll is initiated, each instrument returns a status bit via one of the DIO data lines. Devices may respond either individually using a separate DIO line or collectively on a single data line. Data line assignments are made by the controller via a Parallel Poll Configure (PPC) sequence.

In the following example, the command "INE 1" enables the event "new signal acquired" in the INR to be reported to the INB bit of the status byte STB. The PaRallel poll Enable register (PRE) determines which events will be summarized in the IST status bit. The command "*PRE 1" enables the INB bit to set the IST bit whenever it is set. Once parallel polling has been established, the parallel poll status is examined until a change on data bus line DI02 takes place.

Stage 1: Enable the INE and PRE registers, configure the controller for parallel poll and instruct the oscilloscope to respond on data line 2 (DI02)

```

CMD1$="?_@$"
CALL IBCMD(BRD0%,CMD1$)
CMD$="INE 1;*PRE 1"
CALL IBWRT(BRD0%,CMD$)
CMD4$=CHR(&H5)+CHR(&H69)+"?"
CALL IBCMD(BRD0%,CMD4$)
    
```

Stage 2: Parallel poll the instrument until DI02 is set

```

LOOP% = 1
WHILE LOOP%
    CALL IBRPP(BRD0%,PPR%)
    IF (PPR% AND &H2) = 2 THEN LOOP% = 0
WEND
    
```

Stage 3: Disable parallel polling (hex 15) and clear the parallel poll register

```

CMD5$=CHR$(&H15)
CALL IBCMD(BRD0%,CMD5$)
CALL IBCMD(BRD0%,CMD1$)
CMD$="*PRE 0"
CALL IBWRT(BRD0%,CMD$)
    
```

Note 1: In the example above, board-level GPIB function calls are used. It is assumed that the controller (board) and oscilloscope (device) are respectively located at addresses 0 and 4. The listener and talker addresses for the controller and oscilloscope are:

Logic device	Listener address	Talker address
controller	32 (ASCII<space>)	64 (ASCII @)
oscilloscope	32+4=36 (ASCII \$)	64+4=68 (ASCII D)

Note 2: The characters "?" and "_" appearing in the command strings stand for unlisten and untalk respectively. They are used to set the devices to a "known" state.

Note 3: To shorten the size of the program examples, device talking and listening initialization instructions have been grouped into character chains. They are:

```
CMD1$ = "?_@$" 'Unlisten, Untalk, PC talker, DSO listener
```

Note 4: The remote message code for executing a parallel response in binary form is 01101PPP where PPP specifies the data line. Since data line 2 is selected, the identification code is 001 which results in the code 01101001 (binary) or &H69 (hex). See Table 38 of the IEEE 488-1978 Standard for further details.

***IST Poll**

The state of the Individual Status bit (IST) returned in parallel polling can also be read by sending the "*IST?" query. To enable this poll mode, the oscilloscope must be initialized as for parallel polling by writing into the PRE register. Since *IST polling emulates parallel polling, this method is applicable in all instances where parallel polling is not supported by the controller.

In the following example, the command "INE 1" enables the event "new signal acquired" in the INR to be reported to the INB bit of the status byte STB. The command "*PRE 1" enables the INB bit to set the IST bit whenever it is set. The command "CHDR OFF" suppresses the command header in the response of the instrument, simplifying the interpretation. The status of the IST bit is then continuously monitored until it is set by the instrument.

```
CMD$="CHDR OFF; INE 1; *PRE 1"
CALL IBWRT(SCOPE%,CMD$)
LOOP% = 1
WHILE LOOP%
    CMD$="*IST?"
    CALL IBWRT(SCOPE%,CMD$)
    CALL IBRD(SCOPE%,RD$)
    IF VAL(RD$) = 1 THEN LOOP% = 0
WEND
```

DRIVING A HARD-COPY DEVICE

The oscilloscope can be interfaced to a wide range of plotters and printers and be instructed to directly plot or print the screen contents onto these devices. The devices supported by the unit are listed with the command "HARDCOPY_SETUP" in Section 5.

When the hard-copy device is connected to the GPIB two different configurations should be considered depending on whether or not a GPIB controller is available.

Plotting/Printing without a GPIB Controller

When only the oscilloscope and the hard-copy device are connected to the GPIB, the oscilloscope must be configured as talker-only and the hard-copy device as listener-only to ensure proper data transfer. The oscilloscope can be configured as a talker-only by using the thumbwheel switch at the rear of the instrument to select an address larger than 30. The hard-copy device manufacturer usually specifies an address which forces the instrument into the listening mode.

- Select the oscilloscope's address to be larger than 30.
- Switch on the oscilloscope.
- Configure the "Hardcopy" sub menu in the "Auxiliary Setups" menu specifying "GPIB" as hard copy port.
- Put the hard-copy device in listener-only mode.
- Press the screen dump button on the front panel of the instrument.

Plotting/Printing with a GPIB Controller

If a controller is connected to the GPIB, data transfers must be supervised by the controller. The oscilloscope must be set to an address between 0 and 30 which differs from the controller's and the hard-copy device's address. Different schemes can be used to transfer the screen contents:

1. The controller reads the data into internal memory and then sends them to the printer/plotter. This alternative can be done with simple high-level GPIB function calls.
2. The oscilloscope sends data to both the controller and the printer/plotter.
3. The controller goes into a standby state. The oscilloscope becomes a talker and sends data directly to the printer/plotter.

1. Data read by controller and sent to printer/plotter

The controller stores the full set of printer/plotter instructions and sends them afterwards to the graphics device. This method is the most straightforward way of transferring screen contents but it requires a large amount of buffer storage (110K for 4 traces).

```
CMD$ = "SCDP"  
CALL IBWRT(SCOPE%,CMD$)  
FILE$="PLOT.DAT"  
CALL IBRDF(SCOPE%,FILE$)  
CALL IBWRTF(PLOTTER%,FILE$)
```

2. Oscilloscope sends data to controller and printer/plotter

The oscilloscope puts the printer/plotter instructions on to the bus. The data is directly plotted out and saved in scratch memory in the controller. The contents of the scratch file can be deleted later on.

Stage 1: Controller talker, oscilloscope listener. Issue the screen dump command

```
CMD1$="?_@$": CALL IBCMD(BRD0%,CMD1$)  
CMD$="SCDP": CALL IBWRT(BRD0%,CMD$)
```

Stage 2: Oscilloscope talker, controller and plotter listeners. Plot data while storing data in scratch file SCRATCH.DAT

```
CMD2$="?_D%": CALL IBCMD(BRD0%,CMD2$)  
FILE$="SCRATCH.DAT": CALL IBRDF(BRD0%,FILE$)
```


3. Oscilloscope talks directly to plotter/printer

The controller goes into stand-by and resumes GPIB operations once the data have been plotted, that is when an EOI is detected.

Stage 1: Controller talker, oscilloscope listener. Issue the screen dump command

```
CMD1$="?_@$": CALL IBCMD(BRD0%,CMD1$)
CMD$="SCDP": CALL IBWRT(BRD0%,CMD$)
```

Stage 2: Oscilloscope talker, plotter listener. Put controller in stand-by

```
CMD2$="?_D%": CALL IBCMD(BRD0%,CMD2$)
V%=1: CALL IBGTS(BRD0%,V%)
```

Note 1: In schemes 2 and 3, board-level GPIB function calls are used. It is assumed that the controller (board), the oscilloscope and the plotter are respectively located at addresses 0, 4 and 5. The listener and talker addresses for the controller, oscilloscope and plotter are:

Logic device	Listener address	Talker address
controller	32 (ASCII<space>)	64 (ASCII @)
oscilloscope	32+4=36 (ASCII \$)	64+4=68 (ASCII D)
hard-copy dev.	32+5=37 (ASCII %)	64+5=69 (ASCII E)

Note 2: The characters "?" and "_" appearing in the command strings stand for unlisten and untalk respectively. They are used to set the devices to a "known" state.

Note 3: To shorten the size of the program examples, device talking and listening initialization instructions have been grouped into character chains. They are:

```
CMD1$ = "?_@$" 'Unlisten, Untalk, PC talker, DSO listener
CMD2$ = "?_D" 'Unlisten, Untalk, PC listener, DSO talker
```



INTRODUCTION

LeCroy oscilloscopes may be remotely controlled using a host, either a terminal or a computer, via the RS-232-C port. For this purpose the oscilloscope must be set at an address higher than 30 using the thumbwheel switch at the rear of the instrument.

All the commands described in Section 5 are supported but waveform transfer is only possible in HEX mode. The default value for COMM_FORMAT is set appropriately. The syntax of the response to WF? is identical to the GPIB case.

In this section some special RS-232-C commands are defined either for configuring the oscilloscope, or simulating GPIB 488.1 messages such as setting the oscilloscope into remote or local modes.

Notation

Throughout this section, characters which cannot be printed in ASCII will be represented by their mnemonics.

Example

<LF> is the ASCII line feed character whose decimal value is 10

<BS> is the ASCII backspace character whose decimal value is 8

CTRL_U means that the control key and the U key are pressed simultaneously.

RS-232-C
PIN ASSIGNMENTS

The remote RS-232-C pin assignments (indicated on the rear panel) are as follows:

Pin #		Description
2	T × D	Transmitted data (from the oscilloscope).
3	R × D	Received data (to the oscilloscope).
4	RTS	Request to send (from the oscilloscope). If the software Xon/Xoff handshake is selected it is always TRUE. Otherwise (hardware handshake) it is TRUE when the oscilloscope is able to receive characters and FALSE when the oscilloscope is unable to receive characters.
5	CTS	Clear to send (to the oscilloscope). When true, the oscilloscope can transmit, when false, transmission stops. It is used for the oscilloscope output hardware handshake.
20	DTR	Data terminal ready (from oscilloscope). Always TRUE.

- | | | |
|---|---------|--------------------|
| 1 | GND | Protective Ground. |
| 7 | SIG GND | Signal Ground. |

RS-232-C CONFIGURATION

The RS-232-C port is configured in full duplex. This means that the two sides (i.e. the controller and the oscilloscope) can both send and receive messages at the same time. However, when the oscilloscope receives a new command, it stops outputting.

Transmission of long messages to the oscilloscope should be done while the oscilloscope is in a triggered mode with no acquisition in progress. This is especially important when sending waveforms or front-panel setups into the oscilloscope.

The behavior of the RS-232-C port may be set according to the user's needs. For this purpose, in addition to the basic setup on the front-panel menu there are "immediate commands" as well as a special command "COMM_RS232". Immediate commands consist of the ASCII ESCape character <ESC> (whose decimal value is 27), followed by another character. Such commands are interpreted as soon as the second character has been received.

*Note: The RS-232-C baud rate, parity, character length and number of stop bits are among the parameters that are saved or recalled by the front-panel "SAVE" or "RECALL" button, or by the remote commands "*SAV", "*RCL" or "PANEL_SETUP". When recalling, care must be taken to ensure that these parameters are set at the same value as the actual ones. Otherwise, the host may no longer be able to communicate with the oscilloscope and a manual reconfiguration would be necessary.*

Echo of Received Characters by the Oscilloscope

The serial port may echo the received characters. Echo is useful if the oscilloscope is attached to a terminal. Echoing can be turned on or off by sending the two character sequence <ESC>] or <ESC>[respectively. Echoing is on by default.

Note: The host must not echo characters received from the oscilloscope.

Handshake Control

When the oscilloscope input buffer becomes almost full, the instrument sends a handshake signal to the host telling it to stop transmitting. When this buffer has enough room to receive more characters another handshake signal will be sent. The handshake signals are either the CTRL-S (or <XOFF>) and CTRL-Q (<XON>) characters or a signal level on the RTS line (pin 4). This is selected by sending the two-character sequence <ESC>) for XON/XOFF handshake - this is the default - or <ESC>(for RTS handshake.

The flow of characters coming from the oscilloscope may be controlled either by a signal level on the CTS line (pin 5) or by the <XON>/<XOFF> pair of characters.

Editing Features

When the oscilloscope is directly connected to a terminal, the following features will facilitate the correction of typing errors:

<BS> or <DELETE> Delete the last character.
CTRL_U Delete the last line.

Message Terminators

“Message terminators” are markers that indicate to the receiver that a message has been completed.

On input to the oscilloscope, the Program Message Terminator is one character which can be selected by the user. A good choice would be a character that is never used for anything else. The character is chosen using the command COMM_RS232 and the keyword EI. The default Program Message Terminator is the ASCII character <CR>, whose decimal value is 13.

The oscilloscope appends a Response Message Terminator to the end of each of its responses. It is a string, like a computer prompt, chosen by the user. This string must not be empty. The default Response Message Terminator is “\n\r” which means <LF><CR>.

Examples

(1) COMM_RS232 EI,3

This command informs the oscilloscope that each message it receives will be terminated with the ASCII character <ETX> which corresponds to 3 in decimal.

(2) COMM_RS232 EO,“\r\nEND\r\n”

This command indicates to the oscilloscope that it must append the string “\r\nEND\r\n” to each response.

After these settings, a host command will look like:

```
TDIV?<ETX>
```

The oscilloscope responds:

```
TDIV 1. S
END
```

Note: Having sent a COMM_RS232 command, the host must wait for the oscilloscope to change its behavior before sending a command in the new mode. A safe way to do this is to include a query on the line which contains the COMM_RS232 command and wait until the response is received. For example,

```
COMM_RS232 EI,3; *STB?
```

SRQ Message Each time the Master Summary Status (MSS) bit of the Status Byte (STB) is set, the SRQ message (a string of characters) is sent to the host to indicate that the oscilloscope requests service. The RS-232-C SRQ message has the same meaning as the GPIB SRQ message. If the string is empty, no message will be sent. This is the default setting. Note that no response message terminator is added at the end of the SRQ message.

Example COMM_RS232 SRQ,"`\r\n\nSRQ\r\n\a`"
When the MSS bit is set, the oscilloscope will send
a <CR> followed by 2 <LF>s
 SRQ
a <CR> followed by 1 <LF>
and the buzzer will sound.

Long Line Splitting Line splitting is a feature provided for hosts that cannot accept lines with more than a certain number of characters. The oscilloscope may be configured to split responses into many lines. This feature is very useful for waveform or front-panel setup transfers although it is applicable to all response messages. Two parameters control this feature:

Line Separator: Off messages will not be
 split into lines
 <CR>,<LF> or <CR><LF> possible line terminators.

Line length: the maximum number of characters in a line.

Example COMM_RS232 LS,LF,LL,40
The line separator is the ASCII character <LF>, the line is a maximum of 40 characters long (excluding the line separator).

If the oscilloscope receives the command PNSU?, it may answer:
PNSU #9000001496
AAAA5555000655AA4030005800190000000000001
000000000000000000000000000000C1B0100580000
00
...

Remarks Long commands sent to the oscilloscope may not be split into lines. If a command sent to the oscilloscope is the response to a previous query, the line split characters (<LF> and/or <CR>) must be removed.

This also applies to line split characters inside strings sent to the oscilloscope.

However, hex-ASCII data sent to the oscilloscope may contain line split characters. If you wish to use line splitting, ensure that neither the input message terminator characters nor the line split characters occur in the data.

COMMANDS SIMULATING GPIB COMMANDS

<ESC>C or <ESC>c

Device clear command.

This command clears the input and output buffers. It has the same meaning as the GPIB DCL or SDC interface messages.

<ESC>R or <ESC>r

Set to remote command
(REN)

This command puts the oscilloscope into the remote mode. Its function is the same as GPIB asserting the REN line and setting the oscilloscope to listener.

<ESC>L or <ESC>l

Set to local command

This command puts the oscilloscope into local mode. It clears local lockout. It has the same function as GPIB setting the REN line to false.

<ESC>F or <ESC>f

Set local lockout command

This command disables the front-panel "LOCAL" button either immediately if the oscilloscope is already in the remote mode or later when the oscilloscope is next set to remote control. This disabling of the front-panel "LOCAL" button is called "Local Lockout" and can only be cancelled with the <ESC>L command. <ESC>F has the same meaning as the GPIB LLO interface message.

<ESC>T or <ESC>t

Trigger command (GET)

This command rearms the oscilloscope while it is in "SINGLE" or in "SEQUENCE" mode (valid only while the oscilloscope is in the remote mode). It has the same meaning as the "*TRG" command, and also the same meaning as the GPIB GET interface message.



ORGANIZATION

This section of the manual lists all commands and queries recognized by the oscilloscope. For easy reference the listings are arranged in alphabetical order. Each command starts on a new page and the name (header) of the command is given in both the long and short forms. Below each name (header) it is indicated whether it denotes a command only, a command as well as a query, or a query only. For those headers that may be used to command an action, for example to modify a setup parameter, or to obtain some information such as the current value of a setup parameter, the query form is derived by appending a question mark (?) immediately to the header without intervening spaces.

The description of each command starts with a short explanation of the function performed by it, followed by a presentation of the formal syntax. In the formal syntax the header appears in mixed mode characters with the characters used to construct the short form shown in upper case.

Where applicable, the syntax of the query form is given along with the format of the response the oscilloscope will produce.

For most commands the description terminates with a short example illustrating a typical use of the command. The GPIB examples assume that the controller is equipped with a National Instruments interface board, and they show calls to the National Instruments interface subroutines in BASIC. The device name of the oscilloscope has been defined as "SCOPE%".

COMMAND SUMMARY

The following is an overview of the commands grouped according to their functionality.

Acquisition

To control the acquisition of waveforms:

ARM_ACQUISITION, AUTO_SETUP, BANDWIDTH_LIMIT, INTERLEAVED, SAMPLE_CLOCK, SEGMENTS, STOP, *TRG, WAIT.

To select vertical input parameters to capture waveforms:

ATTENUATION, COUPLING, OFFSET, VOLT_DIV.

To select time-base parameters to capture waveforms:

TIME_DIV, TRIG_DELAY.

To select trigger conditions to capture waveforms:

TRIG_COUPLING, TRIG_LEVEL, TRIG_MODE, TRIG_PATTERN, TRIG_SELECT, TRIG_SLOPE.

Communication

To set communication characteristics:

COMM_FORMAT, COMM_HEADER, COMM_HELP, COMM_ORDER, COMM_RS232, PERSIST_SETUP.

Cursor	To perform measurements: CURSOR_MEASURE, CURSOR_SET, CURSOR_VALUE?, PARAMETER_VALUE?, PER_CURSOR_SET, PER_CURSOR_VALUE, XY_CURSOR_ORIGIN, XY_CURSOR_SET, XY_CURSOR_VALUE?.
Display	To display waveforms: DISPLAY, DUAL_ZOOM, GRID, HOR_MAGNIFY, HOR_POSITION, INTENSITY, MULTI_ZOOM, PERSIST, SELECT, TRACE, VERT_MAGNIFY, VERT_POSITION, XY_ASSIGN, XY_DISPLAY, ZOOM. To display messages to a local user: CALL_HOST, KEY, MESSAGE.
Function	To perform mathematical operations on waveforms: DEFINE, FUNCTION_RESET, FUNCTION_STATE.
Hard Copy	To plot or print the contents of the display screen: HARDCOPY_SETUP, HARDCOPY_TRANSMIT, SCREEN_DUMP.
Save/Recall Setup	To preserve and restore front-panel settings: PANEL_SETUP, *RCL, RECALL_PANEL, *RST, *SAV, STORE_PANEL.
Status	To obtain status information and set up service requests: ALL_STATUS?, *CLS, CMR?, DDR?, *ESE, *ESR?, EXR?, INE, INR?, *IST?, *OPC, *PRE, *SRE, *STB?, URR?, *WAI.
Waveform Transfer	To preserve and restore waveforms: INSPECT?, RECALL, STORE, STORE_SETUP, STORE_TEMPLATE, TEMPLATE?, WAVEFORM, WAVEFORM_SETUP, WAVEFORM_TEXT.
Miscellaneous	To control the calibration and test the instrument: AUTO_CALIBRATE, *CAL?, *TST?. To control the built-in buzzer: BUZZER. To control the real-time clock: DATE.

To delete a file from the memory card:

DELETE_FILE

To obtain a directory listing of the memory card:

DIRECTORY_LIST?

To format the memory card:

FORMAT_CARD

To identify the instrument:

*IDN?, *OPT?.

COMMAND EXECUTION

Before attempting to execute a command or query, the oscilloscope scans it to verify its correctness and that sufficient information is given to perform the requested action. To protect the local user from changes in the oscilloscope's behavior which are beyond his control, the remote user must set the oscilloscope to the remote state to execute commands that affect the operation of the instrument as an oscilloscope. If such a command is received while the oscilloscope is operating in the local state, an execution permission error is generated and the execution of the command is denied. Vice versa, the local user cannot interfere with the remote user because all front-panel controls are disabled while the oscilloscope is in the remote state.

Since interrogating the oscilloscope does not change its internal state, it may be queried at any time, independently of local or remote operation. There are only two exceptions to this rule: the queries *CAL? and *TST? both recalibrate the oscilloscope and are therefore executed in the remote state only.

Commands that only affect the remote behavior are executed independently of whether the oscilloscope is in the local or remote state. In this category are all commands that modify communication parameters (COMM_FORMAT, COMM_HEADER, COMM_HELP, COMM_ORDER, COMM_RS232), all commands affecting status information (*CLS, *ESE, INE, *OPC, *PRE, *SRE, *WAI), and the commands used to display messages on the screen to the local user (CALL_HOST, KEY, MESSAGE).

In the description of each command, only exceptions to the rule that a command is executed only in the remote state and a query is executed in both the local and remote states are mentioned.

COMMAND NOTATION

The following notation is used in the description of the individual commands :

< > Angular brackets enclose words that are used as placeholders. There are two types of placeholders: (1) the header path, (2) a data parameter of a command.

- :=** A colon followed by an equals sign separates a placeholder from the description of the type and range of values that may be used in a command instead of the placeholder.
- { }** Braces enclose a list of choices from which one must be selected.
- []** Square brackets enclose optional items.
- ...** An ellipsis indicates that the items to the left and to the right of the ellipsis may be repeated zero or more times.

As an example, consider the syntax notation for the command to set the vertical input sensitivity:

```
<channel>:VOLT_DIV <v_gain>  
<channel> := {C1, C2}  
<v_gain> := 5.0 mV to 2.5 V
```

The first line shows the formal appearance of the command with `<channel>` denoting the placeholder for the header path, and `<v_gain>` denoting the placeholder for the data parameter specifying the desired vertical gain value. The second line indicates that either "C1" or "C2" must be chosen for the header path, and the third line explains that the actual vertical gain can be set to any value between 5 mV and 2.5 V.

STATUS**ALL_STATUS?, ALST?**

Query

DESCRIPTION

The ALL_STATUS? query reads and clears the contents of all status registers: STB, ESR, INR, DDR, CMR, EXR, and URR except the MAV bit (bit 6) of the STB register. For an interpretation of the contents of each register, refer to the appropriate status register.

The ALL_STATUS? query is useful if a complete overview of the state of the instrument is required.

QUERY SYNTAX

ALL_Status?

Response format

ALL_Status STB,<value>,ESR,<value>,INR,<value>,
DDR,<value>, CMR<value>,EXR,<value>,URR,<value>
<value> := 0 to 65535

EXAMPLE (GPIB)

The following instruction reads the contents of all the status registers.

```
CMD$="ALST?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
ALST STB,000000,ESR,000052,INR,000005,DDR,000000,
EXR,000024,CMR,000004,URR,000000
```

RELATED COMMANDS

*CLS, CMR?, DDR?, *ESR?, EXR?, *STB?, URR?

ACQUISITION

ARM_ACQUISITION, ARM

Command

DESCRIPTION

The ARM_ACQUISITION command enables the signal acquisition process by changing the acquisition state from "triggered" to "ready".

COMMAND SYNTAX

ARM_acquisition

EXAMPLE

The following command enables signal acquisition.
CMD\$ = "ARM": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

STOP, *TRG, TRIG_MODE, WAIT

ACQUISITION**ATTENUATION, ATTN**

Command/Query

DESCRIPTION

The ATTENUATION command selects the vertical attenuation factor of the probe. Values of 1, 10, 100, 1000 or 10000 may be specified.

The ATTENUATION? query returns the attenuation factor of the specified channel.

COMMAND SYNTAX

<channel>:ATTenuation <attenuation>

<channel> := {C1, C2, C3‡, C4‡}

<attenuation>:= {1, 10, 100, 1000, 10000}

QUERY SYNTAX

<channel>:ATTenuation?

Response format

<channel>:ATTenuation <attenuation>

EXAMPLE (GPIB)

The following command sets the attenuation factor of channel 1 to 100.

CMD\$="C1:ATTN 100": CALL IBWRT(SCOPE%,CMD\$)

‡ 4-channel oscilloscopes only

MISCELLANEOUS

AUTO_CALIBRATE, ACAL

Command/Query

DESCRIPTION

The AUTO_CALIBRATE command is used to enable or disable automatic calibration of the instrument. At power-up, auto-calibration is turned ON, i.e. all input channels are periodically calibrated for the current input amplifier and time-base settings.

The automatic calibration may be disabled by issuing the command ACAL OFF. Whenever it is convenient, a *CAL? query may be issued to fully calibrate the oscilloscope. When the oscilloscope is returned to local control, the periodic calibrations will be resumed.

The response to the AUTO_CALIBRATE? query indicates whether auto-calibration is enabled.

COMMAND SYNTAX

Auto_CALibrate <state>

<state> := {ON, OFF}

QUERY SYNTAX

Auto_CALibrate?

Response Format

Auto_CALibrate <state>

EXAMPLE (GPIB)

The following instruction disables auto-calibration.

```
CMD$="ACAL OFF": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

*CAL?

ACQUISITION**AUTO_SETUP, ASET**

Command

DESCRIPTION

The AUTO_SETUP attempts to display the input signal(s) by adjusting the vertical, time-base and trigger parameters. Auto-setup operates only on the channels whose traces are currently turned on. The only exception occurs when no traces are turned on, in which case AUTO_SETUP operates on all channels and turns on all of the traces.

If signals are detected on several channels, the lowest numbered channel with a signal determines the selection of the time base and trigger source.

If only one input channel is turned on, the time base will be adjusted for that channel.

COMMAND SYNTAX

Auto_SETUP

EXAMPLE

The following command instructs the oscilloscope to perform an auto-setup.

```
CMD$ = "ASET": CALL IBWRT(SCOPE%,CMD$)
```

ACQUISITION

BANDWIDTH_LIMIT, BWL

Command/Query

DESCRIPTION

The BANDWIDTH_LIMIT command enables or disables the bandwidth limiting low pass filter.

The response to the BANDWIDTH_LIMIT? query indicates if the bandwidth filter is on or off.

COMMAND SYNTAX

BandWidth_Limit <mode>

<mode> := {ON, OFF}

QUERY SYNTAX

BandWidth_Limit?

Response Format

BandWidth_Limit <mode>

EXAMPLE

The following command turns the bandwidth filter on.

```
CMD$ = "BWL ON": CALL IBWRT(SCOPE%,CMD$)
```

MISCELLANEOUS**BUZZER, BUZZ**

Command

DESCRIPTION

The BUZZER command controls the built-in piezo-electric buzzer. This may be useful to attract the attention of a local operator in an interactive working application. The buzzer may either be activated for short beeps (about 400 msec long in BEEP mode) or continuously for a certain time interval selected by the user by turning the buzzer ON or OFF. A beep request which immediately follows another beep request will be held off for approximately 200 msec.

Note: This command is always accepted (local and remote).

COMMAND SYNTAX

BUZZer := {BEEP, ON, OFF}

EXAMPLE (GPIB)

Sending the following code will cause the oscilloscope to sound two short tones.

```
CMD$ = "BUZZ BEEP; BUZZ BEEP";
```

```
CALL IBWRT(SCOPE%, CMD$)
```

MISCELLANEOUS

***CAL?**

Query

DESCRIPTION

The *CAL? query performs a complete internal calibration. This calibration sequence is the same as that which occurs at power-up. At the end of the calibration, the response indicates how the calibration terminated. When the calibration is finished, the instrument returns to the state it was in prior to the query.

Hardware failures are identified by a unique binary code in the returned <status> number (see Table 1). A "0" response indicates that no failures occurred.

Note: This query is only accepted in remote mode.

QUERY SYNTAX

*CAL?

Response Format

*CAL <diagnostics>

<diagnostics> := 0 calibration successful

BIT	BIT VALUE	DESCRIPTION
0	1	CH1 failure
1	2	CH2 failure
2	4	CH3 failure‡
3	8	CH4 failure‡
4	16	TDC failure
5	32	Trigger circuit failure

Failures
Table 1

EXAMPLE (GPIB)

The following instruction forces a self-calibration.

```
CMD$="*CAL?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RD$): PRINT RD$
```

Response message (if no failure)

```
*CAL 0
```

RELATED COMMANDS

AUTO_CALIBRATE

‡ 4-channel oscilloscopes only, reserved in the 2-channel oscilloscopes

DISPLAY**CALL_HOST, CHST**

Command/Query

DESCRIPTION

The CALL_HOST command allows the user to manually generate a service request (SRQ). Once the CALL_HOST command has been received, the message "Call Host" will be displayed next to the lowest button (10) in the menu field (II). Pressing this button while in the root menu causes the User Request status Register (URR) and the URQ bit of the Event Status Register to be set. This can generate a SRQ in local mode provided that the service request mechanism has been enabled.

The response to the CALL_HOST? query indicates whether call host is enabled (on) or disabled (off).

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

Call_HoST <state>
 <state> := {ON, OFF}

QUERY SYNTAX

Call_HoST?

Response format

Call_HoST <state>

EXAMPLE (GPIB)

After executing the following code an SRQ request will be generated whenever button 10 is pressed. It is assumed that SRQ servicing has already been enabled.

```
CMD$="CHST ON": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

URR?

STATUS

***CLS**
Command

DESCRIPTION

The *CLS command clears all the status data registers.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

***CLS**

EXAMPLE (GPIB)

The following command causes all the status data registers to be cleared.

```
CMD$ = "*CLS": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

ALL_STATUS?, CMR?, DDR?, *ESR?, EXR?, *STB?, URR?

STATUS**CMR?**

Query

DESCRIPTION

The CMR? query reads and clears the contents of the CoMmand error Register (CMR). The CMR register (Table 2) specifies the last syntax error type detected by the instrument.

QUERY SYNTAX

CMR?

Response format

CMR <value>

<value> := 1 to 13

EXAMPLE (GPIB)

The following instruction reads the contents of the CMR register.

```
CMD$="CMR?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

CMR 0

RELATED COMMANDS

ALL_STATUS?, *CLS

Value	Description
1	Unrecognized command/query header
2	Illegal header path
3	Illegal number
4	Illegal number suffix
5	Unrecognized keyword
6	String error
7	GET embedded in another message
10	Arbitrary data block expected
11	Non-digit character in byte count field of arbitrary data block
12	EOI detected during definite length data block transfer
13	Extra bytes detected during definite length data block transfer

Command Error Status Register Structure (CMR)

Table 2

COMMUNICATION

COMM_FORMAT, CFMT

Command/Query

DESCRIPTION

The COMM_FORMAT command selects the format which the oscilloscope will use to send waveform data. The available options allow (1) the block format, (2) the data type and (3) the encoding mode to be modified from the default settings.

Note: This command can be executed in both local and remote modes.

The COMM_FORMAT? query returns the currently selected waveform data format.

COMMAND SYNTAX

Comm_ForMaT <block_format>,<data_type>,<encoding>
<block_format> := {DEF9, IND0, OFF}
<data_type> := {BYTE, WORD}
<encoding> := {BIN, HEX}
(GPIB uses both encoding forms, RS-232-C always uses HEX)

Initial settings (i.e. after power on) are:

DEF9, WORD, BIN for GPIB
DEF9, WORD, HEX for RS-232-C

QUERY SYNTAX

Comm_ForMaT?

Response Format

Comm_ForMaT <block_format>,<data_type>,<encoding>

EXAMPLE (GPIB)

The following code redefines the transmission format of waveform data. The data will be transmitted as a block of indefinite length. Data will be coded in binary and represented as 8-bit integers.

```
CMD$="CFMT IND0,BYTE,BIN"  
CALL IBWRT(SCOPE%,CMD$)
```

EXPLANATION

BLOCK FORMAT

DEF9: Uses the IEEE 488.2 definite length arbitrary block response data format. The digit 9 indicates that the byte count consists of 9 digits. The data block directly follows the byte count field.

For example, a data block consisting of 3 data bytes would be sent as:

```
WF DAT1,#9000000003<DAB><DAB><DAB>
```

where <DAB> represents an 8-bit binary data byte.

IND0: Uses the IEEE 488.2 indefinite length arbitrary block response data format.

A <NL^END> (new line with EOI) signifies that block transmission has ended.

The same data bytes as above would be sent as:

WF DAT1,#0<DAB><DAB><DAB><NL^END>

OFF: Same as IND0. In addition, the data block type identifier and the leading #0 of the indefinite length block will be suppressed. The data presented above would be sent as:

WF <DAB><DAB><DAB><NL^END>

Note: The format OFF does not conform to the IEEE 488.2 standard and is only provided for special applications where the absolute minimum of data transfer may be important.

DATA TYPE

BYTE: Transmits the waveform data as 8-bit signed integers (1 byte).

WORD: Transmits the waveform data as 16-bit signed integers (2 bytes).

Note: The data type BYTE transmits only the high order bits of the internal 16-bit representation. The precision contained in the low order bits is lost.

ENCODING

BIN: Binary encoding (GPIB only)

HEX: Hexadecimal encoding (bytes are converted to 2 hexadecimal ASCII digits (0, ... 9, A, ... F))

RELATED COMMANDS

WAVEFORM?

COMMUNICATION

COMM_HEADER, CHDR

Command/Query

DESCRIPTION

The COMM_HEADER command controls the way the oscilloscope will format responses to queries. The instrument provides three response formats: (1) LONG format, i.e. responses start with the long form of the header word; (2) SHORT format, i.e. responses start with the short form of the header word; (3) OFF, i.e. headers are omitted from the response and suffix units in numbers are suppressed. Until the user requests otherwise, the SHORT response format is used.

This command does not affect the interpretation of messages sent to the oscilloscope. Headers may be sent in their long or short form regardless of the COMM_HEADER setting.

Querying the vertical sensitivity of Channel 1 may result in one of the following responses:

COMM_HEADER	RESPONSE
LONG	CHANNEL_1:VOLT_DIV 200E-3 V
SHORT	C1:VDIV 200E-3 V
OFF	200E-3

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

Comm_HeaDeR <mode>

<mode> := {SHORT, LONG, OFF}

Note: The default mode, i.e. the mode just after power on, is SHORT.

QUERY SYNTAX

Comm_HeaDeR?

Response Format

Comm_HeaDeR <mode>

EXAMPLE (GPIB)

The following code sets the response header format to short.
 CMD\$ = "CHDR SHORT"; CALL IBWRT(SCOPE%,CMD\$)

COMMUNICATION**COMM_HELP, CHLP**

Command/Query

DESCRIPTION

The COMM_HELP command enables the help diagnostics utility to assist remote program debugging. When turned on, this utility displays all message transactions occurring between the controller and the oscilloscope on a terminal, printer or similar recording device connected to the RS-232-C port. Errors detected by the instrument can be directly viewed.

Note: This command can be executed in both local and remote modes.

The COMM_HELP? query indicates if the diagnostics utility has been enabled.

COMMAND SYNTAX

Comm_HeLP <target>

<target> := {RS, OFF}

The initial <target>, (i.e. after power on) is OFF.

QUERY SYNTAX

Comm_HeLP?

Response Format

Comm_HeLP <target>

EXAMPLE (GPIB)

The following code turns on the remote control diagnostics utility.
CMD\$="CHLP RS"; CALL IBWRT(SCOPE%,CMD\$)

COMMUNICATION

COMM_ORDER, CORD

Command/Query

DESCRIPTION

The COMM_ORDER command controls the byte order of waveform data transfers. Waveform data may be sent with the most significant byte (MSB) or the least significant byte (LSB) in the first position. The default mode is the MSB first.

COMM_ORDER applies equally to the waveform's descriptor and time blocks. In the descriptor some values are 16 bits long ("word"), 32 bits long ("long" or "float"), or 64 bits long ("double"). In the time block all values are floating values, i.e. 32 bits long. When "COMM_ORDER HI" is selected the most significant byte is sent first. When "COMM_ORDER LO" is specified the least significant byte is sent first.

The COMM_ORDER? query returns the byte transmission order currently in use.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

Comm_ORDer <mode>

<mode> := {HI, LO}

Note: The initial mode, i.e. the mode after power on, is HI.

QUERY SYNTAX

Comm_ORDer?

Response Format

Comm_ORDer <mode>

EXAMPLE

The order of transmission of waveform data depends on the data type. Table 3 illustrates the different possibilities.

Type	CORD HI	CORD LO
Word	<MSB><LSB>	<LSB><MSB>
Long/float	<MSB><byte 2><byte 3><LSB>	<LSB><byte 3><byte 2><MSB>
Double	<MSB><byte 2> ... <byte 7><LSB>	<LSB><byte 7> ... <byte 2><MSB>

Waveform Data Transmission Order

Table 3

RELATED COMMANDS **WAVEFORM?**

COMMUNICATION

COMM_RS232, CORS

Command/Query

DESCRIPTION

The command COMM_RS232 sets the parameters of the RS-232-C port for remote control.

The COMM_RS232? query reports the settings of the parameters.

Note: This command is ONLY valid if the oscilloscope is being remotely controlled via the RS-232-C port.

The parameters are:

- a. DUPLEX behavior mode.
- b. End Input character. When received by the oscilloscope, this character will be interpreted as the END-of-a-command message marker. The commands received will be parsed and executed.
- c. End Output string. The oscilloscope will add this string at the end of a response message. When the host computer receives this string, it knows that the oscilloscope has completed its response.
- d. Line Length. This parameter defines the maximum number of characters that will be sent to the host in a single line. Remaining characters of the response will be output in separate additional lines. This parameter is only applicable if a line separator has been selected.
- e. Line Separator. This parameter is used to select the line splitting mechanism and to define the characters used to split the oscilloscope response messages into many lines. Possible line separators are: CR, LF, CRLF. A <CR>, a <LF> or a <CR> followed by a <LF> will be sent to the host computer after <line_length> characters.
- f. SRQ string. This string is sent each time the oscilloscope wants to signal an SRQ to the host computer.

Note: Some parameters of this command require ASCII strings as actual arguments. In order to facilitate the embedding of non-printable characters into such strings, escape sequences may be

used. the back-slash character ('\') is used as an escape character. The following escape sequences are recognized:

- "\a": Bell character,
- "\b": Back space character,
- "\e": Escape character,
- "\n": Line feed character,
- "\r": Carriage return character,
- "\t": Horizontal tab character,
- "\\": The back-slash character itself
- "\ddd": ddd represents one to three decimal digit characters giving the code value of the corresponding ASCII character. This allows any ASCII code in the range 1 to 127 to be inserted.

Before using the string, the oscilloscope will replace the escape sequence by the corresponding ASCII character.

For example, the escape sequences "\r", "\13" and "\013" are all replaced by the single ASCII character <Carriage Return>.

Notation			
DUPLEX	duplex	EI	End input character
EO	End output string	LL	Line length
LS	Line separator	SRQ	SRQ Service request

COMMAND SYNTAX

COmm_RS232 DUPLEX,<duplex>,EI,<ei_char>,
EO,'<eo_string>',LL,<line_length>,LS,<Line_sep>,
SRQ,'<srq_string>'

<duplex> := FULL (only full duplex is currently implemented)

<ei_char> := 1 to 126 (default: 13 = <CR>)

<eo_string> := A non-empty ASCII string of up to 20 characters.
(default: "\n\r")

<line_length> := 40 to 1024 (default: 256)

<line_sep> := {OFF, CR, LF, CRLF} (default: OFF)

<srq_string> := An ASCII string which may be empty.
(default: empty string)

QUERY SYNTAX

COmm_RS232?

Response Format

COmm_RS232 DUPLEX,<duplex>,EI,<ei_char>,
EO,"<eo_string>",
LL,<line_length>,LS,<line_sep>,SRQ,"<srq_string>"

EXAMPLE

After executing the command

```
COMM_RS232 EI,3,EO,"R\NEND\R\N"
```

the oscilloscope will assume that it has received a complete message each time the <ETX> (decimal value 3) is detected. Response messages will be terminated by sending the character sequence "<CR><LF>END<CR><LF>".

ACQUISITION**COUPLING, CPL**

Command/Query

DESCRIPTION

The **COUPLING** command selects the coupling mode of the specified input channel.

The **COUPLING?** query returns the coupling mode of the specified channel.

COMMAND SYNTAX

<channel>:CouPLing <coupling>

<channel> := {C1, C2, C3‡, C4‡}

<coupling> := {A1M, D1M, D50, GND}

QUERY SYNTAX

<channel>:CouPLing?

Response format

<channel>:CouPLing <coupling>

EXAMPLE GPIB)

The following command sets the coupling of Channel 2 to 50 Ω DC.

```
CMD$="C2:CPL D50": CALL IBWRT(SCOPE%,CMD$)
```

‡ 4-channel oscilloscopes only

CURSOR

CURSOR_MEASURE, CRMS

Command/Query

DESCRIPTION

The CURSOR_MEASURE command specifies the type of cursor to be displayed.

The CURSOR_MEASURE? query indicates which cursors are currently displayed.

Notation

HABS	Horizontal absolute	HREL	Horizontal relative
VABS	Vertical absolute	VREL	Vertical relative
PARAM	Parameters	OFF	Cursors off
PASS	Pass test	FAIL	Fail test
SHOW	Extended parameters display		

Note: The PARAM mode is turned OFF when the XY mode is ON.

COMMAND SYNTAX

CuRsor_MeaSure <mode>

<mode> := {HABS, VABS, HREL, VREL, PARAM, OFF, PASS, FAIL, SHOW}

QUERY SYNTAX

CuRsor_MeaSure?

Response Format

CuRsor_MeaSure <mode>

EXAMPLE (GPIB)

The following command switches on the vertical relative cursors.

```
CMD$="CRMS VREL": CALL IBWRT(SCOPE%,CMD$)
```

The following command determines which cursor is currently turned on.

```
CMD$="CRMS?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RD$): PRINT RD$
```

Example of response message
CRMS OFF

RELATED COMMANDS

CURSOR_MEASURE, CURSOR_SET,
PASS_FAIL_COUNTER, PASS_FAIL_DO,
PASS_FAIL_MASK, PARAMETER_VALUE?

CURSOR**CURSOR_SET, CRST**

Command/Query

DESCRIPTION

The **CURSOR_SET** command allows the user to position any one of the eight independent cursors at a given screen location. The positions of the cursors can be modified or queried even if the required cursor is not currently displayed on the screen.

When setting a cursor position, a trace must be specified, relative to which the cursor will be positioned.

The **CURSOR_SET?** query indicates the current position of the cursor(s). The values returned depend on the grid type selected.

Note 1: When the oscilloscope is in the dual grid mode, traces are assigned to either the upper grid (EA, MC, FE, C1, C3‡) or lower grid (EB, MD, FF, C2, C4‡). The trace specified determines whether a vertical cursor will be placed relative to the upper or lower grid.

In quad grid mode‡ each channel is permanently assigned to its respective grid with C1 at the top and C4 at the bottom. All other traces may be re-positioned anywhere on the screen using the vertical position control.

*Note 2: If the parameter display is turned on (or the pass/fail display or the extended parameters display), the parameters of the specified trace will be shown unless the newly chosen trace is not displayed or has been acquired in sequence mode (these conditions will produce an environment error, see Table 6, page 84). To only change the trace without repositioning the cursors, the **CURSOR_SET** command may be given with no argument, (e.g. **EB:CRST**).*

Notation

HABS	Horizontal absolute		
VABS	Vertical absolute		
HREF	Horizontal reference	HDIF	Horizontal difference
VREF	Vertical reference	VDIF	Vertical difference
PREF	Parameter reference	PDIF	Parameter difference

‡ 4-channel oscilloscopes only

COMMAND SYNTAX

`<trace>:CuRsor_SeT <cursor>,<position>[<cursor>,<position>,
...<cursor>,<position>]`

`<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}`

`<cursor> := {HABS, VABS, HREF, HDIF, VREF, VDIF,
PREF, PDIF}`

`<position> := 0 to 10 DIV (horizontal)
-13 to 13 DIV (vertical)`

Note 1: The suffix DIV is optional.

Note 2: Parameters are grouped in pairs. The first parameter specifies the cursor to be modified and the second one indicates its new value. Parameters may be grouped in any order and may be restricted to those items to be changed.

QUERY SYNTAX

`<trace>:CuRsor_SeT? [<cursor>,...<cursor>]`

`<cursor>:= {HABS, VABS, HREF, HDIF, VREF, VDIF, PREF,
PDIF, ALL}`

Response Format

`<trace>:CuRsor_SeT <cursor>,<position>[<cursor>,<position>,
...<cursor>,<position>]`

If `<cursor>` is not specified, ALL will be assumed. If the position of a cursor cannot be determined in a particular situation, its position will be indicated as UNDEF.

EXAMPLE (GPIB)

The following command positions the VREF and VDIF cursors at +3 DIV and -7 DIV respectively, using Function E as a reference.

```
CMD$="FE:CRST VREF,3DIV,VDIF,-7DIV"  
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

CURSOR_MEASURE, CURSOR VALUE?,
PASS_FAIL_COUNTER, PASS_FAIL_DO,
PASS_FAIL_MASK, PARAMETER_VALUE?

‡ 4-channel oscilloscopes only

CURSOR**CURSOR_VALUE?, CRVA?**

Query

DESCRIPTION

The CURSOR_VALUE? query returns the values measured by the specified cursors for a given trace. (The PARAMETER_VALUE? query is used to obtain measured waveform parameter values.)

Notation

HABS	Horizontal absolute	HREL	Horizontal relative
VABS	Vertical absolute	VREL	Vertical relative

QUERY SYNTAX

```
<trace>:CuRsor_VAlue? [<mode>,...<mode>]
```

```
<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}
```

```
<mode> := {HABS, VABS, HREL, VREL, ALL}
```

Response Format

```
<trace>:CuRsor_VAlue? <mode>[,<hor_value>],<ver_value>  
[,...,<mode>[,<hor_value>],<ver_value>]
```

For horizontal cursors, both horizontal as well as vertical values are given, whereas for vertical cursors only vertical values are given.

Note: If <mode> is not specified or equals ALL, all the measured cursor values for the specified trace are returned. If the value of a cursor could not be determined in the current environment, the value UNDEF will be returned.

EXAMPLE (GPIB)

The following query reads the measured absolute horizontal value of the cross-hair cursor (HABS) on Channel 2.

```
CMD$="C2:CRVA? HABS": CALL IBWRT(SCOPE%,CMD$)
```

```
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
C2:CRVA HABS,34.2 US,244 MV
```

RELATED COMMANDS

CURSOR_SET

‡ 4-channel oscilloscopes only

MISCELLANEOUS

DATE

Command/Query

DESCRIPTION

The DATE command changes the date/time of the oscilloscope's internal real time clock.

The DATE? query returns the current date/time setting.

COMMAND SYNTAX

DATE <day>,<month>,<year>,<hour>,<minute>,<second>

<day> := 1 to 31

<month> := {JAN, FEB, MAR, APR, MAY, JUN, JUL,
AUG, SEP, OCT, NOV, DEC}

<year> := 1987 to 2500

<hour> := 0 to 23

<minute> := 0 to 59

<second> := 0 to 59

Note: It is not always necessary to specify all the DATE parameters. Only the parameters up to and including the parameter to be changed need to be specified, i.e. to change the "year" setting specify day, month and year together with the required settings. The time settings will remain unchanged. To change the "second" setting all the DATE parameters must be specified with the required settings.

QUERY SYNTAX

DATE?

Response format

DATE <day>,<month>,<year>,<hour>,<minute>,<second>

EXAMPLE (GPIB)

This example will change the date to October 1, 88 and the time to 1:21:16 p.m. (13:21:16 in 24 hour notation).

```
CMD$="DATE 1,OCT,1988,13,21,16"  
CALL IBWRT(SCOPE%,CMD$)
```

<i>STATUS</i>	DDR? Query
DESCRIPTION	The DDR? query reads and clears the contents of the Device Dependent or device specific error Register (DDR). In the case of a hardware failure, the DDR register specifies the origin of the failure. Refer to Table 4, page 66, for further details.
QUERY SYNTAX	DDR?
Response format	DDR <value> <value> := 0 to 65535
EXAMPLE (GPIB)	The following instruction reads the contents of the DDR register. CMD\$="DDR?": CALL IBWRT(SCOPE%,CMD\$) CALL IBRD(SCOPE%,RSP\$): PRINT RSP\$ Response message DDR 0
RELATED COMMANDS	ALL_STATUS?, *CLS

Bit	Bit Value	Description
15..14		0 Reserved
13	8192	1 time-base hardware failure is detected
12	4096	1 a trigger hardware failure is detected
11	2048	1 a Channel 4‡ hardware failure is detected.
10	1024	1 a Channel 3‡ hardware failure is detected.
9	512	1 a Channel 2 hardware failure is detected
8	256	1 a Channel 1 hardware failure is detected
7..4		0 Reserved
3	8	1 a Channel 4‡ overload condition is detected
2	4	1 a Channel 3‡ overload condition is detected ¹
1	2	1 a Channel 2 overload condition is detected
0	1	1 a Channel 1 overload condition is detected

Device Specific Register Structure (DDR)

Table 4

‡ 4-channel oscilloscopes only, reserved in the 2-channel oscilloscopes

FUNCTION**DEFINE, DEF**

Command/Query

Standard Oscilloscopes**DESCRIPTION**

The DEFINE command specifies the mathematical expression to be evaluated by a function.

Notation

MAXPTS	maximum number of points		
SWEEPS	maximum number of sweeps		
EQN	equation	AVGS	average summed
C1	Channel 1	C2	Channel 2
C3	Channel 3‡	C4	Channel 4‡

COMMAND SYNTAX

<function>:DEFine EQN,'<equation>',MAXPTS,<max_points>,
SWEEPS,<max_sweeps>

<function> := {MC‡, MD‡, FE, FF}

<equation> := {<source>, - <source>, <source> + <source>,
<source> - <source>, AVGS (<source>)}

<source> := {C1, C2, C3‡, C4‡}

<max_points> :=	9410/14	All others
	50 to 10000	50 to 50000

<max_sweeps> := 1 to 1000

Note 1: Parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be given in any order and may be restricted to the variables to be changed.

Note 2: The pair SWEEPS,<max_sweeps> applies only when averaging (AVGS) has been chosen. Otherwise it is ignored.

QUERY SYNTAX

<function>:DEFine?

Response format

<function>:DEFine EQN,'<equation>',MAXPTS,<max_points>
[,SWEEPS,<max_sweeps>]

‡ 4-channel oscilloscopes only

5 System Commands

EXAMPLE (GPIB)

The following command defines Function E (FE) to compute the summed average of Channel 1 using 5000 points over 200 sweeps.

```
CMD$="FE:DEF EQN,'AVGS(C1)',  
MAXPTS, 5000,SWEEPS,200"  
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

FUNCTION_RESET, FUNCTION_STATE‡, INR?

‡ 4-channel oscilloscopes only

FUNCTION**DEFINE, DEF**

Command/Query

Oscilloscopes fitted with the WP01 Option

DESCRIPTION

An oscilloscope fitted with the Waveform Processing option (WP01) accepts additional forms of the DEFINE command:

Processing Notation

ABS	Absolute Value
AVGC	Continuous Average
AVGS	Summed Average
DERI	Derivative
EXP	Exponential (power of e)
EXP10	Exponential (power of 10)
EXTR	Extrema
FLOOR	Floor (Extrema only)
ERES	Enhanced Resolution Filter
INTG	Integral
LOG10	Logarithm base 10
LN	Logarithm base e
ROOF	Roof (Extrema only)
SQR	Square
SQRT	Square Root
+	Identity or Add
-	Negation or Subtract
*	Multiply
/	Ratio
1/	Reciprocal

Key words

BITS	Resolution enhancement bits (Enhanced Resolution only)
DITHER	Dither (Summed Average only)
MAXPTS	Maximum number of points
REJECT	Reject overflow/underflow (Summed Average only)
SWEEPS	Maximum number of sweeps (Average and Extrema only)
WEIGHT	Weight (Continuous Average only)

COMMAND SYNTAX

```

<function>:DEFine EQN,'<equation>', MAXPTS,<max_points>,
SWEEPS,<max_sweeps>, DITHER,<off_on>,
REJECT,<off_on>, WEIGHT,<weight>, BITS,<bits>
<function> := {MC‡, MD‡, FE, FF}
<equation> := AVGS(<source>)           Summed Average
<equation> := AVGC(<source>)           Continuous Average
<equation> := <paren_source_expr>       Identity
<equation> := +<paren_source_expr>     Identity
<equation> := -<paren_source_expr>     Negation
<equation> := 1/<paren_source_expr>     Reciprocal
<equation> := <paren_source_expr> + <source> Addition
<equation> := <paren_source_expr> - <source> Subtraction
<equation> := <paren_source_expr> * <source> Multiplication
<equation> := <paren_source_expr> / <source> Ratio
<equation> := EXTR(<source>)           Extrema (R+F)

```

‡ 4-channel oscilloscopes only

<equation> := FLOOR(EXTR(<source>)) Floor
 <equation> := ROOF(EXTR(<source>)) Roof
 <equation> := SQR(<source_expr>) Square
 <equation> := SQRT(<source_expr>) Square Root
 <equation> := LN(<source_expr>) Logarithm base e
 <equation> := LOG10(<source_expr>) Logarithm base 10
 <equation> := EXP(<source_expr>) Power of e
 <equation> := EXP10(<source_expr>) Power of 10
 <equation> := INTG(<source_expr>) Integral
 <equation> := DERI(<source_expr>) Derivative
 <equation> := ABS(<source_expr>) Absolute Value
 <equation> := ERES(<source>) Enhanced Resolution
 <paren_source_expr> := (<source_expr>)
 <paren_source_expr> := <source>
 <source_expr> := <multiplier> * <source> {+, -} <addend>
 <source_expr> := <multiplier> * <source>
 <source_expr> := <source> {+, -} <addend>
 <source_expr> := <source>
 <multiplier> := 0.001e-33 to 999.999e33
 <addend> := - 999.999e33 to 999.999e33
 <source> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<max_points> :=	9410/14	All others
	50 to 10000	50 to 50000

<max_sweeps> := 1 to 1000000
 <off_on> := {OFF, ON}
 <weight> := {1, 3, 7, 15, 31, 63, 127, 255, 511, 1023}
 <bits> := {0.5, 1.0, 1.5, 2.0, 2.5, 3.0}

Note: Space (blank) characters inside equations are optional.

‡ 4-channel oscilloscopes only

QUERY SYNTAX	<function>:DEFINE?
Response format	<function>:DEFine EQN,'<equation>', MAXPTS,<max_points>, SWEEPS,<max_sweeps>, DITHER,<off_on>, REJECT,<off_on>, WEIGHT,<weight>, BITS,<bits>
EXAMPLE (GPIB):	The following command defines Function E to compute the product of (Channel 1 multiplied by 2.1 and augmented by 3.3) and Channel 2, using a maximum of 10000 input points: CMD\$="FE:DEF EQN,'(2.1*C1+3.3)*C2', MAXPTS,10000" CALL IBWRT(SCOPE%,CMD\$)
RELATED COMMANDS	FUNCTION_RESET, FUNCTION_STATE‡, INR?

‡ 4-channel oscilloscopes only

FUNCTION**DEFINE, DEF**

Command/Query

Oscilloscopes fitted with the WP02 Option**DESCRIPTION**

An oscilloscope fitted with the FFT option (WP02) accepts additional forms of the DEFINE command.

Notation

WINDOW	FFT window function
FFT	Fast Fourier Transform (complex result)
REAL	Real part of complex result
IMAG	Imaginary part of complex result
MAG	Magnitude of complex result
PHASE	Phase angle (degrees) of complex result
PS	Power Spectrum
PSD	Power Density
AVGP	Power Average
RECT	Rectangular window
HANN	von Hann window
HAMM	Hamming window
FLTP	Flat Top window
BLHA	Blackman-Harris window
DCSUP	DC component suppression

COMMAND SYNTAX (FFT) <function> :DEFine EQN,'<equation>',
 MAXPTS,<max_points>,WINDOW,<window_type>,
 DCSUP,<off_on>

<function> := {MC‡, MD‡, FE, FF}
 <equation> := FFT(<source_expr>)

<equation> := REAL(FFT(<source_expr>))
 <equation> := IMAG(FFT(<source_expr>))
 <equation> := MAG(FFT(<source_expr>))
 <equation> := PHASE(FFT(<source_expr>))
 <equation> := PS(FFT(<source_expr>))
 <equation> := PSD(FFT(<source_expr>))

<source_expr> := <multiplier> * <source> {+, -} <addend>
 <source_expr> := <multiplier> * <source>

‡ 4-channel oscilloscopes only

<source_expr> := <source> {+, -} <addend>
 <source_expr> := <source>
 <multiplier> := 0.001e-33 to 999.999e33
 <addend> := -999.999e33 to 999.999e33
 <source> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}
 <window_type> := {RECT, HANN, HAMM, FLTP, BLHA}
 <off_on> := {OFF, ON}

	9410/14	All others
<max_points> :=	50 to 10000	50 to 50000

Note: The source waveform must be a time domain signal.

QUERY SYNTAX

<function>:DEFine?

Response Format

<function>:DEFine EQN,'<equation>',MAXPTS,<max_points>, WINDOW,<window_type>,DCSUP,<off_on>

EXAMPLE (GPIB)

The following command defines Function E to compute the Power Spectrum of the FFT of Channel 1. Prior to FFT computation, Channel 1 is multiplied by 1.018 and 0.055 (units of Channel 1, i.e. Volts) is added. A maximum of 1000 points will be used for the input. The window function is Rectangular. The DC component of the input is not suppressed.

```

CMD$="FE:DEF EQN,'PS(FFT(1.018*C1 + 0.055))',
MAXPTS,1000,WINDOW,RECT,DCSUP,OFF"
CALL IBWRT(SCOPE%,CMD$)
  
```

COMMAND SYNTAX

(FFT Power Average)

<function>:DEFine EQN,'<equation>',
 SWEEPS,<max_sweeps>
 <equation> := MAG(AVGP(<source>))
 <equation> := PS(AVGP(<source>))
 <equation> := PSD(AVGP(<source>))
 <source> := {MC‡, MD‡, FE, FF}
 <max_sweeps> := 1 to 50000

Note: The source waveform must be another function defined as a Fourier transform.

‡ 4-channel oscilloscopes only

QUERY SYNTAX	<function>:DEFine?
Response Format	<function>:DEFine EQN,'<equation>',SWEEPS,<max_sweeps>
EXAMPLE (GPIB)	<p>The following command defines Function F to compute the Power Spectrum of the Power Average of the FFT being computed by the Function E, over a maximum of 244 sweeps.</p> <pre>CMD\$="FF:DEF EQN,'PS(AVGP(FE))',SWEEPS,244" CALL IBWRT(SCOPE%,CMD\$)</pre>
RELATED COMMANDS	FUNCTION_RESET, FUNCTION_STATE‡, INR?

‡ 4-channel oscilloscopes only

MISCELLANEOUS

DELETE_FILE, DELF

Command

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The DELETE_FILE command deletes all waveforms, only "Auto-stored" waveforms, or only a single file from the memory card.

Notation	
WF	all waveforms
AUTOWF	all "autostored" waveforms

COMMAND SYNTAX

DELEte_file FILE, <mode>

<mode> := {WF, AUTOWF, '<filename>'}

<filename> := an alphanumeric string of up to 8 characters, followed by a dot and an extension of up to 3 digits.

EXAMPLE (GPIB)

The following commands first delete a front-panel setup, and then delete all the "autostored" waveforms from the card:

```
CMD$="DELF FILE, 'P001.PNL'; DELF FILE, AUTOWF"  
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

DIRECTORY_LIST, FORMAT_CARD

MISCELLANEOUS**DIRECTORY_LIST?, DIR?**

Query

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The DIRECTORY_LIST? query produces a directory listing of the memory card. The response consists of a double-quoted string containing a DOS-like listing of the directory. If no memory card is present, or if it is not formatted, the string will be empty.

QUERY SYNTAX

DIRectory_list?

Response format

DIRectory_list? "<directory>"

<directory> := a variable length string detailing the file content of the memory card.

EXAMPLE (GPIB)

The following code asks for a listing of the directory of the memory card.

```
CMD$="DIR?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD (SCOPE%,RSP$): PRINT RSP$
```

Response message

DIR "

Memory card of 25-JAN-1991 12:10:40

1 SC1 001 29-JAN-1991 16:33:06 20361

2 SC2 001 29-JAN-1991 16:34:32 20361

2 File(s) 84992 bytes free

"

DISPLAY

DISPLAY, DISP

Command/Query

DESCRIPTION

The DISPLAY command controls the display screen of the oscilloscope. When the user is remotely controlling the oscilloscope and does not need to use the display, it may be useful to switch off the display via the DISPLAY OFF command. This improves instrument response time since the waveform graphic generation procedure is suppressed.

The response to the DISPLAY? query indicates the display state of the oscilloscope.

Note: When the display has been set to OFF, the real time clock and the message field are updated. However, the waveforms and associated texts remain unchanged.

COMMAND SYNTAX

DISPlay <state>

<state> := {ON, OFF}

QUERY SYNTAX

DISPlay?

Response Format

DISPlay <state>

EXAMPLE (GPIB)

The following instruction turns off the display generation.

```
CMD$="DISP OFF": CALL IBWRT(SCOPE%,CMD$)
```

*DISPLAY***DUAL_ZOOM, DZOM**

Command/Query

DESCRIPTION

By setting DUAL_ZOOM ON, the horizontal magnification and positioning controls apply to all expanded traces simultaneously. This command is useful if the contents of all expanded traces are to be examined at the same time.

The DUAL_ZOOM? query indicates whether multiple zoom is enabled or not.

Note: This command has the same effect as MULTI_ZOOM.

COMMAND SYNTAX

Dual_ZOoM <mode>
<mode> := {ON, OFF}

QUERY SYNTAX

Dual_ZOoM?

Response format

Dual_ZOoM <mode>

EXAMPLE (GPIB)

The following example turns dual zoom on.

CMD\$="DZOM ON": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

HOR_MAGNIFY, HOR_POSITION, MULTI_ZOOM, ZOOM

STATUS

***ESE**

Command/Query

DESCRIPTION

The *ESE command sets the standard Event Status Enable register (ESE). This command allows one or more events in the ESR register to be reflected in the ESB summary message bit (bit 5) of the STB register. For an overview of the ESB defined events refer to the ESR table (Table 5, page 82).

The *ESE? query reads the contents of the ESE register.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

*ESE <value>

<value> := 0 to 255

QUERY SYNTAX

*ESE?

Response format

*ESE <value>

EXAMPLE (GPIB)

The following command allows the ESB bit to be set if a user request (URQ bit 6, i.e. decimal 64) and/or a device dependent error (DDE bit 3, i.e. decimal 8) occurs. Summing these values yields the ESE register mask $64+8=72$.

CMD\$="*ESE 72": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

*ESR?

STATUS***ESR?**

Query

DESCRIPTION

The *ESR? query reads and clears the contents of the Event Status Register (ESR). The response represents the sum of the binary values of the register bits 0 to 7. Refer to Table 5, page 82 for an overview of the ESR register structure.

QUERY SYNTAX***ESR?****Response format*****ESR <value>**

<value> := 0 to 255

EXAMPLE (GPIB)

The following instruction reads and clears the contents of the ESR register.

```
CMD$="*ESR?": CALL IBWRT(SCOPE%,CMD$)  
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

ESR 0*RELATED COMMANDS****ALL_STATUS?, *CLS, *ESE**

Bit	Bit Value	Bit Name	Description	Note
15..8			0 Reserved by IEEE 488.2	
7	128	PON	1 a Power off-to-ON transition has occurred	(1)
6	64	URQ	1 a User ReQuest has been issued	(2)
5	32	CME	1 a CoMmand parser Error has been found	(3)
4	16	EXE	1 an Execution Error has been detected	(4)
3	8	DDE	1 a Device Specific Error has occurred	(5)
2	4	QYE	1 a QueRy Error has occurred	(6)
1	2	RQC	0 The Instrument never requests bus control	(7)
0	1	OPC	0 The OPeration Complete bit is not used	(8)

Standard Event Status Register (ESR)

Table 5

Notes:

- (1) The Power On (PON) bit is always turned on (1) when the unit is powered up.
- (2) The User Request (URQ) bit is set true (1) when a soft key is pressed. An associated register URR identifies which key was selected. For further details refer to the URR? query.
- (3) The CoMmand parser Error bit (CME) is set true (1) whenever a command syntax error is detected. The CME bit has an associated CoMmand parser Register (CMR) which specifies the error code. Refer to the query CMR? for further details.
- (4) The EXecution Error bit (EXE) is set true (1) when a command cannot be executed due to some device condition (e.g. oscilloscope in local state) or a semantic error. The EXE bit has an associated Execution Error Register (EXR) which specifies the error code. Refer to query EXR? for further details.
- (5) The Device specific Error (DDE) is set true (1) whenever a hardware failure has occurred at power-up or execution time such as a channel overload condition, a trigger or a time-base circuit defect. The origin of the failure may be localized via the DDR? or the self test *TST? query.
- (6) The Query Error bit (QYE) is set true (1) whenever (a) an attempt is being made to read data from the Output Queue when no output is either present or pending, (b) data in the Output Queue has been lost, (c) both output and input buffers are full (deadlock state), (d) an attempt is made by the controller to read before having sent an <END>, (e) a command is received before the response to the previous query was read (output buffer flushed).
- (7) The ReQuest Control bit (RQC) is always false (0) since the oscilloscope has no GPIB controlling capability.
- (8) The OPeration Complete bit (OPC) is set true (1) whenever *OPC has been received since commands and queries are strictly executed in sequential order. The oscilloscope starts processing a command only once the previous command has been entirely executed.

STATUS**EXR?**

Query

DESCRIPTION

The EXR? query reads and clears the contents of the EXecution error Register (EXR). The EXR register specifies the type of the last error detected during execution. Refer to Table 6, page 84 for further details.

QUERY SYNTAX

EXR?

Response format

EXR <value>

<value> := 21 to 64

EXAMPLE (GPIB)

The following instruction reads the contents of the EXR register.

```
CMD$="*EXR?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message (if no fault)
EXR 0

RELATED COMMANDS

ALL_STATUS?, *CLS

Value	Description
21	Permission error. The command cannot be executed in local mode.
22	Environment error. The instrument is not configured to correctly process a command. For instance, the oscilloscope cannot be set to RIS at a slow time base.
23	Option error. The command applies to an option which has not been installed.
24	Unresolved parsing error.
25	Parameter error. Too many parameters specified.
26	Non-implemented command.
30	Hex data error. A non-hexadecimal character has been detected in a hex data block.
31	Waveform error. The amount of data received does not correspond to descriptor indicators.
32	Waveform descriptor error. An invalid waveform descriptor has been detected.
33	Waveform time error. Invalid RIS or TRIG time data has been detected.
34	Waveform data error. Invalid waveform data have been detected.
35	Panel setup error. An invalid panel setup data block has been detected.
50	No memory card present when user attempted to access the card.*
51	Memory card not formatted when user attempted to access the card.*
52	Memory card was exchanged when user attempted to RECALL the NEXT, PREVIOUS or SAME waveform.*
53	Memory card was write protected when user attempted to create a file, to delete a file, or to format the card.*
54	Bad memory card detected during formatting.*
55	Memory card root directory full. LECROY_1.DIR subdirectory cannot be created.*
56	Memory card full when user attempted to write to it.*
57	Memory card file sequence numbers exhausted (999 reached).*
58	Memory card file not found.*
59	Attempt to retrieve a file from memory card that is neither a waveform nor a panel.*
60	Memory card file is write protected (DOS Read/Only attribute).*
61	Memory card filename not DOS compatible, or waveform filename begins with an "A".*
62	A retrieved panel or waveform contains invalid bytes, or a waveform is too long for the instrument.*
64	Read/Write outside of memory card cluster chains. Can indicate a corrupted FAT table. Can also indicate a waveform or panel file with a length shorter than the internal byte count.*

Execution Error Status Register Structure (EXR)

Table 6

* only oscilloscopes equipped with the memory card option

MISCELLANEOUS**FORMAT_CARD, FCRD**

Command/Query

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The **FORMAT_CARD** command formats the memory card according to the PCMIA/JEIDA standard with a DOS partition.

The **FORMAT_CARD?** query returns the status of the card.

COMMAND SYNTAX

Format_CaRD

QUERY SYNTAX

Format_CaRD?

Response format

Format_CaRD <card_status>[,<read/write>,<free_space>,<card_size>,<battery_status>]

<card_status> := {NC, BAD, BLANK, UNKNOWN_FMT, DIR_MISSING, OK}

<read/write> := {WP, RW}

<free_space> := a decimal number giving the number of bytes still available on the card.

<card_size> := a decimal number giving the total number of bytes on the card.

<battery_status> := {BAT_OK, BAT_LOW, BAT_BAD}

Notation

NC	No card.
BAD	Bad card after formatting.
BLANK	Empty card.
UNKNOWN_FMT	Valid PCMIA format, but not supported.
DIR_MISSING	No "LECROY 1_DIR" subdirectory present. It will be automatically created with the next "store" command.
OK	The card is correctly formatted.
WP	Write protected.
RW	Read/Write authorized.
BAT_OK	The battery is in order.
BAT_LOW	The battery should be replaced.
BAT_BAD	Bad battery or no battery.

EXAMPLE (GPIB)

The following code will first format a memory card and then verify its status.

```
CMD$="FCRD": CALL IBWRT(SCOPE%,CMD$)
CMD$="FCRD?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
FCRD OK,RW,130048,131072,BAT_OK
```

RELATED COMMANDS

DIRECTORY_LIST

FUNCTION**FUNCTION_RESET,† FRST**

Command

DESCRIPTION

The FUNCTION_RESET command resets a waveform processing function. The number of sweeps will be reset to zero and the process restarted.

COMMAND SYNTAX

<function>:Function_ReSeT

EXAMPLE (GPIB)

<function> := {MC†, MD†, FE, FF}

Assuming that Function E (FE) has been defined as the summed average of Channel 1, the following example will restart the averaging process.

```
CMD$="FE:FRST": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

DEFINE, INR?

† 4-channel oscilloscopes only

FUNCTION

FUNCTION_STATE, FSTA‡

Command/Query

DESCRIPTION

The FUNCTION_STATE command allows the user to control or enquire how Functions C, D, E and F are being used. The four waveform processing functions may assume up to three different states:

- MEM static memory of a waveform (no further automatic processing occurs)
- ZOOM expansion of another waveform (updated as the source changes)
- FUNC a mathematical function of one or two other waveforms (updated if one of the sources change)

The two Functions C and D may assume all three states whereas E and F may assume only the states MEM and FUNC. The setup information needed to execute expansions or mathematical waveform processing is memorized separately by the oscilloscope for each function. When the state of a function is changed, the last setup information associated with the new state will be reactivated.

There are three other commands which may cause a state transition. The command ZOOM applied to Functions C or D will automatically switch them into zoom state. The commands STORE (storage from internal waveform) and WAVEFORM (storage from external waveform) applied to any one of the Functions C, D, E or F will automatically switch them into the memory state. There is never an automatic transition into the function state. The command FUNCTION_STATE must be used.

Initially Functions C and D are set to the memory state and Functions E and F are set to the function state.

The query FUNCTION_STATE returns the current state of a waveform processing function.

COMMAND SYNTAX

<function>:Function_State <state>

<function> := {MC, MD, FE, FF}

<state> := {FUNC, MEM, ZOOM}

‡ 4-channel oscilloscopes only

QUERY SYNTAX	<function>:Function_STAt?
Response format	<function>:Function_STAt <state>
EXAMPLE (GPIB)	The following example switches the internal function memory C into the mathematical waveform processing state thereby re-establishing the last valid waveform processing definition. CMD\$="MC:FSTA FUNC": CALL IBWRT(SCOPE%,CMD\$)
RELATED COMMANDS	DEFINE, STORE, WAVEFORM, ZOOM

DISPLAY

GRID

Command/Query

DESCRIPTION

The GRID command specifies whether the grid should be displayed in single, dual or quad‡ mode.

In single grid mode all the traces are displayed on a single grid.

In dual grid mode the screen is split into two distinct grids to separate the traces. In the dual-channel oscilloscopes, Channel 1 is always displayed in the upper grid and Channel 2 in the lower grid. In the 4-channel oscilloscopes, Channels 1 and 2 are always displayed in the upper grid and Channels 3 and 4 in the lower grid. All other waveforms can be vertically positioned anywhere.

In quad grid mode‡ Channel 1 is always displayed in the upper grid, Channel 2 in the second grid, etc. All other waveforms can be vertically positioned anywhere.

The GRID? query returns the grid mode currently in use.

COMMAND SYNTAX

GRID <grid>

<grid> := {SINGLE, DUAL, QUAD‡}

QUERY SYNTAX

GRID?

Response Format

GRID <grid>

EXAMPLE (GPIB)

The following command sets the screen display to dual grid mode.

```
CMD$="GRID DUAL": CALL IBWRT(SCOPE%,CMD$)
```

‡ 4-channel oscilloscopes only

HARD COPY**HARDCOPY_SETUP, HCSU**

Command/Query

DESCRIPTION

The HARDCOPY_SETUP command configures the instrument's hard copy driver. The command enables the user to specify the device type, transmission mode, plot size etc. of the hard-copy unit connected to the oscilloscope.

The command allows one or more individual settings to be changed by specifying the appropriate keyword(s) together with the new value(s). For instance, to select the Graphtec FP5301 plotter with normal speed, the command may be restricted to:

HCSU DEV,FP5301,SPEED,N

Notation

DEV	device	PORT	port
SPEED	plot speed	DENS	print density
PENS	plot pens	PFEED	page feed
PSIZE	paper size	GRID	grid square
LLX	lower left X	LLY	lower left Y
FP5301	Graphtec FP5301	PM8151	Philips PM8151
HP7470A	HP 7470A	HP7550A	HP 7550A
HPQJ	HP QuietJet	HPTJ	HP ThinkJet
HPLJ	HP LaserJet	EPSON	Epson FX80
N	Normal	NS	Non-standard
L	Low		

COMMAND SYNTAX

HardCopy_SetUp DEV,<device>, PORT,<port>
 ,SPEED,<plot_speed>,
 DENS,<print_density>, PENS,<plot_pens>,
 PFEED,<page_feed>, PSIZE,<paper_size>,
 GRID,<grid_square>, LLX,<lower_left_X>,
 LLY,<lower_left_Y>

Note: Parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be given in any order and may be restricted to those variables to be changed.

<device> := {FP5301, PM8151, HP7470A, HP7550A, HPQJ,
 HPTJ, HPLJ, EPSON}
 <port> := {GPIB, RS}
 <plot_speed> := {N, L} for plotters only
 <density> := {SINGLE, DOUBLE, QUADRUPLE,
 HIGH_SPEED, HIGH_RESOLUTION,
 ONE_TO_ONE, TWO_TO_ONE, CRT}
 for printers only
 <plot_pens> := 1 to 8
 <page_feed> := {ON, OFF}
 <paper_size> := {A5, A4, A3, NS}
 <grid_square> := 0.0 to 99.9 MM }
 <lower_left_X> := -999 to 999 MM } for non-standard
 <lower_left_Y> := -999 to 999 MM } paper size only

*Note: For these three parameter values the suffix is optional.
 The suffix M is assumed.*

QUERY SYNTAX

HardCopy_SetUp?

Response format

HardCopy_SetUp DEV,<device>,PORT,<port>,
 SPEED,<plot_speed>,DENS,<print_density>,PENS,<plot_pens>,
 PFEED,<page_feed>,PSIZE,<paper_size>,GRID,<grid_square>,
 LLX,<lower_left_X>,LLY,<lower_left_Y>

EXAMPLE (GPIB)

This example selects a HP 7550A plotter to be driven by the GPIB port.

```
CMD$="HCSU PORT,GPIB,DEV,HP7550A"
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

HARDCOPY_TRANSMIT, SCREEN_DUMP

HARD COPY**HARDCOPY_TRANSMIT, HCTR**

Command

DESCRIPTION

The HARDCOPY_TRANSMIT command sends a string of ASCII characters without modification to the hard-copy unit. This allows the user to control the hard-copy unit by sending device specific control character sequences. It also allows the user to place additional text on a screen dump for documentation purposes. This command accepts the escape sequence "\ddd" like those described under the command COMM_RS232 (see page 56). Before sending the string to the hard-copy unit the escape sequence is converted to the ASCII character code.

COMMAND SYNTAX

HardCopy_TRANSMIT '<string>'
<string> := Any sequence of ASCII or escaped characters.

EXAMPLE (GPIB)

The following code sends documentation data to a printer.

```
CMD$="HCTR 'Data from Oct.15\r\n'"  
CALL IBWRT(SCOPE%,CMD$)
```

The following code sends the same documentation data to an HP7470A plotter using pen 1. The text will be printed at the lower left corner of the paper.

```
CMD$=  
"HCTR 'IN;SP1;PA0,0;PD;LBData from Oct.15 \03IN;SP0;PA0,0'"  
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

HARDCOPY_SETUP, SCREEN_DUMP

DISPLAY

HOR_MAGNIFY, HMAG

Command/Query

DESCRIPTION

The HOR_MAGNIFY command horizontally expands the selected expansion trace by a specified factor. Magnification factors which are not within the range of permissible values will be rounded to the closest legal value.

If multiple zoom is enabled, the magnification factor for all expansion traces is set to the specified factor. If the specified factor is too large for any of the expanded traces (depending on their current source), it is reduced to an acceptable value and only then applied to the traces.

The VAB bit (bit 2) in the STB register (Table 8, page 137) is set if a factor outside the legal range is specified.

The HOR_MAGNIFY query returns the current magnification factor for the specified expansion function.

COMMAND SYNTAX

<exp_trace>:Hor_MAGnify <factor>

<exp_trace> := {EA, EB, MC‡, MD‡}

<factor> :=

9410/14	All others
1 to 200	1 to 1000

QUERY SYNTAX

<exp_source>:Hor_MAGnify?

Response Format

<exp_source>:Hor_MAGnify <factor>

EXAMPLE (GPIB)

The following example horizontally magnifies Expand B (EB) by a factor of 5.

```
CMD$="EB:HMAG 5": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

DUAL_ZOOM, ZOOM

‡ 4-channel oscilloscopes only

*DISPLAY***HOR_POSITION, HPOS**

Command/Query

DESCRIPTION

The HOR_POSITION command horizontally positions the geometric center of the intensified zone on the source trace. Allowed positions range from division 0 through 10. If the source trace was acquired in sequence mode, horizontal shifting will only apply to a single segment at a time.

If the multiple zoom is enabled, the difference between the specified and the current horizontal position of the specified trace is applied to all expanded traces. If this would cause the horizontal position of any expanded trace to go outside the left or right screen boundaries, the difference of positions is adapted and then applied to the traces.

If the sources of expanded traces are sequence waveforms, and the multiple zoom is enabled, the difference between the specified and the current segment of the specified trace is applied to all expanded traces. If this would cause the segment of any expanded trace to go outside of the range of the number of segments of sources, the difference is adapted and then applied to the traces.

The VAB bit (bit 2) in the STB register (Table 8, page 137) is set if a value outside the legal range is specified.

The HOR_POSITION query returns the position of the geometric center of the intensified zone on the source trace.

COMMAND SYNTAX

<exp_trace>:Hor_POSition <hor_position>,<segment>

<exp_trace> := {EA, EB, MC‡, MD‡}

<hor_position> := 0 to 10 DIV

<segment> := 1 to 200

Note 1: The suffix DIV is optional.

Note 2: The segment number is only relevant for waveforms acquired in sequence mode. The segment number is ignored in single waveform acquisitions.

‡ 4-channel oscilloscopes only

5 System Commands

QUERY SYNTAX <exp_trace>:Hor_POSition?

Response Format <exp_trace>:Hor_POSition <hor_position>,[<segment>]

Note: The segment number is only given for sequence waveforms.

EXAMPLE (GPIB)

The following example positions the center of the intensified zone on the trace currently viewed by Expand A (EA) at division 3.

CMD\$="EA:HPOS 3": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS DUAL_ZOOM, ZOOM

MISCELLANEOUS***IDN?**

Query

DESCRIPTION

The *IDN? query is used for identification purposes. The response consists of four different fields providing information on the manufacturer, the scope model, the serial number and the firmware revision level.

QUERY SYNTAX***IDN?****Response format*****IDN LECROY,<model>,<serial_number>,<firmware_level>****<model>** := 5-character model identifier**<serial_number>** := an 8-digit decimal code (94xxxxxx)**<firmware_level>** := 2 digits giving the release level followed by a period and a 1-digit update level (xx.y)**EXAMPLE (GPIB)**

This example issues an identification request to the scope.

```
CMD$="*IDN?": CALL IBWRT(SCOPE%,CMD$)
```

```
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
*IDN LECROY,9450_,94501153,02.2
```

STATUS

INE
Command/Query

DESCRIPTION

The INE command sets the Internal state change Enable register (INE). This command allows one or more events in the INR register to be reflected in the INB summary message bit (bit 0) of the STB register. For an overview of the INR defined events refer to Table 7, page 99.

The INE? query reads the contents of the INE register.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

INE <value>

<value> := 0 to 65 535

QUERY SYNTAX

INE?

Response format

INE <value>

EXAMPLE (GPIB)

The following command allows the INB bit to be set whenever a screen dump has finished (bit 1, i.e. decimal 2) and/or a waveform has been acquired (bit 0, i.e. decimal 1). Summing these two values yields the INE mask 2+1=3.

```
CMD$="INE 3": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

INR?

STATUS**INR?**

Query

DESCRIPTION

The INR? query reads and clears the contents of the Internal state change Register (INR). The INR register (Table 7) keeps track of the completion of various internal operations and state transitions.

QUERY SYNTAX

INR?

Response format

INR <state>
 <state> := 0 to 65535

EXAMPLE (GPIB)

The following instruction reads the contents of the INR register.
 CMD\$="INR?": CALL IBWRT(SCOPE%,CMD\$)
 Response message: INR 1026
 i.e. waveform processing in Function E and a screen dump have both terminated.

RELATED COMMANDS

ALL_STATUS?, *CLS, INE

Bit	Bit Value	Description
15...12		0 Reserved for future use.
11	2048	1 Waveform processing has terminated in Function F.
10	1024	1 Waveform processing has terminated in Function E.
9	512	1 Waveform processing has terminated in Memory D‡.
8	256	1 Waveform processing has terminated in Memory C‡.
7	128	1 A memory card exchange has been detected.*
6	64	1 Memory card has become full in "Autostore Fill" mode.*
5	32	1 Autostore to memory card cannot begin. Check for errors in the EXecution error Register.*
4	16	1 A segment of a sequence waveform has been acquired.
3	8	1 A time-out has occurred in a data block transfer.
2	4	1 A return to the local state is detected.
1	2	1 A screen dump has terminated.
0	1	1 A new signal has been acquired.

Internal State Register Structure (INR)

Table 7

‡ 4-channel oscilloscopes only, reserved in the 2-channel oscilloscopes

* only oscilloscopes equipped with the memory card option

WAVEFORM TRANSFER

INSPECT?, INSP?

Query

DESCRIPTION

The INSPECT? query allows the user to read parts of an acquired waveform in intelligible form. The command is based on the explanation of the format of a waveform given by the template (use the query TEMPLATE? to obtain an up-to-date copy). Each logical block of a waveform may be inspected by giving its name (e.g. TRIGTIME as mentioned in the template) enclosed in quotes as the first (string) parameter.

The special logical block named WAVEDESC may also be inspected in more detail. By giving the name of a variable in the block WAVEDESC enclosed in quotes as the first (string) parameter, it is possible to inspect only the actual value of that variable.

Notation

BYTE: raw data as integers (truncated to 8 m.s.b.*)
 WORD: raw data as integers (truncated to 16 m.s.b.*)
 FLOAT: normalized data (gain, offset applied) as floating point numbers (gives measured values in volts or appropriate units)

* most significant bits

QUERY SYNTAX

<trace>:INSpect? '<string>'[,<data_type>]

<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<string> := a valid name of a logical block or a valid name of a variable contained in block WAVEDESC (see the command TEMPLATE and Section 6).

<data_type> := {BYTE, WORD, FLOAT}

Note: The optional parameter <data_type> applies only for inspecting the data arrays. It selects the representation of the data. The default <data type> is FLOAT.

Response format

<trace>:INSpect "<string>"

<string> := a string giving name(s) and value(s) of a logical block or a variable.

‡ 4-channel oscilloscopes only

EXAMPLES (GPIB)

- 1) The following command reads the value of the time base at which the last waveform in Channel 1 was acquired.

```
CMD$ := "C1:INSP? 'TIMEBASE'"  
CALL IBWRT(SCOPE%,CMD$)  
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
C1:INSP "TIMEBASE: 500 US/DIV"
```

- 2) The following command reads the entire contents of the waveform descriptor block.

```
CMD$ = "C1:INSP? 'WAVEDESC'"
```

RELATED COMMANDS

TEMPLATE, WAVEFORM_SETUP

DISPLAY

INTENSITY, INTS

Command/Query

DESCRIPTION

The INTENSITY command sets the intensity level of the grid or the trace/text provided the local control of the intensity has been turned off. Note that normally the screen intensity is still under manual control when the oscilloscope operates remotely. The local intensity control has to be turned off (using the command INTS LOCAL,OFF) before the intensity levels can be modified remotely.

The intensity level is expressed as a percentage (PCT). A level of 100 PCT corresponds to the maximum intensity while a level of 0 PCT sets the intensity to its minimum value.

The response to the INTENSITY? query indicates the grid and trace intensity levels and their control mode.

COMMAND SYNTAX

INTenSity LOCAL,<mode>,GRID,<value>,TRACE,<value>

<mode> := {ON, OFF}

<value> := 0 to 100 PCT

Note 1: Parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be given in any order and may be restricted to those variables to be changed.

Note 2: The suffix PCT is optional.

QUERY SYNTAX

INTenSity?

Response Format

INTenSity LOCAL,<mode>,TRACE,<value>,GRID,<value>

EXAMPLE (GPIB)

The following instruction enables remote control of the intensity and changes the grid intensity level to 75%.

```
CMD$="INTS LOCAL,OFF,GRID,75"  
CALL IBWRT(SCOPE%,CMD$)
```

The following instruction re-enables local control of the intensity.

```
CMD$="INTS LOCAL,ON"  
CALL IBWRT(SCOPE%,CMD$)
```

ACQUISITION**INTERLEAVED, ILVD**

Command/Query

DESCRIPTION

The INTERLEAVED command enables or disables random interleaved sampling (RIS). An environment error (see Table 6, page 82) will be generated if the user attempts to turn off RIS when the time-base setting allows "ON" only, or to turn on RIS when the time-base setting allows "OFF" only (see table below).

RIS	9410/14	9420/24/30	9450
"ON" only	2 ns/div to 20 ns/div	1 ns/div [‡] to 20 ns/div	1 ns/div to 5 ns/div
Selectable "ON"/"OFF"	50 ns/div to 5 μs/div	50 ns/div to 20 μs/div	10 ns/div to 5 μs/div
"OFF" only	10 μs/div to 1000 s/div	50 μs/div to 5000 s/div	10 μs/div to 5000 s/div

RIS is not available for sequence mode acquisitions and therefore an attempt to turn it on in that mode will also result in an environment error.

The response to the INTERLEAVED? query indicates whether the oscilloscope is in the RIS mode.

COMMAND SYNTAX

InterLeaVeD <mode>
<mode> := {ON, OFF}

QUERY SYNTAX

InterLeaVeD?

Response Format

InterLeaVeD <mode>

EXAMPLE

The following command sets the oscilloscope into RIS mode.

CMD\$ = "ILVD ON": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

TIME_DIV, TRIG_MODE

[‡] 2 ns/div for Model 9430

STATUS

***IST?**

Query

DESCRIPTION

The *IST? (Individual S**T**atus) query reads the current state of the IEEE 488.1 defined "ist" local message. The "ist" individual status message is the status bit sent during a parallel poll operation.

QUERY SYNTAX

*IST?

Response format

*IST <value>

<value> := 0 or 1

EXAMPLE (GPIB)

The following command reads the contents of the IST bit:

```
CMD$ = "*IST?": CALL IBWRT(SCOPE%,CMD$)  
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

*IST 0

RELATED COMMANDS

*PRE


DISPLAY**KEY**

Command

DESCRIPTION

The KEY command is used to display a string in the menu field next to one of the 9 menu buttons. The string may consist of up to 12 characters and may be positioned at 4 different locations: above (L1), opposite (LC), or below (L2) the menu buttons; or (LB) between pairs of buttons, ([1,2], [3,4], [5,6] or [7,8]). See Figure 1 on next page.

Note: The button names shown in Figure 1 are not the same as those indicated in the front-panel figure at the beginning of the Operator's Manual.



Text assigned to the menu buttons will disappear on the next transition to local but reappear when the instrument is switched back into the remote state. Text is cleared at power up, when the rear-

panel RESET button (64) is pressed or if an empty string is assigned to a location (e.g. KEY " ",L1).

Pressing any one of the menu buttons while in remote mode causes the User Request status Register (URR) and the URQ bit of the Event Status Register to be set. This can generate an SRQ provided that the service request mechanism has been enabled.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

KEY <button>,'<string>',<position>

<button> := 1 to 9

<string> := a 12 character string (any ASCII code)

<position> := {L1, LC, L2, LB}

Note: If the position is omitted, LC will be assumed.

EXAMPLE (GPIB)

The example menu shown in Figure 1 was created by issuing the following series of KEY commands:

```
KEY 1, 'Pause',LC; KEY 1,'MEASUREMENT',LB; KEY 2,
'Continue',LC;
KEY 3,'key 3 line 1',LI; KEY 3,'key 3 line 2',L2;
KEY 4,'key 4 line 1',LI; KEY 4,'key 4 line 2',L2;
KEY 5,'key 5 center',LC; KEY 5,'between 5/6',LB; KEY 6,
'key 6 center',LC;
KEY 9,'Exit',LC
```

RELATED COMMANDS

URR?

DISPLAY**MESSAGE, MSG**

Command/Query

DESCRIPTION

The MESSAGE command displays a string of characters in the Message Field above the grid. The string may be up to 45 characters in length. The string is displayed as long as the instrument is in remote mode and no internal status message is generated. Turning the oscilloscope back to local mode deletes the message. After the next transition from local to remote the message will be redisplayed. The message is cleared at power up when the RESET button on the rear panel is pressed or if an empty string is sent (MSG " ").

The MESSAGE? query allows the user to read the last message which was sent.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

MeSsaGe '<string>'

<string> := a string of max. 45 characters

QUERY SYNTAX

MeSsaGe?

Response Format

MeSsaGe "<string>"

EXAMPLE (GPIB)

The following code causes the message "**Connect Probe 1**" to appear in the message field.

```
CMD$="MSG '**Connect Probe 1**"  
CALL IBWRT(SCOPE%,CMD$)
```

DISPLAY

MULTI_ZOOM, MZOM

Command/Query

DESCRIPTION

By setting MULTI_ZOOM ON, the horizontal magnification and positioning controls apply to all expanded traces simultaneously. This command is useful if the contents of all expanded traces are to be examined at the same time.

The MULTI_ZOOM? query indicates whether multiple zoom is enabled or not.

Note: This command has the same effect as DUAL_ZOOM.

COMMAND SYNTAX

Multi_ZOoM <mode>
<mode> := {ON, OFF}

QUERY SYNTAX

Multi_ZOoM?

Response format

Multi_ZOoM <mode>

EXAMPLE (GPIB)

The following example turns the multiple zoom on.
CMD\$="MZOM ON": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

HOR_MAGNIFY, HOR_POSITION, DUAL_ZOOM, ZOOM

ACQUISITION**OFFSET, OFST**

Command/Query

DESCRIPTION

The OFFSET command allows the vertical offset of the specified input channel to be adjusted.

The maximum ranges depend on the fixed sensitivity setting as follows:

9410/14/30		
Fixed Sensitivity	Offset Range	Voltage
2 V	± 6 times	± 12 V
1 V to 10 mV	± 12 times	± 12 V to ± 120 mV
5 mV	± 24 times	± 120 mV
2 mV	± 60 times	± 120 mV
1 mV	± 120 times	± 120 mV

All Others		
Fixed Sensitivity	Offset Range	Voltage
1 V	± 10 times	± 10 V
0.5 V to 20 mV	± 12 times	± 6 V to ± 240 mV
10 mV	± 24 times	± 240 mV
5 mV	± 48 times	± 240 mV

If an out-of-range value is entered, the oscilloscope is set to the closest possible value and the VAB bit (bit 2) in the STB register is set.

Note: The probe attenuation factor is not taken into account for adjusting the offset.

The OFFSET? query returns the DC offset value of the specified channel.

COMMAND SYNTAX

<channel>:OFFseT <offset>

<channel> := {C1, C2, C3‡, C4‡}

<offset> := -12V to 12V (maximum range)

Note: The suffix V is optional.

QUERY SYNTAX

<channel>:OFFseT?

Response format

<channel>:OFFseT <offset>

EXAMPLE (GPIB)

The following command sets the offset of Channel 2 to -3 V.
CMD\$="C2:OFST -3V": CALL IBWRT(SCOPE%,CMD\$)

‡ 4-channel oscilloscopes only

STATUS***OPC**

Command/Query

DESCRIPTION

The *OPC (OPeration Complete) command sets the OPC bit (bit 0) in the standard Event Status Register (ESR) to true. This command has no other effect on the operation of the oscilloscope as the instrument starts parsing a command or query only after it has completely processed the previous command or query.

The *OPC? query always responds with the ASCII character "1" as the oscilloscope responds to the query only once the previous command has been entirely executed.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX***OPC****QUERY SYNTAX*****OPC?****Response format*****OPC 1****RELATED COMMANDS*****WAI**

MISCELLANEOUS

***OPT?**

Query

DESCRIPTION

The *OPT? query identifies oscilloscope options, i.e. additional firmware or hardware options. The response consists of a series of response fields listing all the installed options.

QUERY SYNTAX

*OPT?

Response format

*OPT <option_1>,<option_2>,...,<option_N>

<option_i> := character data

Note: If no option is present, the character 0 will be returned.

EXAMPLE (GPIB)

This example queries the installed options.

```
CMD$="*OPT?": CALL IBWRT(SCOPE%,CMD$)
```

```
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
*OPT 0
```

If the waveform processing options WP01 and WP02 are installed, the response message is

```
*OPT WP01,WP02
```

SAVE/RECALL SETUP**PANEL_SETUP, PNSU**

Command/Query

DESCRIPTION

The PANEL_SETUP command complements the *SAV/*RST commands. The PANEL_SETUP command allows panel setups to be archived in encoded form on external storage media.

Only setup data read by the PNSU? query may be recalled into the oscilloscope. A panel setup error (see Table 6, page 84) will be generated if the setup data block contains invalid data.

Note: The communication parameters (those modified by commands CFMT, CHDR, CHLP, CORD and WFSU) and the enable registers associated with the status reporting system (SRE, PRE, ESE, INE) are not saved by this command.

COMMAND SYNTAX

PaNel_SetUp <setup>

<setup> := A setup block previously read by PNSU?

QUERY SYNTAX

PaNel_SetUp?

Response syntax

PaNel_SetUp <setup>

EXAMPLE (GPIB)

1. The following instruction saves the instrument's current panel setup in the file PANEL.SET.

```
FILE$ = "PANEL.SET": CMD$ = "PNSU?"
CALL IBWRT(SCOPE%,CMD$)
CALL IBRDF(SCOPE%,FILE$)
```

2. The following command recalls the front-panel setup stored previously in the file PANEL.SET into the oscilloscope.

```
CALL IBWRTF(SCOPE%,FILE$)
```

RELATED COMMANDS

*RCL, *SAV

CURSOR

PARAMETER_VALUE?, PAVA?

Query

DESCRIPTION

The PARAMETER_VALUE? query returns the current value(s) of the pulse waveform parameter(s) for the specified trace. Traces do not need to be displayed or selected to obtain the values measured by the pulse parameters.

Pulse parameters cannot be evaluated on waveforms composed of segments acquired in sequence mode. However pulse parameters may be applied to individual segments if they are singled out using the expansion function.

Parameter Names

ALL	all parameters	ALLI	all points in (TRUE=1, FALSE=0)
AMPLI	amplitude	ANYO	any points (TRUE=1, FALSE=0)
AREA	area	BASE	base
CYCL	cycles	DLY	delay
DUTY	duty cycle	FALL	falltime
FREQ	frequency	FRST	first point
LAST	last point	MAX	maximum
MEAN	mean	MIN	minimum
OVSN	negative overshoot	OVSP	positive overshoot
PER	period	PNTS	points
PKPK	peak-to-peak	RISE	risetime
RMS	root mean square	SDEV	standard deviation
TOP	top	WID	width

See Table 4 in Chapter 7 of Operator's Manual

Parameter Computation States

OK	deemed to be determined without problem
AV	averaged over several (up to 100) periods
PT	window has been period truncated
IV	invalid value (insufficient data provided)
NP	no pulse waveform
LT	less than given value
OF	signal partially in overflow
UF	signal partially in underflow
OU	signal partially in overflow and underflow

See Figure 15 in Chapter 5 of Operator's Manual

QUERY SYNTAX	<pre><trace>:PParameter_VAlue? [<parameter>, ... <parameter>] <trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡} <parameter> := {FRST, LAST, PNTS, MIN, MAX, MEAN, SDEV, RMS, DLY, PER, WID, RISE, FALL, ALL}</pre>
Response Format	<pre><trace>:PParameter_VAlue <parameter>,<value>,<state>[,... ,<parameter>,<value>,<state>] <value> := decimal numeric value <state> := {OK, AV, PT, IV, NP, LT, OF, UF, OU}</pre> <p><i>Note: If <parameter> is not specified, or equal to ALL, all the parameters followed by their values and states are returned.</i></p>
EXAMPLE (GPIB)	<p>The following query reads the risetime of Expand B (EB).</p> <pre>CMD\$="EB:PAVA? RISE": CALL IBWRT(SCOPE%,CMD\$) CALL IBRD (SCOPE%,RD\$): PRINT RD\$</pre> <p>Response message EB:PAVA RISE,3.6E-9S,OK</p>
RELATED COMMANDS	<pre>CURSOR_MEASURE, CURSOR_SET, PASS_FAIL_COUNTER, PASS_FAIL_DO, PASS_FAIL_MASK</pre>

‡ 4-channel oscilloscopes only

MISCELLANEOUS

PASS_FAIL_CONDITION, PFCO

Command/Query

DESCRIPTION

The PASS_FAIL_CONDITION command adds a Pass/Fail test condition or an extended parameter at a specified position.

The PASS_FAIL_CONDITION? query indicates the current Pass/Fail test setup or the current selection of extended parameters at a specified position.

Note: Up to four test conditions (or extended parameters) can be specified at four different positions on the screen. The PASS_FAIL_CONDITION command deals with one position at a time.

Notation			
GT	Greater than	LT	Lower than

COMMAND SYNTAX

Pass_Fail_Condition [<pos>,<trace>,<param>[,<rel_op>[,<ref_value>]]]

<pos> := {1,2,3,4}

<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3, C4}

<eparam> := extended parameters; (see list on next page)

<rel_op> := {GT,LT}

<ref_value> := 0.001e-33 to 999.999e33

Note: The PFCO command with no arguments (i.e. "PFCO") deletes all conditions.

QUERY SYNTAX

PFCO? <pos>

Response format

PFCO <pos>[,<trace>,<param>[,<rel_op>[,<ref_value>]]]

EXAMPLE (GPIB)

The following command sets the first test condition in the list to be: "frequency on Channel 1 lower than 10 kHz".

```
CMD$ = "PFCO 1,C1,FREQ,LT,10000";
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

CURSOR_MEASURE, CURSOR_SET,
PASS_FAIL_COUNTER, PASS_FAIL_DO,
PASS_FAIL_MASK, PARAMETER_VALUE

MISCELLANEOUS**PASS_FAIL_COUNTER, PFCT**

Command/Query

DESCRIPTION	The PASS_FAIL_COUNTER command resets the Passed/Failed acquisitions counter. The PASS_FAIL_COUNTER? query returns the current count.
COMMAND SYNTAX	Pass_Fail_CounTer
QUERY SYNTAX	Pass_Fail_CounTer?
Response Format	Pass_Fail_CounTer <pass_fail>,<value>,OF,<value> <pass_fail> := {PASS, FAIL} <value> := 0 to 999999
RELATED COMMANDS	CURSOR_MEASURE, CURSOR_SET, PASS_FAIL_DO, PASS_FAIL_MASK, PARAMETER_VALUE

MISCELLANEOUS

PASS_FAIL_DO, PFDO

Command/Query

DESCRIPTION

The PASS_FAIL_DO command defines the actions that have to be performed by the oscilloscope after a Pass/Fail test. The PASS_FAIL_DO? query indicates which actions are currently selected.

Notation

STOP	Stop acquisition
SCDP	Make a hardcopy
STO	Store in Memory
PULS	Send a pulse from the rear-panel accessory
BEEP	Emit a loud "beep"
STCD	Store to Memory Card

COMMMAND SYNTAX

Pass_Fail_DO [<pass_fail>[,<act>[,<act>...]]]

<pass_fail> := {PASS,FAIL} <act> := {STOP, SCDP, STO, PULS, BEEP, STCD}

Note: The PFDO command with no arguments (i.e. "PFDO") deletes all actions.

QUERY SYNTAX

Pass_Fail_DO?

Response Format

Pass_Fail_DO [<pass_fail>[,<act>[,<act>...]]]

EXAMPLE (GPIB)

The following command forces the oscilloscope to "beep" and stop acquiring, when the test passes.

```
CMD$="Pass_Fail_DO PASS,BEEP,STOP":
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

PASS_FAIL_COUNTER, PASS_FAIL_MASK,
PARAMETER_VALUE?, CURSOR_MEASURE,
CURSOR_SET

MISCELLANEOUS**PASS_FAIL_MASK, PFMS**

Command/Query

DESCRIPTION	The PASS_FAIL_MASK command generates a tolerance mask around a selected waveform and stores the result in Memory C. The PASS_FAIL_MASK? query returns the current tolerance values.
COMMAND SYNTAX	<code><trace>:Pass_Fail_MaSk <htol_minus>,<htol_plus>,<vtol_minus>,<vtol_plus>,<htol_minus>,<htol_plus>,<vtol_minus>,<vtol_plus></code> : float numbers <i>Note: if any arguments are missing, the previous settings will be used.</i>
QUERY SYNTAX	<code>Pass_Fail_MaSk?</code>
Response Format	<code>Pass_Fail_MaSk <htol_minus>,<htol_plus>,<vtol_minus>,<vtol_plus></code>
EXAMPLE (GPIB)	The following command generates a tolerance mask around the Channel 1 trace. The tolerance is set at 1.5 V over the trace and 0.7 V under, 25 nanoseconds before the trace and 12 after. <code>CMD\$ = "C1:PASS_FAIL_MASK 25 NS, 12 NS, 0.7 V, 1.5 V":CALL IBWRT(SCOPE%,CMD\$)</code>
RELATED COMMANDS	<code>CURSOR_MEASURE, CURSOR_SET, PASS_FAIL_COUNTER, PASS_FAIL_DO, PARAMETER_VALUE</code>

DISPLAY

PERSIST, PERS

Command/Query

DESCRIPTION	The PERSIST command enables or disables the persistence display mode.
COMMAND SYNTAX	PERSist <mode> <mode> := {ON, OFF}
QUERY SYNTAX	PERSist?
Response format	PERSist <mode>
EXAMPLE (GPIB)	The following code turns the persistence display ON. CMD\$="PERS ON": CALL IBWRT(SCOPE%,CMD\$)
RELATED COMMANDS	PERSIST_SETUP, XY_DISPLAY

CURSOR**PER_CURSOR_SET, PECS**

Command/Query

DESCRIPTION

The PER_CURSOR_SET command allows the user to position any one of the six independent cursors at a given screen location. The position of the cursor can be modified or queried even if the cursor is not currently displayed on the screen.

The PER_CURSOR_SET? query indicates the current position of the cursor(s).

Notation

HABS	Horizontal absolute	VABS	Vertical absolute
HREF	Horizontal reference	VREF	Vertical reference
HDIF	Horizontal difference	VDIF	Vertical difference

COMMAND SYNTAX

PEr_Cursor_Set <cursor>,<position>[,<cursor>,<position>,...,
<cursor>,<position>]

<cursor> := {HABS, VABS, HREF, HDIF, VREF, VDIF}

<position> := 0 to 10 DIV (horizontal)
 -4 to 4 DIV (vertical)

Note 1: The suffix DIV is optional.

Note 2: Parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be in any order and may be restricted to those variables to be changed.

QUERY SYNTAX

PEr_Cursor_Set? <cursor>[<cursor>,...,<cursor>]

<cursor> := {HABS, VABS, HREF, HDIF, VREF, VDIF,
 ALL}

Note: If <cursor> is not specified, ALL will be assumed. If the position of a cursor cannot be determined in a particular situation, its position will be indicated as UNDEF.

Response Format

PEr_Cursor_Set <cursor>,<position>[,<cursor>,<position>,...,
<cursor>,<position>]

5 *System Commands*

EXAMPLE (GPIB)

The following code positions the HREF and HDIF cursors at +2.6 DIV and +7.4 DIV respectively, using Memory C as a reference.

```
CMD$="MC:PECS HREF,2.6 DIV,HDIF,7.4DIV"
```

RELATED COMMANDS

CURSOR_MEASURE, PERSIST, PER_CURSOR_VALUE

CURSOR**PER_CURSOR_VALUE?, PECV?**

Query

DESCRIPTION

The PER_CURSOR_VALUE? query returns the values measured by the specified cursors for a given trace.

Notation

HABS	Horizontal absolute	HREL	Horizontal relative
VABS	Vertical absolute	VREL	Vertical relative

QUERY SYNTAX

<trace>:PEr_Cursor_Value? <cursor>[,<cursor>,...,<cursor>]

<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<cursor> := {HABS, VABS, HREL, VREL, ALL}

Note: If <cursor> is not specified, ALL will be assumed.

Response format

<trace>:PEr_Cursor_Value <cursor>,<value>[,<cursor>,<value>,...,<cursor>,<value>]

EXAMPLE (GPIB)

The following code returns the value measured with the vertical relative cursor on Channel 1.

```
CMD$="C1:PECV? VREL": CALL IBWRT(SCOPE%,CMD$)
```

```
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

```
C1:PECV VREL,56 MV
```

RELATED COMMANDS

CURSOR_MEASURE, PERSIST, PER_CURSOR_SET

‡ 4-channel oscilloscopes only

COMMUNICATION

PERSIST_SETUP, PESU

Command/Query

DESCRIPTION

The PERSIST_SETUP command sets the number of sweeps to be retained in the persistence display.

The PERSIST_SETUP? query reports the number of sweeps that will currently be retained in the persistence display.

COMMAND SYNTAX

PEr sist_SetUp <sweeps>

<sweeps> := {1, 2, 5, 10, 20, 50, 100, 200, INF}

QUERY SYNTAX

PEr sist_SetUp?

Response format

PEr sist_SetUp <sweeps>

EXAMPLE (GPIB)

The following code sets the number of sweeps contributing to the persistence display to 50:

```
CMD$="PESU 50": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

PERSIST

STATUS***PRE**

Command/Query

DESCRIPTION

The *PRE command sets the PaRallel poll Enable register (PRE). The lowest 8 bits of the Parallel Poll Register (PPR) are composed of the STB bits. The *PRE command allows the user to specify which bit(s) of the parallel poll register will affect the 'ist' individual status bit.

The *PRE? query reads the contents of the PRE register. The response is a decimal number which corresponds to the binary sum of the register bits.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

PRE <value>

<value> := 0 to 65 535

QUERY SYNTAX

*PRE?

Response format

*PRE <value>

EXAMPLE (GPIB)

The following command will cause the 'ist' status bit to become 1 as soon as the MAV bit (bit 4 of STB, i.e. decimal 16) is set. This yields the PRE value 16.

```
CMD$="*PRE 16": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

*IST?

SAVE/RECALL SETUP

***RCL**
Command

DESCRIPTION

The *RCL command sets the state of the instrument using one of the eight non-volatile panel setups by recalling the complete front-panel setup of the instrument. Panel setup 0 corresponds to the default panel setup.

The *RCL command produces the opposite effect of the *SAV command.

If the desired panel setup is not acceptable, the Execution error status Register (EXR) is set and the EXE bit of the standard Event Status Register (ESR) is set.

COMMAND SYNTAX

*RCL <panel_setup>
<panel_setup> := 0 to 7

EXAMPLE (GPIB)

The following code recalls the instrument setup previously stored in panel setup 5

```
CMD$="*RCL 5": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

PANEL_SETUP, *SAV, EXR?

SAVE/RECALL SETUP**RECALL_PANEL, RCPN**

Command

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The RECALL_PANEL command recalls a front-panel setup from the memory card.

COMMAND SYNTAX

ReCall_PaNel FILE,'<filename>'

<filename> := a string of up to 8 characters. Any extension will be ignored. Front-panel setup files are always stored with the extension ".PNL" on the card.

EXAMPLE (GPIB)

The following command recalls the front-panel setup from file P012.PNL

```
CMD$="RCPN FILE,'P012':CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

PANEL_SETUP, *SAV, STORE_PANEL, *RCL

WAVEFORM TRANSFER

RECALL, REC

Command

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The RECALL command recalls a waveform file from the memory card into one of the internal memories Memory C, Memory D, Function E‡, or Function F‡.

COMMAND SYNTAX

<memory>:RECall FILE,<mode>

<memory> := {MC, MD, FE‡, FF‡}

<mode> := {SAME, PREVIOUS, NEXT, NEWEST, OLDEST,
'<filename>'}

<filename> := an alphanumeric string of up to 8 characters, followed by a dot and an extension of up to 3 digits.

EXAMPLE (GPIB)

The following commands first recall a waveform file called "SC1.001" into Memory C, after which the NEXT waveform file in the card directory list is recalled to Memory D.

```
CMD$="MC:REC FILE,'SC1.001';MD:REC NEXT":  
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

FUNCTION_STATE‡, STORE, INR?

‡ 4-channel oscilloscopes only

*SAVE/RECALL SETUP****RST**

Command

DESCRIPTION

The *RST command initiates a device reset. The *RST sets all 8 traces to the GND line, recalls the default setup and causes a calibration to be performed.

COMMAND SYNTAX***RST****EXAMPLE (GPIB)**

This example resets the oscilloscope

```
CMD$="*RST": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS***CAL?, *RCL**

ACQUISITION

SAMPLE_CLOCK, SCLK‡

Command/Query

DESCRIPTION

The SAMPLE_CLOCK command allows the user to control the use of an external time base. The user sets the number of data points that will be acquired when the oscilloscope is using the external clock.

COMMAND SYNTAX

Sample_CLoCk <state>[,<recordlength>]

<state> := {INT,EXT} <recordlength> := {50, 100, 200, 500,
1000, 2000, 5000,
10000, 20000,
50000}

Note: If <recordlength> is not specified the previous value will not be modified. (The parameter <recordlength> is initially set to 50000).

QUERY SYNTAX

Sample_CLoCk?

Response Format

Sample_CLoCk <state>,<recordlength>

EXAMPLE

The following command sets the oscilloscope to use the external clock with 1000 data point records.

CMD\$ = "SCLK EXT,1000": CALL IBWRT(SCOPE%,CMD\$)

‡ not available on 941X

SAVE/RECALL SETUP***SAV**
Command**DESCRIPTION**

The *SAV command stores the current state of the instrument in non-volatile internal memory. The *SAV command stores the complete front-panel setup of the instrument at the time the command is issued.

*Note: The communication parameters (those modified by commands COMM_FORMAT, COMM_HEADER, COMM_HELP, COMM_ORDER and WAVEFORM_SETUP) and the enable registers associated with the status reporting system (*SRE, *PRE, *ESE, INE) are not saved by this command.*

COMMAND SYNTAX

*SAV <panel_setup>
<panel_setup> := 1 to 7

EXAMPLE (GPIB)

The following code saves the current instrument setup in panel setup 5.

```
CMD$="*SAV 5": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

PANEL_SETUP, *RCL

HARD COPY

SCREEN_DUMP, SCDP

Command/Query

DESCRIPTION

The SCREEN_DUMP causes the oscilloscope to dump the screen contents onto the hard copy device. For plotting, this command will not halt oscilloscope activities since plotting is performed in parallel with other tasks, unless it is done over the same port as the remote control. Printing, however, cannot be done in parallel with other oscilloscope operations.

Screen dumps may be aborted by adding [A] to the screen dump command, as shown in the command syntax below.

The time/date stamp which appears on the plot corresponds to the time at which the command was executed.

The SCREEN_DUMP? query indicates whether a screen dump is currently in progress (ON) or has finished (OFF).

COMMAND SYNTAX

SCreen_DumP [A]

Note: The optional parameter "A" may be used to abort a screen dump.

QUERY SYNTAX

SCreen_DumP?

Response format

SCreen_DumP <status>
<status> := {ON, OFF}

EXAMPLE (GPIB)

The following code initiates a screen dump.

```
CMD$="SCDP"; CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

INR?, HARDCOPY_SETUP, HARDCOPY_TRANSMIT

ACQUISITION

SEGMENTS, SEGS

Command/Query

DESCRIPTION

The SEGMENTS command sets the number of segments for sequence mode acquisition.

The response to the SEGMENTS? query indicates the number of segments which is set in the oscilloscope.

COMMAND SYNTAX

SEGmentS <segments>

	9410/14	All others
<segments> :=	{2,5,10,20,50}	{2,5,10,20,50,100,200}

QUERY SYNTAX

SEGmentS?

Response Format

SEGmentS <segments>

EXAMPLE

The following command sets the segment count to 100.

CMD\$ = "SEGS 100": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

TRIG_MODE

DISPLAY

SELECT, SEL
Command/Query

DESCRIPTION

The SELECT command selects the specified trace for manual display control. An environment error (Table 6, page 84) is generated if the specified trace is not displayed.

The SELECT? query returns the selection status of the specified trace.

COMMAND SYNTAX

<trace>:SElect

<trace> := {EA, EB, MC, MD, FE, FF}

QUERY SYNTAX

<trace>:SElect?

Response format

<trace>:SElect <mode>

<mode> := {ON, OFF}

EXAMPLE (GPIB)

The following command selects Expand B (EB).

```
CMD$="EB:SEL": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

TRACE

STATUS***SRE**

Command/Query

DESCRIPTION

The *SRE command sets the Service Request Enable register (SRE). This command allows the user to specify which summary message bit(s) in the STB register will generate a service request. Refer to Table 8, page 137 for an overview of the available summary messages.

A summary message bit is enabled by writing a 1 into the corresponding bit location. Conversely, writing a 0 into a given bit location prevents the associated event from generating a service request (SRQ). Clearing the SRE register disables SRQ interrupts.

The *SRE? query returns a value which when converted to a binary number represents the bit settings of the SRE register. Note that bit 6 (MSS) cannot be set and its returned value is always zero.

Note: This command can be executed in both local and remote modes.

COMMAND SYNTAX

*SRE <value>
<value> := 0 to 255

QUERY SYNTAX

*SRE?

Response format

*SRE <value>

EXAMPLE (GPIB)

The following command allows an SRQ to be generated as soon as the MAV summary bit (bit 4, i.e. decimal 16) and/or the INB summary bit (bit 0, i.e. decimal 1) in the STB register are set. Summing these two values yields the SRE mask $16 + 1 = 17$.

CMD\$="*SRE 17": CALL IBWRT(SCOPE%,CMD\$)

STATUS

***STB?**

Query

DESCRIPTION

The *STB? query reads the contents of the 488.1 defined status register (STB), and the Master Summary Status (MSS). The response represents the values of bits 0 to 5 and 7 of the Status Byte register and the MSS summary message.

The response to a *STB? query is identical to the response of a serial poll except that the MSS summary message appears in bit 6 in place of the RQS message. Refer to Table 8, page 137 for further details on the status register structure.

QUERY SYNTAX

*STB?

Response format

*STB <value>

<value> := 0 to 255

EXAMPLE (GPIB)

The following instruction reads the status byte register.

```
CMD$="*STB?"; CALL IBWRT(SCOPE%,CMD$)  
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message

*STB 0

RELATED COMMANDS

ALL_STATUS?, *CLS, *PRE, *SRE

Bit	Value	Name	Description	Note
7	128	DIO7	0 Reserved for future use	
6	64	MSS/RQS MSS =1 RQS =1	at least 1 bit in STB masked by SRE is 1 service is requested	(1) (2)
5	32	ESB	1 an ESR enabled event has occurred	(3)
4	16	MAV	1 Output queue is not empty	(4)
3	8	DIO3	0 Reserved	
2	4	VAB	1 a command data value has been adapted	(5)
1	2	DIO1	0 Reserved	
0	1	INB	1 an enabled INternal state change has occurred.	(6)

Status Byte Register (STB)

Table 8

Notes:

- (1) The Master Summary Status (MSS) indicates that the instrument requests service while the Service Request status - when set - specifies that the oscilloscope issued a service request. Bit position 6 depends on the polling method:
 Bit 6 = MSS if a *STB? query is received
 = RQS if serial polling is conducted
- (2) Example: If SRE = 10 and STB = 10 then MSS = 1. If SRE = 010 and STB = 100 then MSS=0.
- (3) The Event Status Bit (ESB) indicates whether or not one or more of the enabled IEEE 488.2 events have occurred since the last reading or clearing of the Standard Event Status Register (ESR). ESB is set if an enabled event becomes true (1).
- (4) The Message AVailable bit (MAV) indicates whether or not the Output queue is empty. The MAV summary bit is set true (1) whenever a data byte resides in the Output queue.
- (5) The Value Adapted Bit (VAB) is set true (1) whenever a data value in a command has been adapted to the nearest legal value. For instance, the VAB bit would be set if the time base is redefined as 2.5 μ sec/div since the adapted value is 2 μ sec/div.
- (6) The INternal state Bit (INB) is set true (1) whenever certain enabled internal states are entered. For further information, refer to the INR? query.

ACQUISITION

STOP Command

DESCRIPTION

The STOP command immediately stops the acquisition of a signal. It changes the acquisition state from "ready" to "triggered", and if the trigger mode is AUTO or NORM it will change to trigger mode SINGLE to prevent further acquisition.

COMMAND SYNTAX

STOP

EXAMPLE

The following command stops the acquisition process.

```
CMD$ = "STOP": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

ARM_ACQUISITION, TRIG_MODE, WAIT

WAVEFORM TRANSFER**STORE, STO**

Command

DESCRIPTION

The STORE command stores the contents of the specified trace into one of the internal function memories Memory C, Memory D, Function E‡, Function F‡, or to the memory card.*

COMMAND SYNTAX

```
<memory>:STOre <trace>[, '<filename>']*
<memory> := {MC, MD, FE‡, FF‡, CARD*}
<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}
<filename> := a string of up to 8 characters.
```

Notes applying to oscilloscopes equipped with the memory card option:

1. Any extension supplied with the filename is ignored.
2. If no filename (or an empty string) is supplied, the oscilloscope generates a filename according to its internal rules.
3. If the CARD:STORE command is sent without any argument, all traces currently enabled in the Store Setup will be stored on the card. This setup can be modified using the STORE_SETUP command.

EXAMPLE (GPIB)

The following command stores the contents of Expand B (EB) into Memory D (MD).

```
CMD$="MD:STO EB": CALL IBWRT(SCOPE%,CMD$)
```

The following command stores C1 on the memory card in a file called "DIODE".*

```
CMD$="CARD:STO C1,'DIODE'":
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

FUNCTION_STATE‡, STORE_SETUP*, RECALL

‡ 4-channel oscilloscopes only

* only oscilloscopes equipped with the memory card option

SAVE/RECALL PANEL

STORE_PANEL, STPN

Command

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The STORE_PANEL command stores the complete front-panel setup of the instrument, at the time the command is issued, into a file on the memory card.

*Note: The communication parameters (those modified by commands COMM_FORMAT, COMM_HEADER, COMM_HELP, COMM_ORDER and WAVEFORM_SETUP) and the enable registers associated with the status reporting system (*SRE, *PRE, *ESE, INE) are not saved by this command.*

COMMAND SYNTAX

STore_PaNel FILE, '<filename>'

<filename> := a string of up to 8 characters. Any extension will be ignored. Front-panel setup files are always stored with the extension ".PNL" on the card.

Note: If no filename (or an empty string) is supplied, the oscilloscope generates a filename according to its internal rules.

EXAMPLE (GPIB)

The following code saves the current instrument setup to the memory card in a file called "DIODE.PNL":

```
CMD$="STPN FILE,'DIODE':  
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

PNSU, *SAV, RECALL_PANEL, *RCL

WAVEFORM TRANSFER**STORE_SETUP, STST**

Command/Query

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The STORE_SETUP command controls the way in which traces will be stored on the memory card. Traces may be enabled (FILE) or disabled (DIS) for storage. This applies to autostoring or to the CARD_STORE STST command. Traces may be autostored to the memory card after each acquisition until the card becomes full (FILL), or forever (WRAP), replacing the oldest traces by new ones. They may be stored keeping the full precision of the data (WORD), or with reduced precision (BYTE, most significant bits only) to conserve space on the card.

The STORE_SETUP? query returns the current mode of operation of Autostore, the current data size, and the current trace selection.

COMMAND SYNTAX

STore_SeTup [<trace>,<enable>,...,<trace>,<enable>]
[,AUTO,<mode>][,SIZE,<data size>]

<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<enable> := {FILE, DIS}

<mode> := {OFF, WRAP, FILL}

<data_size := {WORD, BYTE}

Note 1: Parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be given in any order and may be restricted to those variables which have to be changed.

Note 2: The STORE_SETUP command only modifies the current setup for storing traces to the memory card. Any parameters not given in a particular STORE_SETUP command will not be affected, i.e. they will stay at their previous settings.

QUERY SYNTAX

STore_SeTup?

Response format

STore_SeTup <trace>,<enable>,...,<trace>,<enable>,AUTO,
<mode>,SIZE,<size>

‡ 4-channel oscilloscopes only

5 *System Commands*

EXAMPLE (GPIB)

The following command selects Channel 1 and Channel 2 traces to be stored. It prevents Expansion A from being stored (DIS) and enables an "autostore" to the card until no more space is left on the memory card (AUTO, FILL). Every data point will be transferred to the card in BYTE format.

```
CMD$="STST C1,FILE, C2,FILE, EA,DIS, AUTO,FILL,  
SIZE,BYTE":CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

STORE, INR?

WAVEFORM TRANSFER**STORE_TEMPLATE, SITM**

Command

Oscilloscopes fitted with the MC01 Option

DESCRIPTION

The STORE_TEMPLATE command stores the instrument's waveform template on the memory card. A filename is automatically generated in the form of "LECROYvv.TPL" where "vv" is the 2-digit revision number.

Note: As an example, the file name generated will be LECROY21.TPL for revision 2.1.

Refer to Section 6 for further information about the waveform template.

COMMAND SYNTAX

STore_TeMplate

EXAMPLE (GPIB)

The following code stores the current waveform template on the memory card for future reference.

```
CMD$="STTM":CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

TEMPLATE?

WAVEFORM TRANSFER

TEMPLATE?, TEMPL?

Query

DESCRIPTION

The **TEMPLATE?** query produces a copy of the template which formally describes the various logical entities making up a complete waveform. In particular, the template describes in full detail the variables contained in the descriptor part of a waveform. Refer to Section 6 for further information.

QUERY SYNTAX

TeMPLate?

Response format

TeMPLate "<template>"

<template> := A variable length string detailing the structure of a waveform.

RELATED COMMANDS

INSPECT?

ACQUISITION**TIME_DIV, TDIV**

Command/Query

DESCRIPTION

The TIME_DIV command modifies the time-base setting. The new time-base setting may be specified with suffixes NS for nanoseconds, US for microseconds, MS for milliseconds, S for seconds or KS for kiloseconds. An out-of-range value causes the VAB bit (bit 2) in the STB register (Table 8, STB) to be set.

The oscilloscope will force random interleaved sampling (RIS) for time-base settings in the "ON" only range, and single-shot sampling for time-base settings in the "OFF" only range. For time-base settings in the "Selectable" range, the INTERLEAVED command allows the user to choose the required sampling mode.

RIS	9410/14	9420/24/30	9450
"ON" only	2 ns/div to 20 ns/div	1 ns/div [‡] to 20 ns/div	1 ns/div to 5 ns/div
Selectable "ON"/"OFF"	50 ns/div to 5 μs/div	50 ns/div to 20 μs/div	10 ns/div to 5 μs/div
"OFF" only	10 μs/div to 1000 s/div	50 μs/div to 5000 s/div	10 μs/div to 5000 s/div

Sequence mode acquisitions also force single-shot sampling and therefore cannot be performed for time-base settings in the RIS "ON" only range (see table above).

The TIME_DIV? query returns the current time-base setting.

COMMAND SYNTAX

Time_DIV <value>

<value> := 1 NS to 5 KS

Note: The suffix S (seconds) is optional.

QUERY SYNTAX

Time_DIV?

Response Format

Time_DIV <value>

[‡] 2 ns/div for Model 9430

EXAMPLE (GPIB)

The following command sets the time base to 500 μ sec/div.

```
CMD$="TDIV 500US": CALL IBWRT(SCOPE%,CMD$)
```

The following command sets the time base to 2 msec/div.

```
CMD$="TDIV.002": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

INTERLEAVED, TRIG_DELAY, TRIG_MODE

DISPLAY**TRACE, TRA**
Command/Query**DESCRIPTION**

The TRACE command enables or disables the display of a trace. An environment error (Table 6, page 84) is set if an attempt is made to display more than four waveforms.

The TRACE? query indicates whether the specified trace is displayed or not.

COMMAND SYNTAX

<trace>:TRAcE <mode>

<trace> := {C1, C2, C3‡, C4‡, EA, EB, MC, MD, FE, FF}

<mode> := {ON, OFF}

QUERY SYNTAX

<trace>:TRAcE?

Response format

<trace>:TRAcE <mode>

EXAMPLE (GPIB)

The following command displays Function E (FE).

CMD\$="FE:TRA ON": CALL IBWRT(SCOPE%,CMD\$)

‡ 4-channel oscilloscopes only

ACQUISITION

***TRG**
Command

DESCRIPTION

The *TRG command executes an ARM command.

*Note: The *TRG command is the equivalent of the 488.1 GET (Group Execute Trigger) message.*

COMMAND SYNTAX

*TRG

EXAMPLE (GPIB)

The following command enables signal acquisition.

CMD\$="*TRG": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

ARM_ACQUISITION, STOP, WAIT

ACQUISITION**TRIG_COUPLING, TRCP**

Command/Query

DESCRIPTION

The TRIG_COUPLING command sets the coupling mode of the specified trigger source. The trigger slope is automatically changed to positive when the trigger coupling is set to HFDIV.

Note: HFDIV is indicated as HF on the front panel. See the Operator's Manual, Section 5 (9424) or Section 6 (9450/20).

The TRIG_COUPLING? query returns the trigger coupling of the selected source.

COMMAND SYNTAX

<trig_source>:TRig_CouPling <trig_coupling>

	9410/14/30	9420/50	9424	9424E
<trig_source> :=	{C1,C2,EX,EX10}	{C1,C2,EX,EX10}	{C1,C2,C4}	{C1,C2,EX}
<trig_coupling> :=	{AC,DC,HFREJ,LFREJ}	{AC,DC,HFREJ,LFREJ,HFDIV}		

QUERY SYNTAX

<trig_source>:TRig_CouPling?

Response format

<trig_source>:TRig_CouPling <trig_coupling>

EXAMPLE (GPIB)

The following command sets the coupling mode of the trigger source Channel 2 to high frequency reject.

```
CMD$="C2:TRCP HFREJ": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

TRIG_SELECT

ACQUISITION

TRIG_DELAY, TRDL

Command/Query

DESCRIPTION

The TRIG_DELAY command sets the time at which the trigger is to occur with respect to the first acquired data point (displayed at the left hand edge of the screen).

The command expects positive trigger delays to be expressed as a percentage of the full horizontal screen (this mode is called pre-trigger acquisition as data are acquired before the trigger occurs). Negative trigger delays must be given in seconds (this mode is called post-trigger acquisition as the data are acquired after the trigger has occurred).

If a value outside the range $-10\ 000\ \text{div} \times \text{time/div}$ and 100% is specified, the trigger time will be set to the nearest limit and the VAB bit (bit 2) will be set in the STB register.

The response to the TRIG_DELAY? query indicates the trigger time with respect to the first acquired data point. Positive times are expressed as a percentage of the full horizontal screen and negative times in seconds.

COMMAND SYNTAX

TRig_DeLay <value>

<value> := 0.00 PCT to 100.00 PCT (pretrigger)
-20 PS to -50 MAS (post-trigger)

Note: The suffix is optional. For positive numbers the suffix PCT is assumed. For negative numbers the suffix S is assumed. MAS is the suffix for Msec (megaseconds), useful only for extremely large delays at very slow time bases.

QUERY SYNTAX

TRig_DeLay?

Response format

TRig_DeLay <value>

EXAMPLE (GPIB)

The following command sets the trigger delay to - 20 sec (post-trigger)

```
CMD$="TRDL -20S": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

TIME_DIV

ACQUISITION

TRIG_LEVEL, TRLV
Command/Query

DESCRIPTION

The TRIG_LEVEL command adjusts the trigger level of the specified trigger source. An out-of-range value will be adjusted to the closest legal value and will cause the VAB bit (bit 2) in the STB register (Table 8, page 137) to be set.

The range of values is as follows:

- ± 5 times the total V/div setting – with CHAN 1, 2 or 4 as trigger source.
- ± 2 V with EXT as trigger source (± 0.8 V in 9424E)
- ± 20 V with EXT/10 as trigger source.

The TRIG_LEVEL? query returns the current trigger level.

COMMAND SYNTAX

<trig_source>:TRig_LeVel <trig_level>

	2-channel oscilloscopes	4-channel with Ext trigger	4-channel without Ext trigger
<trig_source> :=	{C1,C2,EX,EX10}	{C1,C2,EX, EX10&}	{C1,C2,C4}

<trig_level> := -20V to 20V (maximum range)

Note: The suffix V is optional.

QUERY SYNTAX

<trig_source>:TRig_LeVel?

Response format

<trig_source>:TRig_LeVel <trig_level>

EXAMPLE (GPIB)

The following command adjusts the trigger level of Channel 2 to -3.4 V.

CMD\$="C2:TRLV -3.4V": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

TRIG_SELECT

& Not available in 9424E

ACQUISITION

TRIG_MODE, TRMD

Command/Query

DESCRIPTION

The TRIG_MODE command specifies the trigger mode. An environment error (Table 6, page 84) will be generated when TRMD SEQNCE is received while the instrument is in the interleaved sampling (RIS) acquisition mode. With the mode SINGLE, this command will not arm the trigger. Use the command ARM_ACQUISITION to actually start a single acquisition.

The TRIG_MODE? query returns the current trigger mode.

COMMAND SYNTAX

TRig_MoDe <mode>

<mode> := {AUTO, NORM, SEQNCE, SINGLE, WRAP}

QUERY SYNTAX

TRig_MoDe?

Response format

TRig_MoDe <mode>

EXAMPLE (GPIB)

The following command selects the sequence mode.

```
CMD$="TRMD SEQNCE": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

ARM_ACQUISITION, STOP, TRIG_SELECT

ACQUISITION**TRIG_PATTERN, TRPA**

(2-ch. and 4-ch. oscilloscopes with external trigger)

Command/Query

DESCRIPTION

The TRIG_PATTERN command defines a trigger pattern. The command specifies the logic composition of the pattern sources (Channel 1, Channel 2, External) and the conditions under which a trigger can occur. Note that this command can be used even if the complex trigger mode has not been activated.

Notation

L	Low		
H	High		
X	Don't Care		
PR	pattern present	AB	pattern absent
EN	pattern entered	EX	pattern exited

The TRIG_PATTERN? query returns the current trigger pattern.

COMMAND SYNTAX

TRig_PAttern <C1_state>,<C2_state>,<EX_state>,<trig_condi-
tion>

<C1_state> := {L,H,X}

<C2_state> := {L,H,X}

<EX_state> := {L,H,X}

<trig_condition> := {PR, AB, EN, EX}

QUERY SYNTAX

TRig_PAttern?

Response format

TRig_PAttern <C1_state>,<C2_state>,<EX_state>,<trig_condi-
tion>

EXAMPLE (GPIB)

The following command configures the logic state of the pattern as HLX (CH1 = H, CH2 = L, EX = X) and defines the trigger condition as pattern absent (AB).

CMD\$="TRPA H,L,X,AB": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

TRIG_SELECT

ACQUISITION

TRIG_PATTERN, TRPA

(4-channel oscilloscopes without external trigger)

Command/Query

DESCRIPTION

The TRIG_PATTERN command defines a trigger pattern. The command specifies the logic composition of the pattern sources (Channel 1, Channel 2, and Channel 4) and the conditions under which a trigger can occur. Note that this command can be used even if the complex trigger mode has not been activated.

Notation			
L	Low		
H	High		
X	Don't Care		
PR	pattern present	AB	pattern absent
EN	pattern entered	EX	pattern exited

The TRIG_PATTERN? query returns the current trigger pattern.

COMMAND SYNTAX

TRig_PAttern <C1_state>,<C2_state>,<C4_state>,<trig_condition>

<C1_state> := {L,H,X}

<C2_state> := {L,H,X}

<C4_state> := {L,H,X}

<trig_condition> := {PR, AB, EN, EX}

QUERY SYNTAX

TRig_PAttern?

Response format

TRig_PAttern <C1_state>,<C2_state>,<C4_state>,<trig_condition>

EXAMPLE (GPIB)

The following command configures the logic state of the pattern as HLX (CH1 = H, CH2 = L, CH4 = X) and defines the trigger condition as pattern absent (AB).

CMD\$="TRPA H,L,X,AB": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

TRIG_SELECT

ACQUISITION**TRIG_SELECT, TRSE**

Command/Query

DESCRIPTION

The TRIG_SELECT command selects the condition that will trigger the acquisition of waveforms. Depending on the trigger type, additional parameters have to be specified.

The additional parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be given in any order and may be restricted to those variables to be changed.

Note: The state-qualified, time/event qualified and pattern trigger types use the trigger pattern defined by the command TRIGGER_PATTERN.

The TRIG_SELECT? query returns the current trigger condition.

Trigger Notation

STD	Standard	SNG	Single source
PA	Pattern	SQ	State qualified
TEQ	Time event qualified	TI	Time
PL	Pulse larger	IL	Interval larger
EV	Event	PS	Pulse smaller
IS	Interval smaller	SR	Source
HT	Hold type	HV	Hold value

TV Trigger Notation

FLD	Field	FLDC	Field Count
LINE	Line	CHAR	Characteristics
LPIC	Lines per picture	ILAC	Interlace

SR does not apply to the Pattern trigger.

HT and HV do not apply to the standard trigger.

NON TV TRIGGER**COMMAND SYNTAX**

TRIG_Select <trig_type>,SR,<source>,HT,<hold_type>,
HV,<hold_value>

<trig_type> := {STD, SNG, SQ, TEQ, PA}
 <source> := {C1, C2, LINE, EX, EX10&}‡
 <source> := {C1, C2, LINE, C4}@

‡ 2-ch. and 4-ch. oscilloscopes with ext. trigger, & Not available in 9424E, @ 4-ch. oscilloscopes without ext. trigger

<hold_type> := {TI, EV, PS, PL, IS, IL}
 <hold_value> := 25NS to 20S for TI
 1 to 10⁹ for EV
 2.5NS to 20S for PS and PL*
 10NS to 20S for IS*
 25NS to 20S for IL*

* these values are only valid for single-source and pattern triggers.

Note: The suffix S (seconds) is optional.

QUERY SYNTAX

TRig_Select?

Response format

TRig_Select <trig_type>,SR,<source>,HT,<hold_type>,
HV,<hold_value>

EXAMPLE (GPIB)

The following command selects the single-source trigger with Channel 1 as trigger source. Hold type and hold value are chosen as "Pulse smaller" than 20 μ sec

```
CMD$="TRSE SNG,SR,C1,HT,PS,HV,20 US"
CALL IBWRT(SCOPE%,CMD$)
```

TV TRIGGER

COMMAND SYNTAX

TRig_Select TV,SR,<source>,FLDC,<field_count>,FLD,<field>,
CHAR,<characteristics>,LPIC@,<lpic>,ILAC@,<ilace>,
LINE,<line> LINE,<line>

<source> := EX in 2-channel oscilloscopes

<source> := C3 in 4-channel oscilloscopes

<field_count> := {1, 2, 4, 8}

<field> := 1 to field_count

<characteristics> := { NTSC,PALSEC,CUST50,CUST60}

<lpic> := 1 to 1500

<ilace> := {1, 2, 4, 8}

<line> := 1 to 1500

*Note: The FLD value is interpreted with the current FLDC value.
The LINE value is interpreted with the current FLD and CHAR values.*

QUERY SYNTAX

TRig_Select?

Response format

TRig_Select TV,SR,EX,FLDC,<field_count>,FLD,<field>,
CHAR,<characteristic>,LINE,<line>

@ for CUST50 and CUST60 only

EXAMPLE (GPIB)

CMD\$ = "TRSE TV,SR,EX,FLDC,8,FLD,3,CHAR,
PALSEC,LINE,17"

RELATED COMMANDS

TRIG_COUPLING, TRIG_LEVEL, TRIG_MODE, TRIG_PAT-
TERN, TRIG_SLOPE

ACQUISITION

TRIG_SLOPE, TRSL

Command/Query

DESCRIPTION

The TRIG_SLOPE command sets the trigger slope of the specified trigger source. An environment error (see Table 6, page 84) will be generated when TRSL NEG is received while the trigger coupling is set to HFDIV (see TRIG_COUPLING).

The TRIG_SLOPE? query returns the trigger slope of the selected source.

COMMAND SYNTAX

<trig_source>:TRig_SLOpe <trig_slope>

	2-channel oscilloscopes	4-channel with Ext trigger	4-channel without Ext trigger
<trig_source> :=	{C1,C2,EX,EX10}	{C1,C2,EX, EX10 ^{&} }	{C1,C2,C4}

<trig_slope> := {NEG, POS}

QUERY SYNTAX

<trig_source>:TRig_SLOpe?

Response format

<trig_source>:TRig_SLOpe <trig_slope>

EXAMPLE (GPIB)

The following command sets the trigger slope of Channel 2 to negative.

CMD\$="C2:TRSL NEG": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

TRIG_SELECT

[&] Not available in 9424E

MISCELLANEOUS***TST?**

Query

DESCRIPTION

The *TST? query performs an internal self-test. The response indicates if the self-test detected any errors. The self-test includes testing the hardware of all channels, the time base and the trigger circuits.

Hardware failures are identified by a unique binary code in the returned <status> number (see Table 1, page 46). A "0" response indicates that no failures occurred.

Note: This query is only accepted in remote mode.

QUERY SYNTAX***TST?****Response Format*****TST <status>****<status>:= 0 self-test successful****EXAMPLE (GPIB)**

This example causes a self-test to be performed.

```
CMD$="*TST?": CALL IBWRT(SCOPE%,CMD$)  
CALL IBRD(SCOPE%,RD$): PRINT RD$
```

Response message (if no failure)

```
*TST 0
```

RELATED COMMANDS***CAL?**

STATUS

URR?
Query

DESCRIPTION

The URR? query reads and clears the contents of the User Request status Register (URR). The URR register specifies which button in the menu field was pressed. Refer to Table 9 for further details.

In the remote mode, the URR register indicates the last button (2 ... 10) which was pressed. In local mode, the URR register indicates whether the CALL HOST button has been pressed. If no menu button has been pressed since the last URR? query, the value 0 is returned. Figure 1, (page NO TAG) shows the button assignments on the instrument.

QUERY SYNTAX

URR?

Response format

URR <value>
<value> := 0 to 9, 100

EXAMPLE (GPIB)

The following instruction reads the contents of the URR register.

```
CMD$="URR?": CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RSP$): PRINT RSP$
```

Response message
URR 0

RELATED COMMANDS

CALL_HOST, KEY, ALL_STATUS?, *CLS

Value	Description
0	no button has been pressed
1	button 1 has been pressed.
2	button 2 has been pressed.
3	button 3 has been pressed.
4	button 4 has been pressed.
5	button 5 has been pressed.
6	button 6 has been pressed.
7	button 7 has been pressed.
8	button 8 has been pressed.
9	button 9 has been pressed.
100	The "Call Host" key (button 10 in root menu) has been pressed.

User Request Status Register Structure (URR)

Table 9

DISPLAY**VERT_MAGNIFY, VMAG**

Command/Query

DESCRIPTION

The VERT_MAGNIFY command vertically expands the specified trace. The command is executed even if the trace is not displayed. The VERT_MAGNIFY? query returns the magnification factor of the specified trace.

COMMAND SYNTAX

<trace>:Vert_MAGnify <factor>

<trace> := {EA, EB, MC, MD, FE, FF}

<factor> := 0.2 to 50 (maximum)

QUERY SYNTAX

<trace>:Vert_MAGnify?

Response format

<trace>:Vert_MAGnify <factor>

EXAMPLE (GPIB)

The following command enlarges the vertical amplitude of Function E by a factor of 3.45 with respect to its original amplitude.
CMD\$="FE:VMAG 3.45": CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

VERT_POSITION

DISPLAY

VERT_POSITION, VPOS

Command/Query

DESCRIPTION

The VERT_POSITION command adjusts the vertical position of the specified trace on the screen. The VERT_POSITION command does not affect the original offset value obtained at acquisition time.

The VERT_POSITION? query returns the current vertical position of the specified trace.

COMMAND SYNTAX

<trace>:Vert_POSition <display_offset>

<trace> := {EA, EB, MC, MD, FE, FF}

<display_offset> := -56 DIV to 56 DIV

Note: The suffix DIV is optional.

QUERY SYNTAX

<trace>:Vert_POSition?

Response format

<trace>:Vert_POSition <display_offset>

EXAMPLE (GPIB)

The following command shifts Expand A (EA) upwards by +3 divisions, relative to the position at the time of acquisition.

```
CMD$="EA:VPOS 3DIV": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

VERT_MAGNIFY

ACQUISITION**VOLT_DIV, VDIV**

Command/Query

DESCRIPTION

The VOLT_DIV command sets the vertical sensitivity in Volts/div. The VAB bit (bit 2) in the STB register (Table 8, page 137) is set if an out-of-range value is entered.

Note: The probe attenuation factor is not taken into account for adjusting vertical sensitivity.

The VOLT_DIV? query returns the vertical sensitivity of the specified channel.

COMMAND SYNTAX

<channel>:Volt_DIV <v_gain>

<channel> := {C1, C2, C3‡, C4‡}

<v_gain> :=	9410/14/30	All others
	1 mV to 2.5 V	5.0 mV to 2.5 V

Note: The suffix V is optional.

QUERY SYNTAX

<channel>:Volt_DIV?

Response format

<channel>:Volt_DIV <v_gain>

EXAMPLE (GPIB)

The following command sets the vertical sensitivity of channel 1 to 50 mV/div.

```
CMD$="C1:VDIV 50MV":
CALL IBWRT(SCOPE%,CMD$)
```

‡ 4-channel oscilloscopes only

STATUS

***WAI**
Command

DESCRIPTION

The ***WAI** (WAI to continue) command, required by the IEEE 488.2 standard, has no effect on the oscilloscope as the oscilloscope only starts processing a command when the previous command has been entirely executed.

Note: This command can be executed in both local and remote modes.

Command syntax

***WAI**

RELATED COMMANDS

***OPC**

ACQUISITION**WAIT**
Command**DESCRIPTION**

The WAIT command prevents the instrument from analyzing new commands until the oscilloscope has completed the current acquisition process.

COMMAND SYNTAX

WAIT

EXAMPLE

```
send: "TRMD SINGLE"  
loop {send: "ARM; WAIT;C1:PAVA? MAX"  
      read response  
      process response  
      }
```

This example finds the maximum amplitudes of several signals acquired one after another. ARM starts a new data acquisition. The WAIT command ensures that the maximum is evaluated for the newly acquired waveform.

"C1:PAVA? MAX" instructs the instrument to evaluate the maximum data value in the Channel 1 waveform.

WAVEFORM TRANSFER

WAVEFORM, WF Command/Query

DESCRIPTION

A WAVEFORM? query transfers a waveform from the oscilloscope to the controller, whereas a WAVEFORM command transfers a waveform from the controller to the oscilloscope.

The WAVEFORM command stores an external waveform back into the oscilloscope's internal memory. A waveform consists of several distinct entities:

- (1) the descriptor (DESC),
- (2) the user text (TEXT),
- (3) the time (TIME) descriptor,
- (4) the data (DAT1) block, and optionally
- (5) a second block of data (DAT2).

For further information on the structure of the waveform refer to Section 6. In the 9424 the WAVEFORM command automatically sets the corresponding function to the memory state.

Note: Only complete waveforms queried with "WAVEFORM? ALL" can be restored into the oscilloscope.

The WAVEFORM? query instructs the oscilloscope to transmit a waveform to the controller. The entities may be queried independently. If the "ALL" parameter is specified, all 4 or 5 entities are transmitted in one block in the order enumerated above.

Note: The format of the waveform data depends on the current settings specified by the last WAVEFORM_SETUP command, the last COMM_ORDER command and the last COMM_FORMAT command.

COMMAND SYNTAX

<memory>:WaveForm ALL,<waveform_data_block>

<memory>:= {MC, MD, FE‡, FF‡}

<waveform_data_block> := arbitrary data block

QUERY SYNTAX

<trace>:WaveForm? <block>

<trace>:= {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<block>:= {DESC, TEXT, TIME, DAT1, DAT2, ALL}

Note: If no parameter is given ALL will be assumed.

‡ 4-channel oscilloscopes only

Response format

<trace>:WaveForm <block>,<waveform_data_block>

Note: It may be convenient to disable the response header if the waveform is to be restored. Refer to command COMM_HEADER for further details.

EXAMPLE (GPIB)

- 1) The following command reads the block DAT1 from Memory C and saves it in the file "MEMC.DAT". The path header "MC:" is saved together with the data.

```
FILE$ = "MEMC.DAT"
CMD$ = "MC:WF? DAT1"
CALL IBWRT(SCOPE%,CMD$)
CALL IBRDF(SCOPE%,FILE$)
```

- 2) In the following example, the entire contents of Channel 1 are saved in the file "CHAN1.DAT". The path header "C1:" is skipped to ensure that the data can later be recalled into the oscilloscope.

```
FILE$ = "CHAN1.DAT":RD$=SPACES$(3)
CMD$ = "CHDR SHORT; C1:WF?"
CALL IBWRT(SCOPE%,CMD$)
CALL IBRD(SCOPE%,RD$) Skip first 3 characters "C1:"
CALL IBRDF(SCOPE%,FILE$) Save data in the file
"CHAN1.DAT"
```

- 3) The following example illustrates how the waveform data saved in example 2) can be recalled into Memory C.

```
FILE$ = "CHAN1.DAT"
CMD$ = "MC:TRACE ON"
CALL IBWRT(SCOPE%,CMD$)
CALL IBWRTF(SCOPE%,FILE$)
```

The MC:TRACE ON command ensures that the <trace> is set to "MC". When the data file is sent to the instrument, it first sees the header "WF" (the characters "C1:" having been skipped when reading the file) and assumes the default destination "MC".

RELATED COMMANDS

INSPECT?, COMM_FORMAT, COMM_ORDER, FUNCTION_STATE‡, TEMPLATE?, WAVEFORM_SETUP, WAVEFORM_TEXT,

‡ 4-channel oscilloscopes only

WAVEFORM TRANSFER

WAVEFORM_SETUP, WFSU

Command/Query

DESCRIPTION

The **WAVEFORM_SETUP** command specifies the amount of data in a waveform which will be transmitted to the controller. The command controls the settings of the following parameters:

- a. Sparsing (SP). The sparsing parameter defines the interval (0..25000) between data points. For example:

- SP = 0 reads all data points
 - SP = 1 reads all data points
 - SP = 4 reads every 4th data point

- b. Number of points (NP). The number of points parameter indicates how many points should be transmitted. For example:

- NP = 0 sends all data points
 - NP = 1 sends 1 data point
 - NP = 5 sends a maximum of 5 data points
 - NP = 10 sends a maximum of 10 data points

- c. First point (FP). The first point parameter specifies the address of the first data point to be sent. For waveforms acquired in sequence mode, this refers to the relative address in the given segment. For example:

- FP = 0 corresponds to the first data point
 - FP = 1 corresponds to the second data point
 - FP = 5000 corresponds to data point 5001

- d. Segment number (SN). The segment number parameter indicates which segment should be sent if the waveform was acquired in sequence mode. This parameter is ignored for non-segmented waveforms. For example:

- SN = 0 all segments
 - SN = 1 first segment
 - SN = 23 segment 23

The **WAVEFORM_SETUP?** query returns the transfer parameters currently in use.

Notation			
SP	sparsing	NP	number of points
FP	first point	SN	segment number

COMMAND SYNTAX

WaveForm_SetUp SP,<sparsing>,NP,<number>,FP,<point>,
SN,<segment>

	9410/14	All others
<sparsing> :=	0 to 5000 (0 = no sparsing)	0 to 25000
<number> :=	0 to 10000 (0 = all points)	0 to 50000
<point> :=	0 to 10000	0 to 50000
<segment> :=	0 to 50 (0 = all segments)	0 to 200

Note 1: After power-on, all values are set to 0 (i.e. entire waveforms will be transmitted without sparsing).

Note 2: Parameters are grouped in pairs. The first one names the variable to be modified and the second one gives the new value to be assigned. Pairs may be given in any order and may be restricted to those variables to be changed.

QUERY SYNTAX

WaveForm_SetUp?

Response format

WaveForm_SetUp SP,<sparsing>, NP,<number>,
FP,<point>,SN,<segment>

EXAMPLE (GPIB)

The following command specifies that every 3rd data point (SP=3) starting at address 200 should be transferred.

```
CMD$="WFSU SP,3,FP,200"
CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

INSPECT?, WAVEFORM, TEMPLATE

WAVEFORM TRANSFER

WAVEFORM_TEXT, WFTX

Command/Query

DESCRIPTION

The WAVEFORM_TEXT command is used to document the conditions under which a waveform has been acquired. The text buffer is limited to 400 characters.

The WAVEFORM_TEXT? query returns the text section of the specified trace.

COMMAND SYNTAX

<trace>:WaveForm_TeXt '<text>'

<trace> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<text> := An ASCII message (max. 400 characters long)

QUERY SYNTAX

<trace>:WaveForm_TeXt?

Response format

<trace>:WaveForm_TeXt "<text>"

EXAMPLE (GPIB)

The following example shows how to document Function E (FE).

MSG\$ = "'Averaged pressure signal.

Experiment carried out Oct. 15, 88'"

CMD\$ = "FE:WFTX "+ MSG\$

CALL IBWRT(SCOPE%,CMD\$)

RELATED COMMANDS

INSPECT?, WAVEFORM, TEMPLATE?

‡ 4-channel oscilloscopes only

DISPLAY**XY_ASSIGN, XYAS**

Command/Query

DESCRIPTION

The XY_ASSIGN command assigns traces to the X and Y axis to create an X versus Y display.

The XY_ASSIGN? query returns the traces currently assigned to the XY display. If there is no trace assigned to the X-axis and/or the Y-axis the value UNDEF will be returned instead of the trace name.

COMMAND SYNTAX

XY_ASsign <X_source>,<Y_source>

<X_source> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<Y_source> := {EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

QUERY SYNTAX

XY_ASsign?

Response Format

XY_ASsign <X_source>,<Y_source>

<X_source> := {UNDEF, EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

<Y_source> := {UNDEF, EA, EB, MC, MD, FE, FF, C1, C2, C3‡, C4‡}

EXAMPLE (GPIB)

The following command will assign Channel 1 to X and Channel 2 to Y.

```
CMDSS$="XYAS C1,C2": CALL IBWRT(SCOPE%,CMDSS$)
```

‡ 4-channel oscilloscopes only

CURSOR

XY_CURSOR_ORIGIN, XYCO

Command/Query

DESCRIPTION

The **XY_CURSOR_ORIGIN** command sets the position of the origin for absolute time cursor measurements on the XY display.

Absolute time cursor values may be measured either with respect to the point (0,0) volts (**OFF**) or with respect to the center of the XY grid (**ON**).

The **XY_CURSOR_ORIGIN** query returns the current assignment of the origin for absolute time cursor measurements.

COMMAND SYNTAX

XY_Cursor_Origin <mode>

<mode> := {**ON**, **OFF**}

QUERY SYNTAX

XY_Cursor_Origin?

Response Format

XY_Cursor_Origin <mode>

EXAMPLE (GPIB)

The following command sets the origin for absolute time cursor measurements to the center of the XY grid.

```
CMD5$="XYCO ON": CALL IBWRT(SCOPE%,CMD5$)
```

CURSOR**XY_CURSOR_SET, XYCS**

Command/Query

DESCRIPTION

The XY_CURSOR_SET command allows the user to position any one of the nine independent XY cursors at a given screen location. The positions of the cursors can be modified or queried even if the required cursor is not currently displayed or if the XY display mode is OFF.

The XY_CURSOR_SET? query indicates the current position of the cursor(s). The values returned are quoted relative to the original waveform (time or frequency domain).

Notation

HABS	Horizontal absolute
HREF	Horizontal reference
HDIF	Horizontal difference
XABS	Vertical absolute on X axis
XREF	Vertical reference on X axis
XDIF	Vertical difference on X axis
YABS	Vertical absolute on Y axis
YREF	Vertical reference on Y axis
YDIF	Vertical difference on Y axis

COMMAND SYNTAX

XY_Cursor_Set <cursor>,<position>

[<cursor>,<position><cursor>,<position>]

<cursor> := {HABS, HREF,HDIF, XABS, XREF,XDIF, YABS, YREF, YDIF}

<position> := 0 to 10 DIV (horizontal)

– 4 to 4 DIV (vertical)

Note 1: The suffix DIV is optional.

Note 2: Parameters are grouped in pairs. The first one names the cursor to be modified and the second one indicates its new value. Pairs may be given in any order and may be restricted to those items to be changed.

QUERY SYNTAX

XY_Cursor_Set? [<cursor>...<cursor>]

<cursor> := {HABS, HREF, HDIF, XABS, XREF, XDIF, YABS, YREF, YDIF, ALL}

Note: If <cursor> is not specified, ALL will be assumed.

Response Format

XY_Cursor_Set <cursor>,<position>[<cursor>,<position>...
...,<cursor>,<position>]

EXAMPLE (GPIB)

The following command positions the XREF and YDIF at +3 DIV and -2 DIV respectively.

```
CMDS$="XYCS XREF,3DIV,YDIF,-2DIV"
```

```
CALL IBWRT(SCOPE%,CMDS$)
```

CURSOR**XY_CURSOR_VALUE, XYCV**

Command/Query

DESCRIPTION

The XY_CURSOR_VALUE? query returns the current values of the X versus Y trace parameters. The X versus Y trace does not need to be displayed to obtain these parameters, but valid sources must be assigned to the X and Y axes.

Parameter Names

<cursor type>_X	X
<cursor type>_Y	Y
<cursor type>_RATIO	$\Delta Y/\Delta X$
<cursor type>_PROD	$\Delta Y * \Delta X$
<cursor type>_ANGLE	$\text{arc tan}(\Delta Y/\Delta X)$
<cursor type>_RADIUS	$\text{sqrt}(\Delta X^2 + \Delta Y^2)$
<cursor_type> :=	[HABS, HREL, VABS, VREL]

QUERY SYNTAX

XY_Cursor_Value? [<parameter>,...<parameter>]

<parameter> := {HABS_X, HABS_Y, HABS_RATIO, HABS_PROD, HABS_ANGLE, HABS_RADIUS, HREL_X, HREL_Y, HREL_RATIO, HREL_PROD, HREL_ANGLE, HREL_RADIUS, VABS_X, VABS_Y, VABS_RATIO, VABS_PROD, VABS_ANGLE, VABS_RADIUS, VREL_X, VREL_Y, VREL_RATIO, VREL_PROD, VREL_ANGLE, VREL_RADIUS, ALL}

Note: If <parameter> is not specified or equals ALL, all the measured cursor values are returned. If the value of a cursor could not be determined in the current environment, the value UNDEF will be returned. If no trace has been assigned to either the X axis or the Y axis, an environment error will be generated.

Response Format

XY_Cursor_Value <parameter>,<value>
[...<parameter>,<value>]
<value> := decimal value or UNDEF

EXAMPLE (GPIB)

The following query reads the ratio of the absolute horizontal cursor, the angle of the relative horizontal cursor and the product of the absolute vertical cursor.

```
CMDSS="XYCV? HABS_RATIO,HREL_ANGLE,
VABS_PROD
```

```
CALL IBWRT(SCOPE%,CMDSS)
```

DISPLAY

XY_DISPLAY, XYDS

Command/Query

DESCRIPTION

The XY_DISPLAY command enables or disables the XY display mode.

The XY_DISPLAY? query returns the current mode of the XY display.

COMMAND SYNTAX

XY_DiSplay <mode>
<mode> := {ON, OFF}

QUERY SYNTAX

XY_DiSplay?

Response Format

XY_DiSplay <mode>

EXAMPLE (GPIB)

The following command turns the XY display ON.
CMDSS\$="XYDS ON": CALL IBWRT(SCOPE%,CMDSS\$)

DISPLAY**ZOOM**

Command/Query

DESCRIPTION

The ZOOM command allows the user to select which trace is to be expanded by one of the expansion functions.

The response to the ZOOM? query indicates which trace is currently expanded.

In 4-channel oscilloscopes the ZOOM command automatically switches the function to the expand state.

COMMAND SYNTAX

<exp_trace>:ZOOM <trace>

<exp_trace> := {EA, EB, MC‡, MD‡}

<trace> := {MC, MD, FE, FF, C1, C2, C3‡, C4‡}

QUERY SYNTAX

<exp_trace>:ZOOM?

Response format

<exp_trace>:ZOOM <trace>

EXAMPLE (GPIB)

The following example selects Memory C (MC) as the source for Expand B (EB).

```
CMD$="EB:ZOOM MC": CALL IBWRT(SCOPE%,CMD$)
```

RELATED COMMANDS

DUAL_ZOOM

‡ 4-channel oscilloscopes only

REMOTE CONTROL INDEX OF GPIB COMMANDS

<u>Full Command</u>	<u>Abbrev.</u>	<u>Page</u>	<u>Full Command</u>	<u>Abbrev.</u>	<u>Page</u>
ALL_STATUS?	ALST?	39	PASS_FAIL_CONDITION	PFCO	116
ARM_ACQUISITION	ARM	40	PASS_FAIL_COUNTER	PFCT	117
ATTENUATION	ATTN	41	PASS_FAIL_DO	PFDO	118
AUTO_CALIBRATE	ACAL	42	PASS_FAIL_MASK	PFMS	119
AUTO_SETUP	ASET	43	PERSIST	PERS	120
BANDWIDTH_LIMIT	BWL	44	PER_CURSOR_SET	PECS	121-122
BUZZER	BUZZ	45	PER_CURSOR_VALUE?	PECV?	123
*CAL?	*CAL?	46	PERSIST_SETUP	PESU	124
CALL_HOST	CHST	47	*PRE	*PRE	125
*CLS	*CLS	48	*RCL	*RCL	126
CMR?	CMR?	49	RECALL_PANEL	REC	127
COMM_FORMAT	CFMT	50-51	RECALL	RCPN	128
COMM_HEADER	CHDR	52	*RST	*RST	129
COMM_HELP	CHL	53	SAMPLE_CLOCK	SCLK	130
COMM_ORDER	CORD	54-55	*SAV	*SAV	131
COMM_RS232	CORS	56-58	SCREEN_DUMP	SCD	132
COUPLING	CPL	59	SEGMENTS	SEGS	133
CURSOR_MEASURE	CRMS	60	SELECT	SEL	134
CURSOR_SET	CRST	61-62	*SRE	*SRE	135
CURSOR_VALUE?	CRVA?	63	*STB?	*STB?	136-137
DATE	DATE	64	STOP	STOP	138
DDR?	DDR?	65-66	STORE	STO	139
DEFINE	DEF	67-75	STORE_PANEL	STPN	140
DELETE_FILE	DELF	76	STORE_SETUP	STST	141-142
DIRECTORY_LIST?	DIR?	77	STORE_TEMPLATE	STTM	143
DISPLAY	DISP	78	TEMPLATE?	TMLP?	144
DUAL_ZOOM	DZOM	79	TIME_DIV	TDIV	145-146
*ESE	*ESE	80	TRACE	TRA	147
*ESR?	*ESR?	81-82	*TRG	*TRG	148
EXR?	EXR?	83-84	TRIG_COUPLING	TRCP	149
FORMAT_CARD	FCRD	85-86	TRIG_DELAY	TRDL	150
FUNCTION_RESET	FRST	87	TRIG_LEVEL	TRLV	151
FUNCTION_STATE	FSTA	88-89	TRIG_MODE	TRMD	152
GRID	GRID	90	TRIG_PATTERN (+ ext. trig.)	TRPA	153
HARDCOPY_SETUP	HCSU	91-92	TRIG_PATTERN (- ext. trig.)	TRPA	154
HARDCOPY_TRANSMIT	HCTR	93	TRIG_SELECT	TRSE	155-157
HOR_MAGNIFY	HMAG	94	TRIG_SLOPE	TRSL	158
HOR_POSITION	HPOS	95-96	*TST?	*TST?	159
*IDN?	*IDN?	97	URR?	URR?	160
INE	INE	98	VERT_MAGNIFY	VMAG	161
INR?	INR?	99	VERT_POSITION	VPOS	162
INSPECT?	INSP?	100-101	VOLT_DIV	VDIV	163
INTENSITY	INTS	102	*WAI	WAI	164
INTERLEAVED	ILVD	103	WAIT	WAIT	165
*IST?	*IST?	104	WAVEFORM	WF	166-167
KEY	KEY	105-106	WAVEFORM_SETUP	WFSU	168-169
MESSAGE	MSG	107	WAVEFORM_TEXT	WFTX	170
MULTI_ZOOM	MZOM	108	XY_ASSIGN	XYAS	171
OFFSET	OFST	109-110	XY_CURSOR_ORIGIN	XYCO	172
*OPC	*OPC	111	XY_CURSOR_SET	XYCS	173-174
*OPT?	*OPT?	112	XY_CURSOR_VALUE	XYCV	175
PANEL_SETUP	PNSU	113	XY_DISPLAY	XYDS	176
PARAMETER_VALUE?	PAVA?	114-115	ZOOM	ZOOM	177

INTRODUCTION

This section discusses how to read and write waveforms and understand their contents. Waveforms can be divided into two basic entities, the basic data array (i.e. the raw data values from the ADC's in the acquisition) and the accompanying descriptive information, such as vertical scale, horizontal scale, time of day, which are necessary for a full understanding of the data.

The information in a waveform can be accessed using the `IN-SPECT?` query which interprets it in an easily understood ASCII text form. It can also be more rapidly transferred using the `WAVEFORM?` query or written back into the instrument with the `WAVEFORM` command. The oscilloscope contains a data structure called the template which is a detailed description of how the waveform's information is organized.

LOGICAL DATA BLOCKS OF A WAVEFORM

The template gives a detailed description of the form and contents of the logical data blocks of a waveform. It is provided as a reference to be used by you and your programs. A sample template is given in Appendix B although you are encouraged to use the `TEMPLATE?` query to examine the actual template that your instrument is using. The template may change as the instrument's firmware is enhanced. The template will help provide backward compatibility for the interpretation of waveforms.

Usually, a waveform will contain just a Waveform descriptor block (1.) and a Data array block (5.). In more complicated cases one or more of the other blocks will be present. The data blocks are:

1. **Waveform descriptor block (WAVEDESC)**. This block includes all the information necessary to reconstitute the display of the waveform from the data. This includes:
 - hardware settings at the time of acquisition
 - the exact time of the event
 - the kinds of processing that have been performed
 - the name and serial number of the instrument
 - the encoding format used for the data blocks
 - miscellaneous constants
2. **An optional user-provided text (USERTEXT)**. The `WFTX` command can be used to put a title or description of a waveform into this block. The `WFTX?` query command gives an alternative way to read it. This text block can hold up to 400 characters. However, you should limit the length of each line to about 45 characters, otherwise the text will be wrapped onto the next line after 56 characters.
3. **A block of sequence acquisition times (TRIGTIME)**. This block is needed for sequence acquisitions to record the exact

timing information for each segment. It contains the time of each trigger relative to the trigger of the first segment, as well as the time of the first data point of each segment relative to its trigger.

4. **A block of random interleaved sampling times (RISTIME).** This block is needed for RIS acquisitions to record the exact timing information for each segment.
5. **A data array block (SIMPLE or DATA_ARRAY_1).** This is the basic integer data of the waveform. It can be raw or corrected ADC data or the integer result of waveform processing.
6. **A second data array block (DATA_ARRAY_2).** This second data array is needed to hold the results of processing functions such as the Extrema (WP01 option) or Complex FFT (WP02 option). In such cases, the data arrays contain:

	Extrema	FFT
DATA_ARRAY_1	Roof trace	Real part
DATA_ARRAY_2	Floor trace	Imaginary part

Note: The TEMPLATE also describes an array named DUAL. This is simply a way to allow the INSPECT? command to examine the two data arrays together.

INSPECT? COMMAND

This is the simplest way to examine the contents of a waveform. It can be used on both the data and descriptive parts. The simplest form of the command is:

```
INSPECT? "name"
```

where the template gives the name of a descriptor item or data block. The answer is returned as a single string, but may span many lines. Here is some typical dialogue:

```
question  C1:INSPECT? "VERTICAL_OFFSET"
response  C1:INSP "VERTICAL_OFFSET   : 1.5625e-03   "
question  C1:INSPECT? "TRIGGER_TIME"
response  C1:INSP "
          TRIGGER_TIME : Date = FEB 17, 1989, Time = 4: 4:29.5580
          "
```

The INSPECT? command can also be used to get a readable translation of the full waveform descriptor block with the command:

```
INSPECT? "WAVEDESC"
```

The template dump from your instrument (or from Appendix B) will give details on the interpretation of each of the parameters.

The INSPECT? command is also used to examine the measured data values of a waveform. For an acquisition with 42 points we get:

INSPECT? "SIMPLE"

C1:INSP "

4.68749e-03	1.09375e-02	1.71875e-02	2.03125e-02	2.03125e-02	2.65625e-02
3.28125e-02	3.59375e-02	3.90625e-02	4.53125e-02	5.15625e-02	5.15625e-02
5.78125e-02	6.40625e-02	6.71875e-02	6.71875e-02	7.65625e-02	7.96875e-02
8.59375e-02	8.90625e-02	9.21875e-02	9.53125e-02	1.04687e-01	1.04687e-01
1.07812e-01	1.14062e-01	1.20312e-01	1.20312e-01	1.26562e-01	1.29688e-01
1.32812e-01	1.39062e-01	1.42187e-01	1.51562e-01	1.54687e-01	1.57812e-01
1.60938e-01	1.60938e-01	1.70312e-01	1.73437e-01	1.70312e-01	1.76563e-01

"

These numbers are the fully converted measurements in volts. Of course, when the data block contains thousands of items the string will contain many lines.

Depending on the application, you may prefer to have the data in its raw form as either a BYTE (8 bits) or a WORD (16 bits) for each data value. In this case you must use the relations given below in association with the WAVEFORM? command to interpret the measurement. The command might then say:

INSPECT? "SIMPLE",BYTE

The examination of data values for waveforms with two data arrays can be done as follows:

INSPECT? "DUAL" to get pairs of data values on a single line

INSPECT? "DATA_ARRAY_1"

to get the values of the first data array

INSPECT? "DATA_ARRAY_2"

to get the values of the second data array

It is also possible to examine just a part of the waveform or a sparsed form of the waveform. This is controlled with the WAVEFORM_SETUP command mentioned later in this section.

The INSPECT? command has only a query form. It cannot be used to send a waveform back into the oscilloscope. It is also a very verbose way in which to send the information and is not very fast. Users who need speed or the ability to send the waveform back to the instrument should use the WAVEFORM commands.

BASIC users might find it convenient to combine the capabilities of the inspect facility with the waveform query command in order to construct files containing a human and BASIC readable version of the waveform descriptor together with the full waveform in a format suitable for retransmission to the instrument. This can be

done for a waveform in a memory location by sending the command:

```
MC:INSPECT? "WAVEDESC";WAVEFORM?
```

and putting the response directly into a disk file.

WAVEFORM? COMMAND

The WAVEFORM commands are an efficient way to transfer waveform data using the block formats defined in the IEEE-488.2 standard. You have the possibility of reading all of the logical blocks of the waveform with a single query:

```
C1:WAVEFORM?
```

This is the preferred form for most applications since it is complete and the response can be downloaded back into the instrument using the WAVEFORM command. You can also choose to read any single block with a query like:

```
C1:WAVEFORM? DAT1
```

This can save time and space when you need to read many waveforms all with the same acquisition conditions or if you are only interested in lots of raw integer data. Consult the description of the WAVEFORM command in Section 5 for the names of the various blocks.

Please be aware that a waveform query response can easily be a block containing over 200,000 bytes if it is in binary format and twice as much if the HEX option is used.

Interpreting the waveform descriptor

The binary response to a query command of the form:

```
C1:WAVEFORM? or C1:WAVEFORM? ALL
```

can be put into a disk file and then dumped to show the following hexadecimal and ASCII form: (This was done over GPIB with default settings)

Byte offset	Binary contents in hexadecimal	ASCII translation (b is for uninteresting)
0	4331 CA57 4620 414c 4c2c 2339 3030 3030	C1:WF ALL,#90000
16	3030 3433 3057 4156 4544 4553 4300 0000	00430WAVEDESCbbb
32	0000 0000 004c 4543 524f 595f 315f 3100	bbbbbbLECROY_1_1b
48	0000 0000 0000 0100 0000 0001 5a00 0000	bbbbbbbbbbbbbbbb
64	0000 0000 0000 0000 0000 0000 0000 0000	bbbbbbbbbbbbbbbb
80	0000 0000 5400 0000 0000 0000 0000 0000	bbbbbbbbbbbbbbbb
96	004c 4543 524f 5939 3435 305f 0000 0000	bLECROY0450_bbbb
112	0005 a1f4 b100 0000 0000 0000 0000 0000	...
128	0000 0000 0000 0000 0000 0000 2a00 0000	
144	2800 0000 0000 0000 2900 0000 0000 0000	
160	0100 0000 0000 0000 0100 0000 0100 0000	
176	0037 4ccc cd3a cccd 0046 fe00 00c7 0000	
192	0000 0800 0031 2bcc 77be 49fe 783b e800	
208	00be 4579 8ee0 0000 0056 0000 0000 0000	
224	0000 0000 0000 0000 0000 0000 0000 0000	
240	0000 0000 0000 0000 0000 0000 0000 0000	
256	0000 0000 0000 0000 0053 0000 0000 0000	
272	0000 0000 0000 0000 0000 0000 0000 0000	
288	0000 0000 0000 0000 0000 0000 0000 0000	
304	0000 0000 0000 0000 0000 0000 0040 3d8e	
320	d913 4ab0 0004 0401 0407 d200 0000 0000	
336	0000 0000 0000 0000 0000 0c00 003f 8000	
352	0000 0f00 003f 8000 003c 8000 2000 0002	
368	0004 0006 0007 0007 0009 000b 000c 000d	
384	000f 0011 0011 0013 0015 0016 0016 0019	
400	001a 001c 001d 001e 001f 0022 0022 0023	
416	0025 0027 0027 0029 002a 002b 002d 002e	
432	0031 0032 0033 0034 0034 0037 0038 0037	
448	0039 000a	

It can be seen that the first 10 bytes translate into ASCII and look like the simple beginning of a query response. This is followed by the string "#9000000430". This is the beginning of a binary block where 9 ASCII integers are used to give the length of the block (430 bytes). The waveform itself starts immediately after this at byte number 21 (the first byte is byte 0).

Deciphering the waveform descriptor can be done with the aid of the template (see Appendix B). It states that the first object is a DESCRIPTOR_NAME which is a string of 16 characters with the value WAVEDESC and this is what we see. At byte 16 relative to the beginning of the descriptor (or byte 37 above) we find the next string, the TEMPLATE_NAME with the value LECROY_1_1. Several other parameters follow. We can easily recognize the IN-

STRUMENT_NAME at 76 bytes from the descriptor start (or byte 97 above).

In a similar way we learn that a 4 byte long integer giving the length of the descriptor starts at byte 36 (or byte 57 above):

WAVE_DESCRIPTOR = 15a (hex) = 346

At byte 60 (or byte 81 above) we find another 4 byte integer giving the length of the data array:

WAVE_ARRAY_1 = 54 (hex) = 84

and at byte 116 (or byte 137 above) the number of data points:

WAVE_ARRAY_COUNT = 2a (hex) = 42

Now we know that the data will start at byte 346 from the beginning of the descriptor (or byte 367 above) and that each of the 42 data points will be represented by two bytes. The waveform has a total length of $346 + 84$ which is the same as the ASCII string told us at the beginning of the block. The final 0a at byte 451 is the NL character associated with the GPIB message terminator <NL><EOI>.

The data can be easily seen starting at byte 367 above. Since the example was taken using an oscilloscope with an 8-bit ADC we see those 8 bits followed by a 0 byte for each data point. It should be noted that for oscilloscopes with 10-bit ADCs or for many other kinds of waveform this second byte will not be zero and contains interesting information. The data is coded in signed form (two's complement) with values ranging from $-32768 = 8000$ (hex) to $32767 = 7fff$ (hex). If we had chosen to use the BYTE option for the data format the values would have been signed integers in the range $-128 = 80$ (hex) to $127 = 7f$ (hex).

Interpreting the waveform vertical data

Now that we know how to decipher the data it would be useful to convert it to the appropriate measured values. The vertical reading for each data point depends on the vertical gain and the vertical offset given in the descriptor. For acquisition waveforms this corresponds to the volts/div and voltage offset selected after conversion for the data representation being used. The template tells us that the vertical gain and offset can be found at bytes 156 and 160 respectively of the descriptor and that they are stored as floating point numbers in the IEEE 32 bit format. An ASCII string giving the vertical unit is to be found in VERTUNIT, byte 196. The vertical value is given by the relationship:

$$\text{value} = \text{VERTICAL_GAIN} * \text{data} - \text{VERTICAL_OFFSET}$$

A data value of 0 is normally displayed as a point in the middle of the grid.

In the case of the data shown above we find:

VERTICAL_GAIN = 1.22070314e-05 from the floating point number 374c cccd at byte 177

VERTICAL_OFFSET = 1.56250596e-03 from the floating point number 3acc cd00 at byte 181

VERTICAL_UNIT = V = volts from the string 5600 ... at byte 217

and therefore:

since data[0] = 512 from the hexadecimal word 0200 at byte 367

value[0] = 0.00468 V as stated in the inspect command above

and

since data[1] = 1024 from the hexadecimal word 0400 at byte 369

value[0] = 0.0109 V as stated in the inspect command above.

If your computer or available software is incapable of understanding the IEEE floating point values you can find a description of this format in the template (see Appendix B).

The data values in a waveform may not all correspond to measured points. The parameters, FIRST_VALID_PNT and LAST_VALID_PNT give the necessary information. The descriptor also records the SPARSING_FACTOR, the FIRST_POINT, and the SEGMENT_INDEX to aid interpretation if the options of the WAVEFORM_SETUP command have been used.

For sequence acquisitions the data values for each segment are given in their normal order and the segments are read out one after the other. The important descriptor parameters are the WAVE_ARRAY_COUNT and the SUBARRAY_COUNT, giving the total number of points and the number of segments.

For waveforms such as the extrema and the complex FFT there will be two arrays one after the other for the two arrays of the result.

Calculating the horizontal position of a data point

Each vertical data value has a corresponding horizontal position, usually measured in time or in frequency units. The calculation of this position depends on the type of waveform being examined. We will treat separately the single sweep, the sequence, and the interleaved (RIS) waveform. Each data value has a position, i , in

the original waveform with $i = 0$ corresponding to the first data point acquired. The descriptor parameter `HORUNIT` gives a string with the name of the horizontal unit.

- **Single-sweep waveforms**

$$x[i] = \text{HORIZ_INTERVAL} * i + \text{HORIZ_OFFSET}$$

For acquisition waveforms this time is from the trigger to the data point in question. It will be different from acquisition to acquisition since the `HORIZ_OFFSET` is measured for each trigger.

In the case of the data shown above this means:

`HORIZ_INTERVAL` = 2.5000e-09 from the floating point number 312b cc77 at byte 197

`HORIZ_OFFSET` = -1.21044098e-08 from the double precision floating point number be49 fe78 3be8 0000 at byte 201

`HORUNIT = S` = seconds from the string 5300 ... at byte 265

which gives

$$x[0] = -1.210e-08 \text{ S}$$

$$x[1] = -0.960e-08 \text{ S}$$

- **Sequence waveforms**

Since sequence waveforms are really many independent acquisitions, each segment will have its own horizontal offset. These can be found in the `TRIGTIME` array. For the n 'th segment

$$x[i,n] = \text{HORIZ_INTERVAL} * i + \text{TRIGGER_OFFSET}[n]$$

The `TRIGTIME` array can contain up to 200 segments of timing information with two 8 byte double precision floating point numbers for each segment.

- **Interleaved (RIS) waveforms**

These waveforms are composed of many acquisitions interleaved together. The descriptor parameter, `SWEEPS_PER_ACQ` gives the number of acquisitions. The i 'th point will belong to the m 'th segment where

$$m = i \text{ modulo } (\text{SWEEPS_PER_ACQ})$$

will have a value between 0 and `SWEEPS_PER_ACQ` - 1.

Then with

$$j = i - m$$

$$x[i] = x[j,m] = \text{HORIZ_INTERVAL} * j + \text{RIS_OFFSET}[m]$$

where the `RIS_OFFSET`'s can be found in the `RISTIME` array. There can be up to 100 8 byte double precision floating point numbers in this block. The instrument tries to get segments with times such that

$$\text{RIS_OFFSET}[i] \approx \text{PIXEL_OFFSET} + (i - 0.5) * \text{HORIZ_INTERVAL}$$

Thus, taking as an example a RIS with `SWEEPS_PER_ACQ` = 10 `HORIZ_INTERVAL` = 1 ns and `PIXEL_OFFSET` = 0.0, we might find for a particular event that:

<code>RIS_OFFSET</code> [0] = -0.5 ns	<code>RIS_OFFSET</code> [1] = 0.4 ns
<code>RIS_OFFSET</code> [2] = 1.6 ns	<code>RIS_OFFSET</code> [3] = 2.6 ns
<code>RIS_OFFSET</code> [4] = 3.4 ns	<code>RIS_OFFSET</code> [5] = 4.5 ns
<code>RIS_OFFSET</code> [6] = 5.6 ns	<code>RIS_OFFSET</code> [7] = 6.4 ns
<code>RIS_OFFSET</code> [8] = 7.6 ns	<code>RIS_OFFSET</code> [9] = 8.5 ns

and therefore:

$$\begin{aligned} x[0] &= \text{RIS_OFFSET}[0] &= &-0.5 \text{ ns} \\ x[1] &= \text{RIS_OFFSET}[1] &= &0.4 \text{ ns} \end{aligned}$$

...

$$\begin{aligned} x[9] &= \text{RIS_OFFSET}[9] &= &8.5 \text{ ns} \\ x[10] &= 1 \text{ ns} * 10 + (-0.5) &= &9.5 \text{ ns} \\ x[11] &= 1 \text{ ns} * 10 + 0.4 &= &10.4 \text{ ns} \end{aligned}$$

...

$$\begin{aligned} x[19] &= 1 \text{ ns} * 10 + 8.5 &= &18.5 \text{ ns} \\ x[20] &= 1 \text{ ns} * 20 + (-0.5) &= &19.5 \text{ ns} \end{aligned}$$

...

WAVEFORM COMMAND

Waveforms that have been read in their entirety with the `WAVEFORM?` command can be sent back into the instrument. Since the descriptor contains all of the necessary information, you do not have to be careful about any of the communication format parameters. The instrument can learn all that it needs to know from the waveform.

If you want to synthesize waveforms for display or comparison purposes, you are encouraged to read out a waveform of the appropriate size and then replace the data with the desired values. This will assure that the descriptor is coherent.

Note: You are only allowed to send back waveforms to memory traces (MC or MD for 2-channel instruments or MC, MD, FE and FF for 4-channel instruments). This means that you may have to remove or change the prefix (C1 or CHANNEL_1) in the response to the WF? query. The examples for the WF command in Section 5 show how this can be done.

MORE CONTROL OF WAVEFORM QUERIES

There are many different ways for you to use the WAVEFORM? command which may simplify or speed up your work. Among them are:

- **Partial readout of waveform**

The WAVEFORM_SETUP command allows you to specify a short part of a waveform for readout. It also lets you select a sparsing factor to read only every n'th data point.

- **Byte swapping**

The COMM_ORDER command allows you to swap the two bytes of a 16-bit word. In fact, byte swapping is done for all numbers represented by more than one byte. This is the case for the descriptor, the time blocks, and WORD arrays, thereby simplifying data interpretation for some computer systems (e.g. INTEL based, or DEC)

- **Data length, block format, and encoding**

The COMM_FORMAT command gives you control over these parameters. If you do not need the extra precision of the lower order byte of the standard data value, the BYTE option lets you save a factor of two on the amount of data to be transmitted or stored. If your computer is not able to read binary data, the HEX option allows a response form where the value of each byte is given by a pair of hexadecimal digits.

- **Data only transfers**

The COMM_HEADER OFF mode will allow you to get a response to WF? DAT1 with the data only (the C1:WF DAT1 will disappear).

If you have also specified COMM_FORMAT OFF,BYTE,BIN, you will just get a response of data bytes (the #90000nnnnn will disappear).

- **Formatting for RS-232 users**

The COMM_RS232 command can help you by splitting the very long WF? response into individual lines

HIGH-SPEED WAVEFORM TRANSFER

In order to achieve the maximum continuous data transfer rates from the oscilloscope to your instrument you will have to optimize many factors. The single most important point is to limit the work done in your computer. This means avoiding having to write the data to disk, minimizing the per data point computations, minimiz-

ing the number of calls to the IO system, etc. You can let the instrument help by reducing the number of points to be transferred and the number of data bytes per point. The pulse parameter capability and the processing functions can save you lots of computing and lots of data transfer time if employed creatively. Two other very important principles are:

- Try to overlap waveform acquisition with waveform transfer. The oscilloscope is capable of transferring an already acquired or processed waveform after a new acquisition has been started. This can also considerably increase the total time that the oscilloscope will be able to acquire events if it has to wait for triggers (livelime).
- Minimize the number of waveform transfers by using the sequence mode to accumulate many triggers for each transfer. This is preferable to using the WAVEFORM_SETUP command to reduce the number of data points to be transferred. It also reduces the oscilloscope transfer overhead significantly.

Here is an example of the type of commands to be given:

```
ARM                to acquire the first event or sequence
WAIT;ARM;C1:WF?   to wait for the event, start the next acquisition and then transfer the data.
```

This second line can be repeated by your program as soon as it has finished reading the waveform.



An extensive set of status registers allows the user to quickly determine the oscilloscope's internal processing status at any time. The status registers as well as the status reporting system have been designed to comply with IEEE 488.2 recommendations.

Related functions are grouped together in common status registers. Some, such as the Status Byte Register (STB) or the Standard Event Status Register (ESR), are required by the IEEE 488.2 standard. However, other registers are device specific. They include the Command Error Register (CMR) or the Execution Error Register (EXR). Commands associated with IEEE 488.2 mandatory status registers are preceded by an asterisk <*>.

OVERVIEW OF STATUS AND SERVICE REQUEST REPORTING


Figure 2 shows the organization of the status registers. The central reporting structure is the Status Byte Register (STB). It consists of 8 bits, three of which are not used.

The Standard Event Status Bit (ESB) and the Internal Status Change Bit (INB) in the Status Byte Register are summary bits of the Standard Event Status Register (ESR) and the Internal State Change Register (INR). The Message Available Bit (MAV) is set whenever there are data bytes in the output queue. The Value Adapted Bit (VAB) indicates that a parameter value was adapted during a previous command interpretation (e.g. if the command "TDIV 2.5 US" is received, the time base is set to 2 μ sec/div and the VAB bit is set).

The Master Summary Status bit (MSS) indicates that the instrument requests service. The MSS bit can only be set if any of the other bits of STB are enabled with the Service Request Enable Register (SRE).

All Enable registers (SRE, ESE and INE) are used to generate a bitwise AND with their associated status registers. The logical OR of this operation is reported to the STB register. At power-on, all Enable registers are zero, inhibiting any reporting to the STB.

The Standard Event Status Register (ESR) mostly summarizes errors, whereas the Internal State Change Register (INR) reports internal changes in the instrument. Additional details of the errors reported by ESR can be obtained with the queries "CMR?", "DDR?", "EXR?" and "URR?".



STATUS REGISTER STRUCTURE

Figure 2

The register structure contains one more register (not shown in Figure 2). It is the Parallel Poll Enable Register (PRE) which acts exactly like the Service Request Enable Register (SRE), but it sets the “ist” bit (not shown in Figure 2), used in the Parallel Poll. The “ist” bit can also be read with the “*IST?” query.

Example of status reporting

If an erroneous remote command, e.g. “TRIG_MAKE SINGLE”, is transmitted to the instrument, it rejects the command and sets the Command Error Register (CMR) to the value 1 (unrecognized command/query header). The non-zero value of CMR is reported to bit 5 of the Standard Event Status Register (ESR) which is then set.

Nothing further happens unless the corresponding bit 5 of the Standard Event Status Enable Register (ESE) is set (with the command “*ESE 32”), enabling the fact that bit 5 of ESR is set to be reported to the summary bit ESB of the Status Byte Register (STB).

If setting of the ESB summary bit in STB is enabled, again nothing happens unless further reporting is enabled by setting the corresponding bit in the Service Request Enable Register (with the command “*SRE 32”). In this case, the generation of a non-zero value of CMR ripples through to the Master Summary Status bit (MSS), generating a Service Request (SRQ).

The value of CMR can be read and simultaneously reset to zero at any time with the command “CMR?”. The occurrence of a command error can also be detected by analyzing the response to “*ESR?”. However, if several types of potential errors must be surveyed, it is usually much more efficient to enable propagation of the errors of interest into the STB with the enable registers ESE and INE.

Summary

A command error (CMR) sets bit 5 of ESR:

- if bit 5 of ESE is set, ESB of STB is also set.
- if bit 5 of SRE is set, MSS/RQS of STB is also set and a Service Request is generated.

STATUS BYTE REGISTER (STB)

The Status Byte Register is the instrument’s central reporting structure. The STB is composed of 8 single-bit summary messages (of which 3 are unused) which reflect the current status of the associated data structures implemented in the instrument.

Bit 0 is the summary bit INB of the Internal State Change Register. It is set if any of the bits of the INR are set, provided that they are enabled by the corresponding bit of the INE register.

Bit 2 is the Value Adapted Bit, indicating that a parameter value was adapted during a previous command interpretation.

Bit 4 is the Message Available (MAV) bit, indicating that the interface output queue is not empty.

Bit 5 of STB is the summary bit ESB of the Standard Event Status Register. It is set if any of the bits of the ESR are set, provided that they are enabled by the corresponding bit of the ESE register.

Bit 6 of the Status Byte Register (STB) is alternatively called the Master Summary Status bit (MSS) or the Request for Service bit (RQS) because the STB can be read in two different ways. The command “*STB?” reads and clears the STB in the query mode, in which case bit 6 of the STB is the MSS bit, indicating if the instrument has any reason for requesting service. The other way of reading the STB is the serial poll (see Section 3, page 22, for the GPIB serial poll procedure). In this case, bit 6 of the STB is the RQS bit, indicating that the instrument has actually activated the SRQ line on the GPIB. The serial poll only clears the RQS bit. Therefore, the MSS bit of the STB (and any other bits which caused MSS to be set) will stay set after a serial poll. The controller must reset these bits.

The Status Byte Register may be read via the query “*STB?”. The response represents the binary weighted sum of the register bits. The register is cleared by “*STB?”, “ALST?”, “*CLS” or after the instrument has been powered up.

STANDARD EVENT STATUS REGISTER (ESR)

The ESR is a 16-bit register reflecting the occurrence of events. The register bit assignments have been standardized by IEEE 488.2. Only the lower 8 bits are currently in use.

The Standard Event Status Register may be read via the query “*ESR?”. The response is the binary weighted sum of the register bits. The register is cleared with an “*ESR?” or “ALST?” query, a “*CLS” command or after power-on.

Example

The response message “*ESR 160” indicates that a command error occurred and that the ESR is being read the first time after power-on. The value 160 can be broken down into 128 (bit 7) plus 32 (bit 5). See Table 5, page 82, for a description of the conditions corresponding to the bits set.

The “Power ON” bit appears only on the first “*ESR?” query after power-on because the query clears the register. The type of command error can be determined by reading the Command Error Status Register with the query “CMR?”. Note that it is not necessary to read (and simultaneously clear) this register in order to be able to set the CMR bit in the ESR on the next command error.

STANDARD EVENT STATUS ENABLE REGISTER (ESE)

The ESE allows one or more events in the Standard Event Status Register to be reported to the ESB summary bit in the STB.

The Standard Event Enable Register is modified with the command “*ESE”. It is cleared with the command “*ESE 0”, or after power-on. It may be read with the query “*ESE?”.

Example

“*ESE 4” sets bit 2 (i.e. binary 4) of the standard event enable register, enabling query errors to be reported.

SERVICE REQUEST ENABLE REGISTER (SRE)

The Service Request Enable Register specifies which summary bit(s) in the Status Byte Register will cause a service request. The Service Request Enable Register consists of 8 bits. Setting a bit in the register allows the summary bit located at the same bit position in the Status Byte Register to generate a service request provided that the associated event becomes true. Bit 6 (MSS) cannot be set and is always reported as zero in response to the query “*SRE?”.

The Standard Event Enable Register is modified with the command “*SRE”. It is cleared with the command “*SRE 0”, or after power-on. It may be read with the query “*SRE?”.

PARALLEL POLL ENABLE REGISTER (PRE)

The Parallel Poll Enable Register specifies which summary bit(s) in the Status Byte Register will set the “ist” individual local message. This register is quite similar to the Service Request Enable Register (SRE), but it is used to set the parallel poll “ist” bit rather than MSS.

The value of the “ist” may also be read without a Parallel Poll via the query “*IST?”. The response indicates if the “ist” message has been set or not (values are 1 or 0).

The Parallel Poll Enable Register is modified with the command “*PRE”. It is cleared with the command “*PRE 0”, or after power-on. It may be read with the query “*PRE?”. (See Section 3, page 23, for the GPIB parallel poll procedure.)

Example

“*PRE 5” sets bits 2 and 0 (decimal 4 and 1) of the Parallel Poll Enable Register.

INTERNAL STATE CHANGE STATUS REGISTER (INR)

The INR reports the completion of a number of internal operations. The events tracked by this 16-bit-wide register are listed with the command “INR?” in Section 5.

The Internal State Change Status Register may be read via the query "INR?". The response is the binary weighted sum of the register bits. The register is cleared with an "INR?" or "ALST?" query, a "*CLS" command or after power-on.

INTERNAL STATE CHANGE ENABLE REGISTER (INE)

The INE allows one or more events in the Internal State Change Status Register to be reported to the INB summary bit in the STB. The Internal State Change Enable Register is modified with the command "INE". It is cleared with the command "INE 0", or after power-on. It may be read with the query "INE?".

COMMAND ERROR STATUS REGISTER (CMR)

The Command Error Status register contains the code of the last command error detected by the instrument. Command error codes are listed with the command "CMR?" in Section 5.

The Command Error Status Register may be read via the query "CMR?". The response is the error code. The register is cleared with a "CMR?" or "ALST?" query, a "*CLS" command or after power-on.

DEVICE DEPENDENT ERROR STATUS REGISTER (DDR)

The DDR indicates the type of hardware errors affecting the instrument. Individual bits in this register report specific hardware failures. They are listed with the command "DDR?" in Section 5.

The Device Dependent Error Status Register may be read via the query "DDR?". The response is the binary weighted sum of the error bits. The register is cleared with a "DDR?" or "ALST?" query, a "*CLS" command or after power-on.

EXECUTION ERROR STATUS REGISTER (EXR)

The Execution Error Status Register contains the code of the last execution error detected by the instrument. Execution error codes are listed with the command "EXR?" in Section 5.

The Execution Error Status Register may be read via the query "EXR?". The response is the error code. The register is cleared with a "EXR?" or "ALST?" query, a "*CLS" command or after power-on.

USER REQUEST STATUS REGISTER (URR)

The URR contains the identification code of the last menu button which was pressed. The codes are listed with the command "URR?" in Section 5.

The User Request Status Register may be read via the query "URR?". The response is the decimal code associated with the selected menu button. The register is cleared with a "URR?" or "ALST?" query, a "*CLS" command or after power-on.



APPENDIX A

EXAMPLE 1: USE OF THE INTERACTIVE GPIB PROGRAM 'IBIC'

This example assumes the use of an IBM PC, or a compatible computer, equipped with a National Instruments GPIB interface card. It also assumes that the GPIB-driver is left in the default state so that the device name "dev4" corresponds to the GPIB address 4 which is assumed to be the address of the oscilloscope. All text entered by the user is underlined.

ibic<CR>

program announces itself

: ibfind<CR>

enter board/device name: dev4<CR>

dev4: ibwrt<CR>

enter string: "tdiv?"<CR>

[0100] (cmpl)

count: 5

dev4: ibrd<CR>

enter byte count: 10<CR>

[0100] (cmpl)

count: 10

54 44 49 56 20 35 30 45

2D 39

T D I V 5 0 E

- 9

dev4: ibwrt<CR>

enter string: "c1:cpl?"<CR>

[0100] (cmpl)

count: 7

dev4: ibrd<CR>

enter byte count: 20<CR>

[2100] (end cmpl)

count: 11

43 31 3A 43 50 4C 20 44

35 30 0A

C 1 : C P L D

5 0 ◆

dev4: q<CR>

for quitting the program

**EXAMPLE 2:
GPIB PROGRAM FOR
IBM PC (HIGH-LEVEL
FUNCTION CALLS)**

The following BASICA program allows full interactive control of the 9420/24/50 using an IBM PC as GPIB controller. It is again assumed that the controller is equipped with a National Instruments GPIB interface card. All the remote control commands listed in Section 5 can be used by simply entering the text string of the command, i.e. "c1:vdiv 50 mv" (without the quotes). The program automatically displays the information sent back by the oscilloscope in response to queries.

In addition, a few utilities have been provided for convenience. The commands ST and RC enable waveform data to be stored on or retrieved from disk if proper drive and file names are provided. The command LC returns the oscilloscope to local mode. Responses sent back by the oscilloscope are interpreted as character strings and are thus limited to a maximum of 255 characters.

Note 1: It is assumed that the National Instruments GPIB driver GPIB.COM is in its default state. This means that the interface board can be referred to by its symbolic name 'GPIB0' and that devices on the GPIB with addresses 1 to 16 can be called by the symbolic name 'DEV1' to 'DEV16'.

Note 2: Lines 1 - 99 are a copy of the file DECL.BAS supplied by National Instruments. The first 6 lines are required for the initialization of the GPIB handler. DECL.BAS requires access to the file BIB.M during the GPIB initialization. BIB.M is one of the files supplied by National Instruments, and must exist in the directory currently in use.

Note 3: The first 2 lines of DECL.BAS each contain a string "XXXXX" which must be replaced by the number of bytes which determine the maximum workspace for BASICA (computed by subtracting the size of BIB.M from the currently available space in BASICA). For example, if the size of BIB.M is 1200 bytes and when BASICA is loaded it reports "60200 bytes free", you would replace "XXXXX" by the value 59000 or less.

Note 4: The default timeout of 10 seconds is modified to 300 msec during the execution of this program. However, the default value of the GPIB handler is not changed. Whenever a remote command is entered by the user, the program sends it to the instrument with the function call IBWRT. Afterwards, it always executes an IBRD call, independently of whether or not a response is expected. If a response is received it is immediately displayed. If there is no response, the program waits until time-out and then asks for the next command.

```
1-99 <DECL.BAS>
100 CLS
110 PRINT "Control of the 9450 via GPIB and IBM PC"
115 PRINT ""
120 PRINT "Options :EX to exit   LC local mode"
125 PRINT "           ST store dataRC recall data"
130 PRINT ""
140 LINE INPUT "GPIB-address of oscilloscope (1...16)? :",ADDR$
145 DEV$ = "DEV" + ADDR$
150 CALL IBFIND(DEV$,SCOPE%)
155 IF SCOPE% < 0 THEN GOTO 830
160 TMO% = 10 'timeout = 300 msec (rather than default 10 sec)
165 CALL IBTMO(SCOPE%,TMO%)
170 '
200 LOOP% = 1
205 WHILE LOOP%
210     LINE INPUT "Enter command (EX --> Exit) : ",CMD$
220     IF CMD$ = "ex" OR CMD$ = "EX" THEN LOOP% = 0 : GOTO 310
230     IF CMD$ = "st" OR CMD$ = "ST" THEN GOSUB 600 : GOTO 300
240     IF CMD$ = "rc" OR CMD$ = "RC" THEN GOSUB 700 : GOTO 300
250     IF CMD$ = "lc" OR CMD$ = "LC" THEN GOSUB 400 : GOTO 300
260     IF CMD$ = "" THEN GOTO 300
270     CALL IBWRT(SCOPE%,CMD$)
275     IF IBSTA% < 0 THEN GOTO 840
280     GOSUB 500
300 WEND
310 GOSUB 400
320 END
400 '
405 'SUBROUTINE LOCAL_MODE
410 '
420 CALL IBLOC(SCOPE%)
425 PRINT ""
430 RETURN
500 '
505 'SUBROUTINE GET_DATA
510 'If there are no data to read, simply wait until timeout occurs
515 '
520 CALL IBRD(SCOPE%,RD$)
525 I = IBCNT% 'IBCNT% is the number of characters read
530 FOR J = 1 TO I
535     PRINT MID$(RD$,J,1);
540 NEXT J
545 PRINT ""
550 RETURN
600 '

```

```

605  `SUBROUTINE STORE_DATA
610  `
615  RD1$=SPACE$(3)
620  LINE INPUT "Specify trace (EA,EB,MC,MD,FE,FF,C1,C2): ",TRACE$
625  LINE INPUT "Enter filename : ",FILE$
630  CMD$="WFSU NP,0,SP,0,FP,0,SN,0; CHDR SHORT"
640  CALL IBWRT(SCOPE%,CMD$)
645  CMD$=TRACE$+":WF?"
650  CALL IBWRT(SCOPE%,CMD$)
660  CALL IBRD(SCOPE%,RD1$)           `Discard first 3 chars of response
665  CALL IBRDF(SCOPE%,FILE$)
670  IF IBSTA% < 0 THEN GOTO 840
675  PRINT ""
680  RETURN
700  `
705  `SUBROUTINE RECALL_DATA
710  `
715  LINE INPUT "Specify target memory (MC,MD): ",MEM$
720  LINE INPUT "Enter filename : ",FILE$
730  CMD$=MEM$+":TRACE ON"
735  CALL IBWRT(SCOPE%,CMD$)
740  CALL IBWRTF(SCOPE%,FILE$)
745  IF IBSTA% < 0 THEN GOTO 840
750  PRINT ""
755  RETURN
800  `
810  `ERROR HANDLER
820  `
830  PRINT "IBFIND ERROR"
835  END
840  PRINT "GPIB ERROR -- IBERR: ";IBERR%;"IBSTA: ";HEX$(IBSTA%)
845  END

```

**EXAMPLE 3:
GPIB PROGRAM FOR
IBM PC (LOW-LEVEL
FUNCTION CALLS)**

The following example has the same function as example 2, but it is written with low level function calls.

The program assumes that the controller (board) and oscilloscope (device) are at addresses 0 and 4 respectively. The decimal listener and talker addresses of the controller and the device thus are:

	Listener address	Talker address
controller	32 (ASCII <space>)	64 (ASCII @)
device	32+4=36 (ASCII \$)	64+4=68 (ASCII D).


```
1-99 <DECL.BAS>
100 CLS
110 PRINT "Control of the 9450 (address 4) via GPIB and IBM PC"
115 PRINT "": PRINT "Options :  EX to exit      LC local mode"
120 PRINT "          ST store data      RC recall data": PRINT""
125 LOOP=1
130 CMD1$ = "?_@$" 'Unlisten, Untalk, Board talker, Device listener
135 CMD2$ = "?_D" 'Unlisten, Untalk, Board listener, Device talker
140 BDNAME$= "GPIBO": CALL IBFIND(BDNAME$,BRDO%)
145 IF BRDO% < 0 THEN GOTO 420
150 CALL IBSIC(BRDO%): IF IBSTA% < 0 THEN GOTO 425
155 WHILE LOOP
160     LINE INPUT "Enter command (EX --> Exit) : ",CMD$
165     V% = 1: CALL IBSRE(BRDO%,V%)
170     IF CMD$ = "ex" OR CMD$ = "EX" THEN LOOP = FALSE: GOTO 205
175     IF CMD$ = "st" OR CMD$ = "ST" THEN GOSUB 285: GOTO 200
180     IF CMD$ = "rc" OR CMD$ = "RC" THEN GOSUB 365: GOTO 200
185     IF CMD$ = "lc" OR CMD$ = "LC" THEN GOSUB 240: GOTO 200
190     IF CMD$ = "" THEN GOTO 200
195     CALL IBCMD(BRDO%,CMD1$): CALL IBWRT(BRDO%,CMD$): GOSUB 270
200 WEND
205 CALL IBSIC(BRDO%): V%=0: CALL IBSRE(BRDO%,V%)
210 CALL IBSIC(BRDO%)
215 END
220 '
230 'LOCAL MODE
235 '
240 V% = 0: CALL IBSRE(BRDO%,V%): PRINT ""
245 RETURN
250 '
260 'SUBROUTINE GET_DATA
265 '
270 CALL IBCMD(BRDO%,CMD2$): CALL IBRD(BRDO%,RD$): I=IBCNT%
275 FOR J=1 TO I: PRINT MID$(RD$,J,1);: NEXT J: PRINT ""
280 RETURN
285 '
290 'SUBROUTINE STORE_DATA
295 '
300 RD1$=SPACE$(3)
305 LINE INPUT "Specify trace (EA,EB,MC,MD,FE,FF,C1,C2): ",TRACES$
310 LINE INPUT "Enter filename : ",FILE$
315 CALL IBCMD(BRDO%,CMD1$)
320 CMD$="WFSU NP,0,SP,0,FP,0,SN,0;CHDR SHORT"
321 CALL IBWRT(BRDO%,CMD$)
325 CMD$=TRACES$+"WF?": CALL IBWRT(BRDO%,CMD$)
330 CALL IBCMD(BRDO%,CMD2$): CALL IBRD(BRDO%,RD1$)
```

Appendix A

```
335 CALL IBRDF(BRDO%,FILE$)
340 IF IBSTA% < 0 THEN GOTO 430
345 PRINT ""
350 RETURN
355 '
360 'SUBROUTINE RECALL_DATA
365 '
370 LINE INPUT "Specify target memory (MC,MD): ",MEM$
375 LINE INPUT "Enter filename : ",FILE$
380 CALL IBCMD(BRDO%,CMD1$)
385 CMD$=MEM$+":TRACE ON": CALL IBWRT(BRDO%,CMD$)
390 CALL IBWRTF(BRDO%,FILE$)
395 IF IBSTA% < 0 THEN GOTO 430
400 PRINT ""
405 RETURN
410 '
415 'ERROR HANDLER
420 '
425 PRINT "IBFIND ERROR": STOP
430 PRINT "CPIB ERROR -- IBERR : ";IBERR%;"IBSTA : ";HEX$(IBSTA%)
435 STOP
440 END
```

APPENDIX B

THE WAVEFORM TEMPLATE

This is the response of the instrument to a command of the form
TMPL?

```
TMPL "  
/00  
000000          LECROY_2_1:  TEMPLATE  
                8 68 109  
;  
; Explanation of the formats of waveforms and their descriptors on the  
; LeCroy Digital Oscilloscopes,  
;   Software Release 41.1.1.2, 90/10/04.  
;  
; A descriptor and/or a waveform consists of one or several logical data blocks  
; whose formats are explained below.  
; Usually, complete waveforms are read: at the minimum they consist of  
;   the basic descriptor block WAVEDESC  
;   a data array block.  
; Some more complex waveforms, e.g. Extrema data or the results of a Fourier  
; transform, may contain several data array blocks.  
; When there are more blocks, they are in the following sequence:  
;   the basic descriptor block WAVEDESC  
;   the history text descriptor block USERTEXT (may or may not be present)  
;   the time array block (for RIS and sequence acquisitions only)  
;   data array block  
;   auxiliary or second data array block  
; With the exception of the data and time arrays, every block starts with  
; an 8-character name which identifies which kind of block it is.  
;  
; In the following explanation, every element of a block is described by a  
; single line in the form  
;  
; <byte position>  <variable name>: <variable type> ; <comment>  
;  
; where  
;  
; <byte position> = position in bytes (decimal offset) of the variable,  
;                  relative to the beginning of the block.  
;  
; <variable name> = name of the variable.  
;  
; <variable type> = string          up to 16-character name  
;                  terminated with a null byte  
;                  byte            8-bit signed data value
```

Appendix B

```

;           word           16-bit signed data value
;           long          32-bit signed data value
;           float         32-bit IEEE floating point value
;                           with the format shown below
;                           31 30 .. 23 22 ... 0 bit position
;                           s exponent fraction
;                           where
;                           s = sign of the fraction
;                           exponent = 8 bit exponent e
;                           fraction = 23 bit fraction f
;                           and the final value is
;                            $(-1)**s * 2**(e-127) * 1.f$ 
;           double        64-bit IEEE floating point value
;                           with the format shown below
;                           63 62 .. 52 51 ... 0 bit position
;                           s exponent fraction
;                           where
;                           s = sign of the fraction
;                           exponent = 11 bit exponent e
;                           fraction = 52 bit fraction f
;                           and the final value is
;                            $(-1)**s * 2**(e-1023) * 1.f$ 
;           enum          enumerated value in the range 0 to N
;                           represented as a 16-bit data value.
;                           The list of values follows immediately.
;                           The integer is preceded by an _.
;           time_stamp    double precision floating point number,
;                           for the number of seconds and some bytes
;                           for minutes, hours, days, months and year.
;
;           double seconds (0.00 to 59.999999)
;           byte minutes (0 to 59)
;           byte hours (0 to 23)
;           byte days (1 to 31)
;           byte months (1 to 12)
;           word year (0 to 16000)
;           word unused
;
;           There are 16 bytes in a time field.
;           data          byte, word or float, depending on the
;                           read-out mode reflected by the WAVEDESC
;                           variable COMM_TYPE, modifiable via the
;                           remote command COMM_FORMAT.
;           text          arbitrary length text string

```

```

;                                     (maximum 400)
;         unit_definition              a unit definition consists of a 48 character
;                                     ASCII string terminated with a null byte
;                                     for the unit name.
;
;=====
;
WAVEDESC: BLOCK
;
; Explanation of the wave descriptor block WAVEDESC;
;
;
< 0>         DESCRIPTOR_NAME: string ; the first 8 chars are always WAVEDESC
;
< 16>        TEMPLATE_NAME: string
;
< 32>        COMM_TYPE: enum          ; chosen by remote command COMM_FORMAT
              _0      byte
              _1      word
              endenum
;
< 34>        COMM_ORDER: enum
              _0      HIFIRST
              _1      LOFIRST
              endenum
;
;
; The following variables of this basic wave descriptor block specify
; the block lengths of all blocks of which the entire waveform (as it is
; currently being read) is composed. If a block length is zero, this
; block is (currently) not present.
;
;
;BLOCKS :
;
< 36>        WAVE_DESCRIPTOR: long    ; length in bytes of block WAVEDESC
< 40>        USER_TEXT: long         ; length in bytes of block USERTTEXT
< 44>        RES_DESC1: long         ;
;
;ARRAYS :
;
< 48>        TRIGTIME_ARRAY: long    ; length in bytes of TRIGTIME array
;

```

Appendix B

```
< 52>      RIS_TIME_ARRAY: long      ; length in bytes of RIS_TIME array
;
< 56>      RES_ARRAY1: long          ; an expansion entry is reserved
;
< 60>      WAVE_ARRAY_1: long        ; length in bytes of 1st simple
; data array. In transmitted waveform,
; represent the number of transmitted
; bytes in accordance with the NP
; parameter of the WFSU remote command
; and the used format (see COMM_TYPE).
;
< 64>      WAVE_ARRAY_2: long        ; length in bytes of 2nd simple
; data array
;
< 68>      RES_ARRAY2: long
< 72>      RES_ARRAY3: long          ; 2 expansion entries are reserved
;
; The following variables identify the instrument
;
< 76>      INSTRUMENT_NAME: string
;
< 92>      INSTRUMENT_NUMBER: long
;
< 96>      TRACE_LABEL: string       ; identifies the waveform.
;
<112>     RESERVED1: word
<114>     RESERVED2: word           ; 2 expansion entries
;
; The following variables describe the waveform and the time at
; which the waveform was generated.
;
<116>     WAVE_ARRAY_COUNT: long     ; number of data points in the data
; array. If there are two data
; arrays (FFT or Extrema), this number
; applies to each array separately.
;
<120>     PNTS_PER_SCREEN: long      ; nominal number of data points
; on the screen
;
<124>     FIRST_VALID_PNT: long      ; count of number of points to skip
; before first good point
; FIRST_VALID_POINT = 0
; for normal waveforms.
```

```
;  
<128>      LAST_VALID_PNT: long      ; index of last good data point  
                                                ; in record before padding (blanking)  
                                                ; was started.  
                                                ; LAST_VALID_POINT = WAVE_ARRAY_COUNT-1  
                                                ; except for aborted sequence  
                                                ; and rollmode acquisitions  
;  
<132>      FIRST_POINT: long        ; for input and output, indicates  
                                                ; the offset relative to the  
                                                ; beginning of the trace buffer.  
                                                ; Value is the same as the FP parameter  
                                                ; of the WFSU remote command.  
;  
<136>      SPARSING_FACTOR: long    ; for input and output, indicates  
                                                ; the sparsing into the transmitted  
                                                ; data block.  
                                                ; Value is the same as the SP parameter  
                                                ; of the WFSU remote command.  
;  
<140>      SEGMENT_INDEX: long      ; for input and output, indicates the  
                                                ; index of the transmitted segment.  
                                                ; Value is the same as the SN parameter  
                                                ; of the WFSU remote command.  
;  
<144>      SUBARRAY_COUNT: long     ; for Sequence, acquired segment count,  
                                                ; between 0 and NOM_SUBARRAY_COUNT  
;  
<148>      SWEEPS_PER_ACQ: long     ; for Average or Extrema,  
                                                ; number of sweeps accumulated  
                                                ; else 1  
;  
<152>      OBSOLETE1: long  
;  
<156>      VERTICAL_GAIN: float  
;  
<160>      VERTICAL_OFFSET: float   ; to get floating values from raw data :  
                                                ; VERTICAL_GAIN * data - VERTICAL_OFFSET  
;  
<164>      MAX_VALUE: float         ; maximum allowed value. It corresponds  
                                                ; to the upper edge of the grid.  
;  
<168>      MIN_VALUE: float         ; minimum allowed value. It corresponds
```

Appendix B

```

; to the lower edge of the grid.
;
<172>    NOMINAL_BITS: word      ; a measure of the intrinsic precision
;                                     ; of the observation: ADC data is 8 bit
;                                     ;   averaged data is 10-12 bit, etc.
;
<174>    NOM_SUBARRAY_COUNT: word ; for Sequence, nominal segment count
;                                     ; else 1
;
<176>    HORIZ_INTERVAL: float  ; sampling interval for time domain
;                                     ; waveforms
;
<180>    HORIZ_OFFSET: double   ; trigger offset for the first sweep of
;                                     ; the trigger, seconds between the
;                                     ; trigger and the first data point
;
<188>    PIXEL_OFFSET: double   ; needed to know how to display the
;                                     ; waveform
;
<196>    VERTUNIT: unit_definition ; units of the vertical axis
;
<244>    HORUNIT: unit_definition ; units of the horizontal axis
;
<292>    RESERVED3: word
<294>    RESERVED4: word      ; 2 expansion entries
;
<296>    TRIGGER_TIME: time_stamp ; time of the trigger
;
<312>    ACQ_DURATION: float    ; duration of the acquisition (in sec)
;                                     ; in multi-trigger waveforms.
;                                     ; (e.g. sequence, RIS, or averaging)
;
<316>    RECORD_TYPE: enum
        _0    single_sweep
        _1    interleaved
        _2    histogram
        _3    trend
        _4    filter_coefficient
        _5    complex
        _6    extrema
        _7    sequence_obsolete
        endenum
;

```



```
<318>      PROCESSING_DONE: enum
          _0      no_processing
          _1      fir_filter
          _2      interpolated
          _3      sparsed
          _4      autoscaled
          _5      no_result
          _6      rolling
          _7      cumulative
          endenum
;
<320>      RESERVED5: word          ; expansion entry
;
<322>      RIS_SWEEPS: word          ; for RIS, the number of sweeps
          ; else 1
;
; The following variables describe the basic acquisition
; conditions used when the waveform was acquired
;
<324>      TIMEBASE: enum
          _0      1_ps/div
          _1      2_ps/div
          _2      5_ps/div
          _3      10_ps/div
          _4      20_ps/div
          _5      50_ps/div
          _6      100_ps/div
          _7      200_ps/div
          _8      500_ps/div
          _9      1_ns/div
          _10     2_ns/div
          _11     5_ns/div
          _12     10_ns/div
          _13     20_ns/div
          _14     50_ns/div
          _15     100_ns/div
          _16     200_ns/div
          _17     500_ns/div
          _18     1_us/div
          _19     2_us/div
          _20     5_us/div
          _21     10_us/div
          _22     20_us/div
```

Appendix B

```
_23 50_us/div
_24 100_us/div
_25 200_us/div
_26 500_us/div
_27 1_ms/div
_28 2_ms/div
_29 5_ms/div
_30 10_ms/div
_31 20_ms/div
_32 50_ms/div
_33 100_ms/div
_34 200_ms/div
_35 500_ms/div
_36 1_s/div
_37 2_s/div
_38 5_s/div
_39 10_s/div
_40 20_s/div
_41 50_s/div
_42 100_s/div
_43 200_s/div
_44 500_s/div
_45 1_ks/div
_46 2_ks/div
_47 5_ks/div
_100 EXTERNAL
endenum
```

```
;
<328> VERT_COUPLING: enum
_0 DC_50_Ohms
_1 ground
_2 DC_1MOhm
_3 ground
_4 AC,_1MOhm
endenum
```

```
;
<328> PROBE_ATT: float
```

```
;
<332> FIXED_VERT_GAIN: enum
_0 1_uV/div
_1 2_uV/div
_2 5_uV/div
_3 10_uV/div
```

```

    _4  20_uV/div
    _5  50_uV/div
    _6  100_uV/div
    _7  200_uV/div
    _8  500_uV/div
    _9  1_mV/div
   _10  2_mV/div
   _11  5_mV/div
   _12  10_mV/div
   _13  20_mV/div
   _14  50_mV/div
   _15  100_mV/div
   _16  200_mV/div
   _17  500_mV/div
   _18  1_V/div
   _19  2_V/div
   _20  5_V/div
   _21  10_V/div
   _22  20_V/div
   _23  50_V/div
   _24  100_V/div
   _25  200_V/div
   _26  500_V/div
   _27  1_kV/div
endenum
;
<334>    BANDWIDTH_LIMIT: enum
        _0      off
        _1      on
endenum
;
<336>    VERTICAL_VERNIER: float
;
<340>    ACQ_VERT_OFFSET: float
;
<344>    WAVE_SOURCE: enum
        _0      CHANNEL_1
        _1      CHANNEL_2
        _2      CHANNEL_3
        _3      CHANNEL_4
        _9      UNKNOWN
endenum
;
```

Appendix B

```
/OO          ENDBLOCK
;
;-----
;
USERTEXT: BLOCK
;
; Explanation of the descriptor block USERTEXT at most 400 bytes long.
;
;
< 0>        DESCRIPTOR_NAME: string ; the first 8 chars are always USERTEXT
;
< 16>       TEXT: text                ; a list of ASCII characters
;
/OO          ENDBLOCK
;
;-----
;
DATA_ARRAY_1: ARRAY
;
; Explanation of the data array DATA_ARRAY_1.
; The data item is repeated for each acquired or computed data point
; of the first data array of any waveform.
;
< 0>        MEASUREMENT: data          ; the actual format of a data is
;                                       ; given in the WAVEDESC descriptor
;                                       ; by the COMM_TYPE variable.
;
/OO          ENDARRAY
;
;-----
;
DATA_ARRAY_2: ARRAY
;
; Explanation of the data array DATA_ARRAY_2.
; The data item is repeated for each acquired or computed data point
; of the second data array of any dual waveform
; e.g. the imaginary part of a FFT.
;
< 0>        MEASUREMENT: data          ; the actual format of a data is
;                                       ; given in the WAVEDESC descriptor
;                                       ; by the COMM_TYPE variable.
;
/OO          ENDARRAY
```

```
;  
=====;  
;  
TRIGTIME: ARRAY  
;  
; Explanation of the trigger time array TRIGTIME.  
; This data block is repeated for each segment which makes up the acquired  
; sequence record.  
;  
< 0>          TRIGGER_TIME: double      ; for sequence acquisitions,  
                                           ; time in seconds from first  
                                           ; trigger to this one  
;  
< 8>          TRIGGER_OFFSET: double    ; the trigger offset is in seconds  
                                           ; from trigger to zeroth data point  
;  
/OO          ENDARRAY  
;  
=====;  
;  
RISTIME: ARRAY  
;  
; Explanation of the random-interleaved-sampling (RIS) time array RISTIME.  
; This data block is repeated for each sweep which makes up the RIS record  
;  
< 0>          RIS_OFFSET: double        ; seconds from trigger to zeroth  
                                           ; point of segment  
;  
/OO          ENDARRAY  
;  
=====;  
;  
SIMPLE: ARRAY  
;  
; Explanation of the data array SIMPLE.  
; The data item is repeated for each acquired or computed data point  
; of the first data array of any waveform.  
; It is identical to DATA_ARRAY_1 above.  
;  
< 0>          MEASUREMENT: data         ; the actual format of a data is  
                                           ; given in the WAVEDESC descriptor  
                                           ; by the COMM_TYPE variable.  
;  
;
```

Appendix B

```
/OO          ENDARRAY
;
;=====
;
DUAL: ARRAY
;
; Explanation of the DUAL array.
; This data block is repeated for each pair of computed data points
; composing respectively the first and second data array of a waveform,
; e.g the real and imaginary parts of an FFT.
;
< O>        MEASUREMENT_1: data      ; data in the first data array.
;                                           ; the actual format of a data is
;                                           ; given in the WAVEDESC descriptor
;                                           ; by the COMM_TYPE variable.
;
< O>        MEASUREMENT_2: data      ; data in the second data array
;                                           ; the byte offset depends on
;                                           ; the actual format of the data which
;                                           ; is given in the WAVEDESC descriptor
;                                           ; by the COMM_TYPE variable.
;
/OO          ENDARRAY
;
;
000000      ENDTEMPLATE
"
```

INDEX

A

ALL_STATUS?, ALST, Query, 39
ARM_ACQUISITION, ARM, Command, 40
Assistance, 1
ATTENUATION, ATTN, Command/Query,
41
AUTO_CALIBRATE, ACAL, Command/
Query, 42
AUTO_SETUP, ASET, Command, 43

B

BANDWIDTH_LIMIT, BWL, Command/
Query, 44
BASIC(A), 15–22, 181, 200
Binary Blocks, 183
BUZZER, BUZZ, Command, 45

C

*CAL? Query, 46
CALL_HOST, CHST, Command/Query, 47
*CLS Command, 48
CMR – CoMmand error Register, 191, 193,
194, 196
CMR? Query, 49
COMM_FORMAT, CFMT, Command/
Query, 50–51
COMM_HEADER, CHDR, Command/
Query, 52
COMM_HELP, CHLP, Command/Query,
53
COMM_ORDER, CORD, Command/Query,
54–55
COMM_RS232, CORS, Command/Query,
56–58
Command Execution, 37
Command Notation, 37–38

Command Summary, 35–37

Commands

ALL_STATUS?, ALST, 39
ARM_ACQUISITION, ARM, 40
ATTENUATION, ATTN, 41
AUTO_CALIBRATE, ACAL, 42
AUTO_SETUP, ASET, 43
BANDWIDTH_LIMIT, BWL, 44
BUZZER, BUZZ, 45
*CAL?, 46
CALL_HOST, CHST, 47
*CLS, 48
CMR?, 49
COMM_FORMAT, CFMT, 50–51
COMM_HEADER, CHDR, 52
COMM_HELP, CHLP, 53
COMM_ORDER, CORD, 54–55
COMM_RS232, CORS, 56–58
COUPLING, CPL, 59
CURSOR_MEASURE, CRMS, 60
CURSOR_SET, CRST, 61–62
CURSOR_VALUE?, CRVA?, 63
DATE, 64
DDR?, 65–66
DEFINE, DEF (Oscilloscopes fitted with
WP01), 69–72
DEFINE, DEF (Oscilloscopes fitted with
WP02), 73–75
DEFINE, DEF (Standard Oscilloscopes),
67–68
DELETE_FILE, DELF, 76
DIRECTORY_LIST?, DIR?, 77
DISPLAY, DISP, 78
DUAL_ZOOM, DZOM, 79
*ESE, 80
*ESR?, 81–82
EXR?, 83–84
FORMAT_CARD, FCRD, 85
FUNCTION_RESET, FRST, 87
FUNCTION_STATE, FSTA (9424 only),
88–89
GRID, 90
HARDCOPY_SETUP, HCSU, 91–92
HARDCOPY_TRANSMIT, HCTR, 93

- HOR_MAGNIFY, HMAG, 94
HOR_POSITION, HPOS, 95–96
*IDN?, 97
INE, 98
INR?, 99
INSPECT?, INSP?, 100–101
INTENSITY, INTS, 102
INTERLEAVED, ILVD, 103
*IST?, 104
KEY, 105–106
MESSAGE, MSG, 107
MULTI_ZOOM, MZOM, 108
OFFSET, OFST, Command/Query,
109–110
*OPC, 111
*OPT?, 112
PANEL_SETUP, PNSU, 113
PARAMETER_VALUE?, PAVA?,
114–115
PASS_FAIL_CONDITION, PFCO, 116
PASS_FAIL_COUNTER, PFCT, 117
PASS_FAIL_DO, PFDO, 118
PASS_FAIL_MASK, PFMS, 119
PER_CURSOR_SET, PECS, 121
PER_CURSOR_VALUE?, PECV?, 123
PERSIST, PERS, 120
PERSIST_SETUP, PESU, 124
*PRE, 125
*RCL, 126
RECALL, REC, 128
RECALL_PANEL, RCPN, 127
*RST, 129
SAMPLE_CLOCK, SCLK, 130
*SAV, 131
SCREEN_DUMP, SCDP, 132
SEGMENTS, SEGS, 133
SELECT, SEL, 134
*SRE, 135
*STB?, 136–137
STOP, 138
STORE, STO, 139
STORE_PANEL, STPN, 140
STORE_SETUP, STST, 141
STORE_TEMPLATE, STTM, 143
TEMPLATE?, TMPL?, 144
TIME_DIV, TDIV, 145–146
TRACE, TRA, 147
*TRG, 148
TRIG_COUPLING, TRCP, 149–150
TRIG_DELAY, TRDL, 150
TRIG_LEVEL, TRLV, 151–152
TRIG_MODE, TRMD, 152
TRIG_PATTERN, TRPA (2-ch. & 4-ch.
with ext. trigger), 153
TRIG_PATTERN, TRPA (4-ch. without
ext. trigger), 154
TRIG_SELECT, TRSE, 155–157
TRIG_SLOPE, TRSL, 158
*TST?, 159
URR?, 160
VERT_MAGNIFY, VMAG, 161
VERT_POSITION, VPOS, 162
VOLT_DIV, VDIV, 163
*WAI, 164
WAIT, 165
WAVEFORM, WF, 166–167
WAVEFORM_SETUP, WFSU, 168–170
WAVEFORM_TEXT, WFTX, 170
XY_ASSIGN, XYAS, 171
XY_CURSOR_ORIGIN, XYCO, 172
XY_CURSOR_SET, XYCS, 173–174
XY_CURSOR_VALUE, XYCV, 175
XY_DISPLAY, XYDS, 176
ZOOM, 177
Commands and Queries, 4, 6–9
Continuous Polling, 21
Controller Timeout, 4, 13, 19, 22, 200
COUPLING, CPL, Command/Query, 59
CURSOR_MEASURE, CRMS, Command/
Query, 60
CURSOR_SET, CRST, Command/Query,
61–62
CURSOR_VALUE?, CRVA?, Query, 63
Customer Service, 1–2

D

Data

- Arrays, 179, 180
- ASCII forms, 7–9
- Blocks, 179
- Formatting, 181, 183, 188
- HEX mode, 29, 33, 182, 188
- Horizontal position, 185–187
- Interpretation, 182–185
- Sparsing, 188
- Time of, 185–187
- Values, 181, 183–186

DATE Command/Query, 64

DDR – Device Dependent error status Register, 196

DDR? Query, 65–66

DEFINE, DEF, Command/Query (Oscilloscopes fitted with WP01), 69–72

DEFINE, DEF, Command/Query (Oscilloscopes fitted with WP02), 73–75

DEFINE, DEF, Command/Query (Standard Oscilloscopes), 67–68

DELETE_FILE, DELF, Command, 76

Descriptor

- Block, 179, 183
- Values, 181, 183

Diagnostics Help Messages, 4

DIRECTORY_LIST?, DIR?, Query, 77

DISPLAY, DISP, Command/Query, 78

DUAL, 180, 181

DUAL_ZOOM, DZOM, Command/Query, 79

E

Error Messages, 4

*ESE Command/Query, 80

ESE – Standard Event Status Enable register, 20, 191, 193, 195

*ESR? Query, 81–82

ESR – Standard Event Status Register, 20, 191–194

EXR – EXecution error Register, 191, 196

EXR? Query, 83–84

F

FORMAT_CARD, FCRD, Command/Query, 85

FUNCTION_RESET, FRST, Command, 87

FUNCTION_STATE, FSTA, Command/Query (9424 only), 88–89

G

GPIB, 11

- Address switches, 3, 12, 25

- ATN – ATteNtion, 12

- Data lines, 12

- DCL – Device CLear, 13–14

- EOI – End Or Identify, 5, 13

- GET – Group Execute Trigger, 14, 19

- GTL – Go To Local, 14, 18

- Handshake lines, 12

- Hard copies, 25–27

- Hardware Configuration, 15

- IEEE 488.1, 3, 13–15, 24

- IEEE 488.2, 3, 13, 191, 194

- IFC – InterFace Clear, 12, 15

- Interface capabilities, 11

- Interface management lines, 12–13

- Listener address, 24

- LLO – Local LOkout, 14

- MLA – Listen address, 12, 24, 27

- MTA – Talker address, 12, 24, 27

- Polling, 21–25

- Programs for GPIB, 15–18, 21–26, 199–202

- REN – Remote ENable, 13, 14

RQS - ReQuest for Service, 22
SDC - Selected Device Clear, 13, 18
Signals, 12-13
Software Configuration, 15
SRE - Service Request Enable register,
19, 22
SRQ - Service ReQuest, 13, 19-20
Talker address, 24
Transfers, 15
UNL - Universal unlisten, 12, 24-25, 27
UNT - Universal untalk, 12, 24-25, 27
GRID Command/Query, 90

H

Hard Copies. *See* GPIB
HARDCOPY_SETUP, HCSU, Command/
Query, 91-92
HARDCOPY_TRANSMIT, HCTR, Com-
mand, 93
Header, 6-7
Header Path, 7
Help Messages, 4
HOR_MAGNIFY, HMAG, Command/
Query, 94
HOR_POSITION, HPOS, Command/Query,
95-96

I

IBCLR, 18
IBCMD, 23
IBFIND, 16
IBLOC, 18
IBRD, 16
IBRDF, 18
IBRDI, 18-19
IBRPP, 24
IBRSP, 22

IBTMO, 19
IBTRG, 19
IBWAIT, 19, 22
IBWRT, 16
IBWRTF, 18
IBWRTI, 18-19
*IDN? Query, 97
IEEE Standards. *See* GPIB
INE - INternal state change Enable register,
20, 22, 23, 25, 191, 193, 196
INE Command/Query, 98
INR - INternal state change Register, 20,
21, 22, 23, 25, 191, 193, 195
INR? Query, 99
INSPECT? Queries, 180-182
INSPECT?, INSP?, Query, 100-101
INTENSITY, INTS, Command/Query, 102
Interface Messages, 11
INTERLEAVED, ILVD, Command/Query,
103
IST Polling, 24-25, 193, 195
*IST? Query, 104

K

KEY Command, 105-106

L

Line Splitting. *See* RS-232-C
Local State, 5
Logical Data Blocks, 179

M

Maintenance, 1
MESSAGE, MSG, Command/Query, 107
MULTI_ZOOM, MZOM, Command/Query,
108

O

OFFSET, OFST, Command/Query,
109–110
*OPC Command/Query, 111
*OPT? Query, 112

P

PANEL_SETUP, PNSU, Command/Query,
113
PaRallel poll Enable register. *See* PRE
Parallel Polling, 23–24
PARAMETER_VALUE?, PAVA?, Query,
114–115
PASS_FAIL_CONDITION, PFCO, Com-
mand/Query, 116
PASS_FAIL_COUNTER, PFCT, Command/
Query, 117
PASS_FAIL_DO, PFDO, Command/Query,
118
PASS_FAIL_MASK, PFMS, Command/
Query, 119
PER_CURSOR_SET, PECS, Command/
Query, 121
PER_CURSOR_VALUE?, PECV?, Query,
123
PERSIST, PERS, Command/Query, 120
PERSIST_SETUP, PESU, Command/
Query, 124
Pin Assignments. *See* RS-232-C
Polling, 21–25

*PRE Command/Query, 125
PRE – PaRallel poll Enable register, 23, 24,
193, 195
Program Messages, 3–4, 5–6

R

RAN – Return Authorization Number, 2
*RCL Command, 126
RECALL, REC, Command, 128
RECALL_PANEL, RCPN, Command, 127
Remote State, 5
Response Message Form, 9–10
Return Procedure, 2
RISTIME, 180, 187
RS-232-C, 29
 Configuration, 30
 Echoing, 30
 Editing, 31
 Handshake control, 30–31
 Immediate commands, 30
 Line splitting, 32–33
 Message terminators, 31
 Pin assignments, 29–30
 Simulating GPIB commands, 33
 SRQ – Service request, 32
*RST Command, 129

S

SAMPLE_CLOCK, SCLK, Command/
Query, 130
*SAV Command, 131
SCREEN_DUMP, SCDP, Command/Query,
132
SEGMENTS, SEGS, Command/Query, 133
SELECT, SEL, Command/Query, 134
Serial Polling, 22
Service Procedure, 2
Service Requests, 191–193

SIMPLE, 180
*SRE Command/Query, 135
SRE – Service Request Enable register, 19, 191, 193, 195
SRQ – Service ReQuest, 19–21, 32, 193–194
Standard Event Status Enable Register. *See* ESE
Standard Event Status Register. *See* ESR
Status Register Reporting, 191
*STB? Query, 136–137
STB – SStatus Byte register, 20–25, 32, 191–194
STOP Command, 138
STORE, STO, Command, 139
STORE_PANEL, STPN, Command, 140
STORE_SETUP, STST, Command/Query, 141
STORE_TEMPLATE, STTM, Command, 143
Suffix Multipliers, 8–9

T

Template, 179, 180, 183, 184, 185, 205–216
TEMPLATE?, TMPL?, Query, 144
Terminators, 5, 9, 31, 184
Thumbwheel Switch. *See* GPIB Address Switches
TIME_DIV, TDIV, Command/Query, 145–146
TRACE, TRA, Command/Query, 147
Transfers. *See* GPIB
*TRG Command, 148
TRIG_COUPLING, TRCP, Command/Query, 149–150

TRIG_DELAY, TRDL, Command/Query, 150
TRIG_LEVEL, TRLV, Command/Query, 151–152
TRIG_MODE, TRMD, Command/Query, 152
TRIG_PATTERN, TRPA, Command/Query (2-ch. & 4-ch. with ext. trigger), 153
TRIG_PATTERN, TRPA, Command/Query (4-ch. without ext. trigger), 154
TRIG_SELECT, TRSE, Command/Query, 155–157
TRIG_SLOPE, TRSL, Command/Query, 158–159
Trigger Time, 180
TRIGTIME, 179, 186
*TST? Query, 159

U

URR – User Request status Register, 20, 196
URR? Query, 160
USERTEXT, 179

V

VERT_MAGNIFY, VMAG, Command/Query, 161
VERT_POSITION, VPOS, Command/Query, 162
VOLT_DIV, VDIV, Command/Query, 163

W


*WAI Command, 164
WAIT Command, 165
Warning Messages, 4
Warranty, 1
WAVEDESC. *See* Descriptor
WAVEFORM
 Command, 187
 Query, 182–187, 188
 Transfer Optimization, 188–189
Waveform Template, 205–216
WAVEFORM, WF, Command/Query,
 166–167
WAVEFORM_SETUP, WFSU, Command/
 Query, 168–170
WAVEFORM_TEXT, WFTX, Command/
 Query, 170

X

XY_ASSIGN, XYAS, Command/Query, 171
XY_CURSOR_ORIGIN, XYCO, Command/
 Query, 172
XY_CURSOR_SET, XYCS, Command/
 Query, 173–174
XY_CURSOR_VALUE, XYCV, Command/
 Query, 175
XY_DISPLAY, XYDS, Command/Query,
 176

Z

ZOOM, Command/Query, 177



ADDRESSES

US SALES OFFICES

800-5-LeCroy (automatically connects you to your local sales office)

WORLDWIDE

Argentina: Search SA, (01) 394-5882

Australia: Scientific Devices Pty.Ltd,
(03) 579-3622

Austria: Dewetron Elek. Messgeräte GmbH,
(0316) 391804

Benelux: LeCroy B.V. *31-4902-89285

Brazil: A. Santos, (021) 233 5590

Canada: Rayonics, W. Ontario:
(416) 736-1600
Rayonics, E. Ontario/Manitoba:
(613) 521-8251
Rayonics, Quebec:
(514) 335-0105
Rayonics, W. Canada:
(604) 293-1854

Denmark: Lutronic, (42) 459764

Finland: Labtronic OY, (90) 847144

France: LeCroy Sarl (1) 69073897

Germany: LeCroy GmbH, (06221) 831001
(North) (0405) 42713

Greece: Hellenic S/R Ltd., (01) 721 1140

India: Electronic Ent., (022) 4137096

Israel: Ammo, (03) 453157

Italy: LeCroy S.r.l., Roma (06) 300-9700
Milano (02) 2940-5634
(02) 204 7082

Japan: LeCroy Japan,
Tokyo (0081) 3 376 9400
Osaka (0081) 6 330 0961

Korea: Samduk Science & Ind., Ltd.:
(02) 468 04914

Mexico: Nucleoelectronica SA:
(905) 593 6043

New Zealand: E.C. Gough Ltd.,
(03) 798-740

Norway: Avantec A/S, (02) 630520

Pakistan: Electronuclear Corp.,
(021) 418087

Portugal: M.T. Brandao, Lta.,
(02) 691116

Singapore: Singapore Electronics and
Eng. Ltd., (065) 481 8888

Spain: Tempel SA,
(36) 3323.4278

Sweden: MSS AB, (0764) 68100

Switzerland: LeCroy S.A.
(022) 719 21 11

Taiwan: Topward El. Inst., Ltd.,
(02) 601 8801

Thailand: Measuretronix Ltd.,
(02) 374 2516

United Kingdom: LeCroy Ltd.,
(0235) 33 114

LeCROY

CORPORATE HEADQUARTERS

700 Chestnut Ridge Road
Chestnut Ridge, NY 10977-6499
Telephone: (914) 425-2000
TWX:(710) 577-2832
Fax: (914) 425-8967

LeCROY

EUROPEAN HEADQUARTERS

2, rue du Pré-de-la-Fontaine
P.O. Box 341
1217 Meyrin 1/Geneva, Switzerland
Tel.: (022) 719 21 11 Telex: 419 058
Fax: (022) 782 39 15

