


INPUT NORMALIZER

067-1129-00



CHARACTERISTICS
Calibration Accuracy
Maximum Rated Input Voltage

PERFORMANCE REQUIREMENT
Will guarantee 2236 Channel 1 & 2 Input capacitance of 22.0 pf \pm .25 pf when calibration procedure is followed.
140 V (DC + peak AC)

CAUTION

This unit should be handled with care to avoid shocks which could cause a deviation from factory calibrated values. It is recommended that it be returned to Tektronix for calibration if it is accidentally dropped from a height of greater than six inches.

Removing the cover on this calibration fixture will void the certification.

NOTE

For recertification, this calibration fixture must be sent to Tektronix Factory Service (Beaverton).

NO. 062-6780-01

DATE DEC 1983 (R)

COPYRIGHT © 1983
TEKTRONIX INC.
ALL RIGHTS RESERVED