

DPO-UP
Upgrades for MSO70000/C, DSA70000/B/C/D,
DPO70000/B/C/D, DPO7000/C,
MSO5000, and DPO5000 Series Instruments
Instructions

www.tektronix.com



075-1038-00

Tektronix

Copyright © Tektronix. All rights reserved. Licensed software products are owned by Tektronix or its subsidiaries or suppliers, and are protected by national copyright laws and international treaty provisions.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supersedes that in all previously published material. Specifications and price change privileges reserved.

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

Contacting Tektronix

Tektronix, Inc.
14150 SW Karl Braun Drive
P.O. Box 500
Beaverton, OR 97077
USA

For product information, sales, service, and technical support:

- In North America, call 1-800-833-9200.
- Worldwide, visit www.tektronix.com to find contacts in your area.

Kit Description

This kit provides instructions to install upgrades on MSO70000/C, DSA70000/B/C/D, DPO70000/B/C/D, DPO7000/C, MSO5000, and DPO5000 Series instruments. Once installed, the upgrades are equivalent to the factory-installed options.

For a list of available options, see the Tektronix Web site (www.tektronix.com) or your Tektronix representative.

Products

Model	Serial number
MSO70000/C Series	B010100 and above
DSA70000/B/C/D Series	B010100 and above
DPO70000/B/C/D Series	B010100 and above
DPO7000/C Series	B010100 and above
MSO5000 Series	B010100 and above
DPO5000 Series	B010100 and above

Minimum Tool and Equipment List

Required tools and equipment	Part number
Keyboard	119-6633-00 (USB) 119-6906-00 (PS-2) or equivalent

Kit Parts List

Option	Quantity	Part number ¹	Description
All	1 each	NS	LABEL, OPTION INFORMATION
All	1 each	075-1038-00	MANUAL, KIT INSTRUCTIONS
All	1 each	071-2046-XX	MANUAL, LATEST VERSION RELEASES

¹ NS – Not Saleable

Installation Instructions

To install your new option(s), you need the key number that is provided with this kit. This unique key number is valid only for the specific model number and serial number of your instrument; it will not work on any other instrument. This single key number enables all valid options for your instrument; those that were factory installed and any others that you may have previously purchased. In other words, this new key number supersedes any previous key numbers that you may already have.



CAUTION. *To avoid losing your options, do not misplace the paper label that is included with this kit. This label contains the option installation key number and should be installed on the rear panel of the instrument for future reference.*

As a backup, you may want to copy the key number and save it with the other important papers for your instrument. If you ever need to reenable the option, you need to reenter this key number.

Install Option

Follow these steps to enable your new option(s):

1. Connect a keyboard to the USB or PS-2 port of the instrument.
2. Power on the instrument and wait for it to boot.
3. If the instrument is in the button-bar mode, put the instrument into menu-bar mode. In menu-bar mode, there is a PC-style menu bar across the top of the display.
4. Pull down the Utilities menu, and then select Option Installation.
5. Follow the on-screen instructions to enter the alphanumeric key number exactly as it is printed on the label included in this kit.
6. Attach the label to the rear panel over the existing label.
7. Restart the oscilloscope product software.

NOTE. *Verify that you have the latest version of the optional application and firmware. The latest versions are downloadable from www.tek.com/software.*

Verify Option or Application Package Upgrade

Follow these steps to verify that an option or application upgrade was successful:

1. Restart the oscilloscope product software.
2. Pull down the Help menu, and then select About TekScope.
3. Verify that the option or application that you installed is in the list of installed options.

End of document