

TEKTRONIX®

This insert is provided as a supplement to the instruction manual furnished with this modified instrument. The information given in this insert supersedes that given in the manual.

This insert material Copyright © 1975 by Tektronix, Inc., Beaverton, Oregon. Printed in the United States of America. All rights reserved. Contents of this insert may not be reproduced in any form without permission of the copyright owner.

465

MOD XB

MODIFICATION INSERT

465
MOD XB

This manual insert describes the special features of MOD XB as it applies to the Tektronix 465 Oscilloscope. MOD XB deletes the 12/24 V auxiliary transformer winding (terminals 10 through 12).

NOTE

MOD XB is not compatible with field or factory installation of Option 7 (DC Operation), DM40-Series Digital Multimeters, or the MOD UA Transition Counter without first replacing the AC Power Transformer.

Options 4 (EMI) and 5 (TV Triggering) may be specified in conjunction with MOD XB as for the standard instrument.

