



Transmille software supports PDF scanning and automatic linking of data with a specific instrument / certificate



A typical PDF scanner for this application is the FUJI ScanSnap







PROCALTRACK



# Enter the PDF scan path into the settings within ProCalTrack :

Data Paths		
Certificates Database	S:\ProCal\Certificates	
Contacts Database	S:\My documents\Transmille Contacts\Contacts.mdb	
Digital Photos	C:\ProCal-Track	
Sales Database	S:\ProSales	
Stock Database	C:\ProSales	
PDF Scans	C:\Document Scans	

Setup the PDF scanner to store the PDF document to the C:\ScanTemp folder.





Virtual Job Shee	et [System ID : FO	0000001]	List.
Job No. : 000	00001397 ore testmic	Certificate Add.: TRANSMILLE LTD., UNIT 13, GRAVELLY	
Job Info Inst De	tails   Cert Address   F	Returned By Despatch Address Accessories Service Details History	Order
Job Number :	0000001397	Link Service Details Sub Contract Details Do Not Recall	
Date Received :	05/03/1999	Order Service Tupe A: Calibration Only	Mour
Order Number :		View Cert Two A Standard Callbacking	01000
Invoice Number :	RESET 24-6-9	Docs Quote Status : D: Go shoat airon	Docs
Despatch Date :	05/03/1999	Add	
Calibration Date :	05/03/1999		
Calibration Due :	03/03/2000	Link Service By:	Add
Certificate No. :	CM0743	Cert Warranty Mains Safety Fast 1/h	Scanl
Status : P : Ove	rDue (Posted)	▼ Turn-around: 7 ▼ Days Cal. mervel: 52 ▼ Weeks	
Use :		Customer Due letter sent 09/03/2000	Link
Booked in By:	1	Inst. Log     ODue letter sent 16/03/2000	PDF
Close	<u>P</u> rint Menu	Advanced Edit	Cert

The 'Link' buttons automatically copy the LATEST PDF file from the default 'C:\ScanTemp' folder to the shared document folder (as set in program options)

**TRANSNILE** SOLUTIONS IN CALIBRATION

Virtual Job Sheet [System ID : F00000001]								
Job No. : 0000001397 Instrument : Moore testmic		Certificate Add.: TRANSMILLE LTD., UNIT 13, GRAVELLY Customer Add.: ALPHA ELECTRONICS (NORTHERN) LTD.,						
Job Info Inst Details	⊧ │ Cert Address │ Re	eturned By   Despatch Address   Accessories   Service Details   History						
Job Number :	0000001397	Link Service Details Sub Contract Details Do Not Recall						
Date Received :	05/03/1999	PDF Order Causian Turney A Calibration Only						
Order Number :		View Service Type: A. Calibration only						
Invoice Number :	RESET 24-6-9	Docs Ducts Chattan Docs						
Despatch Date :	05/03/1999	Add Utite Status : U: Go-ahead given						
Calibration Date :	05/03/1999	Scan Quote No.						
Calibration Due :	03/03/2000	Link Service By:						
Certificate No. :	CM0743	Cert Warranty Mains Safety Fast 1/H						
Status : P : OverDue (Posted)  Turn-around : 7  Days Cal. Interval : 52  Weeks								
Use : Customer Due letter sent 09/03/2000								
Booked Inst. Log ODue letter sent 16/03/2000								
<u>C</u> lose	<u>P</u> rint Menu	Advanced Edit						

The 'Add Scan' button automatically copies the LATEST PDF file from the default 'C:\ScanTemp' folder to a specific file & folder named using the Instrument ID Number, Manufacturer and model number.



Link PDF

The filename can be modified prior to copying to the server



SOLUTIONS IN CALIBRATION