



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Load Cell
 Shear Beam
 Model: KIS-3
 n_{\max} Single Cell: 3 000
 Capacity: 450 to 45 000 lb
 Accuracy Class: III

Submitted By:

Vishay Transducers
 3 Edgewater Drive
 Norwood, MA 02062
 Tel: 781-298-2200
 Fax: 781-762-3988
 Contact: Bruce Yohr
 Email: bruce.yohr@vishaypg.com
 Web site: www.vishaypg.com


Standard Features and Options


The specific models of load cells covered by this Certificate are identified by the capacity of each load cell represented. The specific load cell capacities, v_{\min} and minimum dead loads are listed below.

Capacity(lb)	v_{\min} (lb)	Minimum Dead Load (lb)
450	0.10	5
1 125	0.25	5
2 250	0.50	5
4 500	1.125	5

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.


 Tim Tyson
 Chairman, NCWM, Inc.


 Randy Jennings
 Chairman, National Type Evaluation Program Committee
 Issued: May 19, 2011

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Vishay Transducers

Load Cell / KIS-3

Application: The load cells may be used in Class III scales for both single and multiple cell applications consistent with the model designations and parameters specified in this certificate.

Test Conditions: This Certificate supersedes Certificate of Conformance number 87-061A1 and was issued without additional testing to reactivate Certificate of Conformance number 87-061A1 without lapse. Changes were also made to update the contact information.

Certificate of Conformance Number 87-061A1: This Certificate supersedes Certificate of Conformance number 87-061 and was issued without additional testing to reactivate Certificate of Conformance number 87-061 without lapse. Changes were also made to update the contact information. Previous test conditions are listed below for reference.

Certificate of Conformance Number 87-061: A 1 125 lb capacity load cell was tested to capacity using a dead weight machine. The data were analyzed for single load cell applications. The cell was tested over a temperature range of -10 °C to 40 °C. Three tests were run on the load cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test was waived due to the insensitivity of the load cell design to changes in barometric pressure. The manufacturer's laboratory was used to collect the test data.

Representatives from the National Bureau of Standards evaluated the manufacturer's test facility, witnessed repeat tests on the load cells, and analyzed the data. The results indicate that the load cells comply with the applicable requirements of NBS Handbook 44.

This certificate replaces Provisional Certificate of Conformance Number 87-061P.

Evaluated By: NIST Force Group, NBS Office of Weights and Measures (Formerly NBS) 87-061

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2011. NCWM, Publication 14: Weighing Devices, 2011.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM) 87-061A1, 87-061A2

Example of Device:

