

| | | | |
|-----------|--------------------------------|---------------------|----------------------------|
| Reference | Fluke | Calibration date | October 04 2020 |
| Ref P/N | 732A | Ambient Temperature | 24.63 °C |
| Serial | XFER | Relative Humidity | 51.12 % |
| ID Number | Calibration test, as received; | Pressure | 1020.65 hPa |
| Notes | Test as received, 5h warmup | Test type | Front PTFE cable terminals |

| Reference standard | Mfg | Model | Options | Serial / Unc | CEID | Calibration date | Due date |
|--------------------|-------|-------|---------------|--------------|------|------------------|----------|
| DCV REFERENCE | Fluke | 732B | 10.00000592 ? | 7350002 | None | Invalid | Invalid |
| DMM | HP | 3458A | | FF-4 | | None | None |

REDACTED. Actual measurement uncertainty available upon request was calculated using the expanded method and is expressed in values at approximately the 95% confidence level using a coverage factor of K= 2.

Certificate statements are based on test results within specified limits without reduction of the uncertainty of the test and/or measurement. The test and measurement data here relate only to the item tested and/or measured. Unit acceptance of failure includes uncertainty data compilation. Calibration due date that appears on the Certificate of Calibration and labels are determined by the customer and does not imply conformance to a standard.



UUT output transferred by manual ratiometric measurement with reference standard.



Fixed 1.000e+01 range is used on the Keysight 3458A/X02 detector. The following test use 10 minute transfer specification with Fluke 732A output source as reference. Gain verified for stability ±0.10 ppm over the test period. Detector zero offset is DUT is nulled prior to the measurement.

Configuration : Battery power STD, NPLC100, NDIG8, Guard is open.

| | Measurement | Unit | Uncertainty | Standard Deviation | DUT Spec / Δ | Degree of freedom / Notes |
|-------------------------------------|-------------------|------------|-------------------|-----------------------------|----------------|---------------------------|
| Transfer reference output | 10.0000059 | VDC | ±0.100 ppm | | | |
| Reference measured output (+) | 10.0000128 | VDC | ±0.100 ppm | $\sigma = 2.317779e-07$ VDC | Δ = 0.688 ppm | 35 |
| Reference measured output (-) | -10.0000118 | VDC | ±0.100 ppm | $\sigma = 2.295406e-07$ VDC | Δ = 0.584 ppm | 35 |
| Reference calculated +/- | 10.0000123 | VDC | ±0.100 ppm | | Δ = 0.636 ppm | |
| Detector zero offset | -0.0000000 | VDC | | $\sigma = 0.000000e+00$ VDC | | |
| UUT measured output (+) | 10.0001402 | VDC | ±0.100 ppm | $\sigma = 3.265449e-07$ VDC | | 35 |
| UUT measured output (-) | -10.0001391 | VDC | ±0.100 ppm | $\sigma = 0.2755$ μVDC | | 35 |
| Ratio positive polarity | 1.00001274 | | ±0.200 ppm | | | Inf |
| Ratio negative polarity | 1.00001273 | | ±0.200 ppm | | | Inf |
| UUT calculated output (+) | 10.0001333 | VDC | ±0.300 ppm | | Δ = 0.005 ppm | |
| UUT calculated output (-) | -10.0001332 | VDC | ±0.300 ppm | | Δ = -0.005 ppm | |
| Temperature Δ | -0.537 | °C | ±1.00 % | | ±1.0 °C | |
| UUT transfer result (Linear) | 10.0001333 | VDC | ±0.300 ppm | | 0.1% | In spec |
| UUT transfer result (RSS) | 10.0001333 | VDC | ±0.224 ppm | | 0.1% | In spec |

[Statistics image data](#)





Test procedure : \$Id\$

Lab temperature maintained +23°C ±1°C

xDevs.com Confidential

| RAW data | Result |
|-------------|---|
| Array Ref P | [10.00001253, 10.00001237, 10.00001243, 10.00001272, 10.00001269, 10.00001269, 10.00001244, 10.0000125, 10.0000128, 10.00001294, 10.0000128, 10.00001258, 10.00001269, 10.00001281, 10.00001255, 10.00001288, 10.00001288, 10.00001265, 10.0000125, 10.00001264, 10.00001297, 10.0000128, 10.00001271, 10.00001285, 10.00001283, 10.00001262, 10.00001253, 10.0000128, 10.0000126, 10.00001281, 10.00001283, 10.00001255, 10.0000131, 10.00001343, 10.00001341] |
| Array Ref N | [-10.00001158, -10.00001228, -10.00001225, -10.00001172, -10.00001174, -10.00001163, -10.00001139, -10.00001139, -10.00001121, -10.00001153, -10.00001172, -10.00001165, -10.00001153, -10.00001128, -10.00001156, -10.00001181, -10.00001193, -10.00001218, -10.00001184, -10.00001188, -10.00001153, -10.00001163, -10.00001176, -10.00001181, -10.00001211, -10.000012, -10.00001186, -10.00001191, -10.00001204, -10.00001179, -10.00001184, -10.0000119, -10.00001176, -10.00001177, -10.00001176] |
| Array UUT P | [10.00013996, 10.00014006, 10.00013999, 10.00013992, 10.00014008, 10.00013996, 10.00013994, 10.00014017, 10.00014052, 10.00013989, 10.00013985, 10.00014017, 10.00014091, 10.00014057, 10.00014022, 10.00014011, 10.00013987, 10.00013959, 10.00013999, 10.0001401, 10.00013978, 10.00013985, 10.00014015, 10.00014019, 10.00013996, 10.00014068, 10.00014054, 10.00014029, 10.00014056, 10.0001402, 10.00014061, 10.0001405, 10.00014075, 10.0001405, 10.00014052] |
| Array UUT N | [-10.00013913, -10.00013944, -10.00013971, -10.00013959, -10.00013913, -10.00013927, -10.0001389, -10.00013915, -10.0001392, -10.00013821, -10.00013923, -10.00013886, -10.00013877, -10.00013881, -10.0001393, -10.00013915, -10.00013969, -10.00013959, -10.00013904, -10.00013907, -10.00013886, -10.00013874, -10.00013888, -10.00013906, -10.00013892, -10.00013918, -10.00013902, -10.00013899, -10.000139, -10.00013904, -10.00013929, -10.00013885, -10.00013907, -10.00013943, -10.00013927] |

Histogramm

