



xDevs.com FX , LTZ1000ACH reference, 10V output stability, V/°C



xDevs.com FX DC Reference




CALIBRATION DATA


Date 8/14/2020
 T_{CAL, air} +23.0°C 46%RH

V_{EMF} 10.0000361 VDC
U = ±2.4 ppm, σ = 0.06 ppm


T_{nominal} : +18 ... +28 °C
 V+ nominal : +10.8 ... +13.6 VDC
 V- nominal : -10.8 ... -13.6 VDC



xDevs.com FX , LTZ1000ACH reference, 10V output stability, V/°C



xDevs.com FX DC Reference



CALIBRATION DATA

Date 8/16/2020
 T_{CAL, air} +23.0°C 46%RH

V_{EMF} 9.9999795 VDC
U = ±2.7 ppm, σ = 0.07 ppm

T_{nominal} : +18 ... +28 °C
 V+ nominal : +10.8 ... +13.6 VDC
 V- nominal : -10.8 ... -13.6 VDC

